[J. Res. Natl. Inst. Stand. Technol. 103, 593 (1998)]

A Problem in AC Quantized Hall Resistance Measurements and a Proposed Solution

Volume 103	Number 6	November-December 1998
M. E. Cage and A. Jeffery	In all experiments reported to date the	to do accurate measurements with high pre-
National Institute of Standards and Technology, Gaithersburg, MD 20899-0001	measured values of the ac quantized Hall resistances $R_{\rm H}$ varied with the frequency of the applied current, and differed signifi- cantly from the dc values of $R_{\rm H}$, making it difficult to use the ac quantum Hall ef- fect as an absolute impedance standard. We analyze the effects due to the large ca- pacitances-to-shields existing in the sample probes on measurements of $R_{\rm H}$ to see if this is the source of the problem. Equivalent electrical circuits are utilized: they con-	cision ac bridges. These large out-of- phase contributions could introduce the lin- ear frequency dependences observed in previous $R_{\rm H}$ measurements. We predict, however, that quadruple-series connec- tions to the quantum Hall devices yield only small out-of-phase contributions to $V_{\rm H}$ which may allow accurate measurements of the quantity $R_{\rm H} - R_x$, where R_x is the longitudinal resistance along the device.
	tain capacitances and leakage resistances to the sample probe shields, longitudinal re- sistances within the quantized Hall effect devices, and multiple connections to the devices. The algebraic solutions for the $R_{\rm H}$	Key words: ac quantum Hall effect; capac- itance-to-shield; equivalent electrical cir- cuit; longitudinal resistance; multi-series connections; quantized Hall resistance.
	values in these circuits reveal large out- of-phase contributions to the quantized Hall	Accepted: September 15, 1998
	voltages $V_{\rm H}$ that would make it difficult	Available online: http://www.nist.gov/jres

1. Background

The quantum Hall effect (QHE) has been successfully used as an intrinsic dc resistance standard. In the integer dc QHE [1–3], the Hall resistance $R_{\rm H}$ of the *i*th plateau of a fully-quantized, two-dimensional electron gas (2DEG) is $R_{\rm H}(i) = V_{\rm H}(i)/I_{\rm T}$, where $V_{\rm H}(i)$ is the quantum Hall voltage measured between potential probes located on opposite sides of the device, and $I_{\rm T}$ is the total current flowing between the source and drain current contacts at the ends of the device. Under ideal conditions, the values of $R_{\rm H}(i) = h/(e^2i) = R_{\rm K}/i$, where *h* is the Planck constant, *e* is the elementary charge, *i* is an integer, and $R_{\rm K}$ is the von Klitzing constant $R_{\rm K} \approx$ 25 812.807 Ω . However, the conditions are not always ideal. The values of $R_{\rm H}(i)$ can vary with the device temperature *T* and with the frequency *f* of the applied current if it is not dc. Thus the measured values of $R_{\rm H}(i)$ are not necessarily equal to $h/(e^2i)$.

The current flow within the 2DEG is nearly dissipationless in the quantum Hall plateau regions of highquality devices, and the longitudinal resistances $R_x(i)$ of this standard become very small over ranges of magnetic field that exhibit quantized Hall resistance plateaus. The dc longitudinal resistance is defined to be $R_x(i) = V_x(i)/I_T$, where $V_x(i)$ is the measured longitudinal voltage drop between potential probes located on the same side of the device.

- [7] A. Hartland, B. P. Kibble, P. J. Rodgers, and J. Bohacek, AC Measurements of the Quantized Hall Resistance, IEEE Trans. Instrum. Meas. 44, 245–248 (1995).
- [8] J. Bohacek, P. Svoboda, and P. Vasek, AC QHE-Based Calibration of Resistance Standards, IEEE Trans. Instrum. Meas. 46, 273–275 (1997).
- [9] A. Hartland, B. P. Kibble, and S. W. Chua, AC Quantized Hall Resistance Experiments at the National Physical Laboratory, BEMC Π97 Conference Digest, (National Physical Laboratory, Teddington, U. K. (1997) pp. 18-1 to 18-4.
- [10] F. Piquemal, G. Trapon, and G. Geneves, AC Measurements of the Minimum Longitudinal Resistance of a QHE Sample from 10 Hz to 10 kHz, IEEE Trans. Instrum. Meas. 45, 918–922 (1996).
- [11] F. Cabiati, L. Callegaro, C. Cassiago, V. D'Elia, and G. M. Reedtz, Effect of the Frequency on the Longitudinal Resistance of a GaAs-AlGaAs Heterostructure, CCE/97-32, 1–5 (1997).
- [12] M. E. Cage and A. Jeffery, Intrinsic Capacitances and Inductances of Quantum Hall Effect Devices, J. Res. Natl. Inst. Stand. Technol. 101, 733–744 (1996).
- [13] F. Delahaye, Series and Parallel Connection of Multiterminal Quantum Hall Effect Devices, J. Appl. Phys. 73, 7915–7920 (1993).
- [14] B. W. Ricketts and P. C. Kemeny, Quantum Hall Effect Devices as Circuit Elements, J. Phys. D: Appl. Phys. 21, 483–487 (1988).
- [15] M. E. Cage, B. F. Field, R. F. Dziuba, S. M. Girvin, A. C. Gossard, and D. C. Tsui, Temperature Dependence of the Quantum Hall Resistance, Phys. Rev. B 30, 2286–2288 (1984).
- [16] A. Jeffery, R. E. Elmquist, and M. E. Cage, Precision Tests of a Quantum Hall Effect Device dc Equivalent Circuit using Double-Series and Triple-Series Connections, J. Res. Natl. Inst. Stand. Technol. 100, 677–685 (1995).
- [17] R. D. Cutkosky, Four-Terminal-Pair Networks as Precision Admittance and Impedance Standards, Communication Electron. 70, 19–22 (1964).
- [18] R. D. Cutkosky, Techniques for Comparing Four-Terminal-Pair Admittance Standards, J. Res. Natl. Inst. Stand. Technol. 74C, 63–78 (1970).
- [19] M. E. Cage, R. F. Dziuba, R. E. Elmquist, B. F. Field, G. R. Jones, P. T. Olsen, W. D. Phillips, J. Q. Shields, R. L. Steiner, B. N. Taylor, and E. R. Williams, NBS Determination of the Fine-Structure Constant and of the Quantized Hall Resistance and Josephson Frequency-to-Voltage Quotient in SI Units, IEEE Trans. Instrum. Meas. 38, 284–289 (1989).
- [20] A. Jeffery, R. E. Elmquist, L. H. Lee, J. Q. Shields, and R. F. Dziuba, NIST Comparison of the Quantized Hall Resistance and the Realization of the SI OHM Through the Calculable Capacitor, IEEE Trans. Instrum. Meas. 46, 264–268 (1997).
- [21] F. Delahaye, Technical Guidelines for Reliable Measurements of the Quantized Hall Resistance, Metrologia 26, 63–68 (1989).

About the authors: M. E. Cage, A. Jeffery, R. E. Elmquist, and K. C. Lee are physicists in the Electricity Division of the NIST Electronics and Electrical Engineering Laboratory. The National Institute of Standards and Technology is an agency of the Technology Administration, U.S. Department of Commerce.

Volume 103, Number 6, November-December 1998 Journal of Research of the National Institute of Standards and Technology



Fig. 1. An equivalent electrical circuit representation of a QHE resistance standard while the quantized Hall resistance is being measured in an ac ratio bridge using four-terminal-pair [14,15] measurement techniques. The ac ratio bridge is not shown, nor is the ac reference resistance standard with which the QHE standard is being compared. The symbols are explained in Sec. 2.1. See Sec. 3.1 for the circuit analysis.

standard and not shown in the figure) assure that the equal/opposite current conditions are satisfied. The current I_{Ot} exits the QHE standard at the Inner/Outer port and enters the ac reference standard (not shown).

A short has been drawn in Fig. 1 between the shield and inner conductor at the Detector coaxial port to indicate four-terminal-pair condition (2). We let the Detector potential be zero, i.e., $V_{\text{Dt}} = 0$. At bridge balance the quantized Hall resistance $R_{\text{H}}(i)$ is defined as

$$V_{\rm Pt} = [1 + \Delta] R_{\rm H}(i) I_{\rm Ot} , \qquad (1)$$

where Δ is the correction factor to $R_{\rm H}(i)$ to be determined in this analysis.

Next we describe the equivalent circuit model of the QHE device. The device has contact pads that provide electrical access to the 2DEG at the source S', the drain D', and the potential pads 1' through 6'. Each contact pad is located at the end of an arm of the device. Every arm in the equivalent circuit has an intrinsic resistor

whose value is $R_{\rm H}(i)/2$. We assume that the device is homogeneous, i.e., that the quantized Hall resistances $R_{\rm H}(i)$ are all measured on plateau regions, that their values are the same on all the Hall potential probe sets, and that they are all measured at the same magnetic flux density value. $R_{\rm H}(i)$ can, however, vary with temperature and frequency.

A positively-charged applied current I_a enters the 2DEG via device drain contact pad D', and current I_d exits the 2DEG via source contact pad S'. The magnetic flux density *B* is directed into the figure. Under these current and magnetic field conditions the drain contact pad D' and the potential probe contact pads 1', 3', and 5' at the device periphery are at higher potentials than contact pads S', 2', 4', and 6'. These current and flux density directions are chosen to be consistent with those we have used in earlier calculations [12,13,17].

Potentials at the contact pads S', 1' through 6', and D' are produced by voltage generators V_{AB} located between

Many laboratories are now attempting to employ the QHE to realize an intrinsic ac resistance standard by using ac ratio bridges to compare the ac quantized Hall resistances R_H with ac reference standards. In experiments reported to date [4-9], the measured values of the ac quantized Hall resistances $R_{\rm H}$ vary with the applied frequency of the current f (usually increasing linearly with f), and differ from the dc value of $R_{\rm H}$ by at least the factor $10^{-7} R_{\rm H}(i)$ at a frequency f of 1592 Hz ($\omega = 2\pi f$ is 10⁴ rad/s). With one exception [10], the reported ac longitudinal resistances $R_x(i)$ are significantly larger than the dc longitudinal resistances in the same device under the same temperature and magnetic field conditions. The ac longitudinal resistances increase with increasing frequency of the applied current, and are of order 1 m Ω at 1592 Hz [4,5,11]. The frequency dependences of $R_{\rm H}$ and $R_{\rm x}$ are reported to be a property of the real, resistive (in-phase) component of the ac impedance measurements.

These results might arise from intrinsic properties of the quantum Hall devices. However, in a previous publication [12] we showed that the intrinsic impedance due to the internal Hall capacitance of the two-dimensional electron gas across the QHE device does not account for the observed frequency dependences of the in-phase ac quantized Hall resistances $R_{\rm H}$. We also showed [12] that the kinetic inductance of the 2DEG and the magnetic inductance of the device provide no plausible intrinsic impedance explanations for the observed frequency dependences of the in-phase ac longitudinal resistances R_x . Other calculations [13] showed that the intrinsic longitudinal resistances of the QHE device have very little effect on $R_{\rm H}$ measurements.

In this paper we investigate what effects the large capacitances-to-shield in the sample probe can have on the measured ac values of $R_{\rm H}$.

2. Equivalent Electrical Circuit of an AC QHE Standard

We are in the process of deriving exact algebraic solutions for the currents and potentials of equivalent electrical circuits containing (a) multiple connections to the device; (b) capacitances to the sample probe shields; (c) longitudinal resistances within the device; and (d) leakage resistances to the probe shields. Finding the exact algebraic equations for $R_{\rm H}$ and R_x is rather difficult, and the solutions are nontrivial.

The exact algebraic solutions for the $R_{\rm H}$ values in these circuits reveal large 90° out-of-phase (reactive) contributions to the quantized Hall voltages $V_{\rm H}$ that would make it very hard to perform accurate in-phase (resistive) measurements with high precision ac bridges because it is difficult to measure to sufficient accuracy the phase defect (the in-phase or resistive part) of the components used to null the out-of-phase (reactive) signal. Preliminary tests at NIST suggest that these large out-of-phase contributions could introduce linear frequency dependences into the in-phase measurements of $R_{\rm H}$.

We find, however, that quadruple-series connections to the quantum Hall devices yield only small out-ofphase contributions to $V_{\rm H}$ that should allow accurate measurements of $R_{\rm H}$ with small uncertainties. Analyzing these equivalent circuits is a long process, so we present this preliminary report of the results because it may be of use to other laboratories that are preparing, or making, ac QHE measurements.

2.1 Circuit Description

Figure 1 shows an equivalent electrical circuit representation of a QHE resistance standard while the standard is being measured in an ac ratio bridge. The quantized Hall resistance $R_{\rm H}(i)$ of the QHE standard is compared with the resistance of an ac reference standard using four-terminal-pair [14,15] measurement techniques. The ac ratio bridge is not shown in the figure, nor is the ac reference standard with which the QHE standard is being compared. Only the QHE standard is shown. This figure is rather detailed, so we explain it one step at a time.

The QHE standard is bounded by an electrical shield indicated by thick lines. This shield is also referred to in the text as outer conductors. To simplify the figure, we label only currents in the inner conductors. A QHE device occupies the central region of the figure. This device is modeled as an equivalent electrical circuit. There are additional components of the ac QHE standard. They are also modeled as circuit elements. These components are described below. The standard has electrical access via coaxial measurement ports labeled Inner/Outer, Detector, Potential, and Drive. (A coaxial port is often referred to in the literature as a terminalpair.)

The ideal four-terminal-pair measurement definition [14,15] of $R_{\rm H}$ is satisfied by the following three simultaneous conditions: (1) The current $I_{\rm Dr}$ at the Drive coaxial port is adjusted so that there are no currents in the inner or outer conductors of the Potential coaxial port, i.e., $I_{\rm Pt} = 0$. (2) The potential difference is zero across the inner and outer conductors of the Detector coaxial port. (3) There are no currents in the inner or outer conductors of the Detector coaxial port.

It is implicit in the four-terminal-pair definition that the ports are treated as terminal-pairs, and that the current in the inner conductor of every port is equal and opposite to the current in the outer conductor (the shield). Coaxial chokes [16] (located outside the QHE

3. Circuit Analyses

We use Kirchoff's rules to sum the currents at branch points and the voltages around loops to obtain exact algebraic equations for equivalent electrical circuits of ac QHE standards. Finding the exact algebraic solutions for all the currents, and for the correction factor Δ as defined by Eq. (1), is rather difficult because there are many coupled equations. Our criteria for obtaining the solutions is that both authors independently derive the same equations, and that computer software verifies the results. It will take time to complete this task, so we present here approximate solutions for some of the currents and for the correction factor Δ . Only the largest terms are included in the approximate solutions. However, it was necessary to carry smaller terms in the intermediate approximate equations because the larger terms sometimes unexpectedly canceled.

3.1 Single-Series Connections

The ac QHE standard shown in Fig. 1 has one current lead connected to the source contact pad S' and another current lead connected to the drain pad D'. We refer to this wiring configuration as two single-series connections to the device.

The shunt currents I_{C_5} , I_{C_1} , I_{C_D} , I_{C_A} , and I_{K_A} are much larger than shunt currents I_{C_2} , I_{C_4} , I_{C_6} , I_{C_5} , I_{C_8} , and I_{K_B} because contact pads 5', 3', 1', and D' are all near the quantum Hall potential, rather than near the shield potential. The room temperature access points 1, 2, 5, and 6 are open-circuited in the figure. There may be significant antenna noise generated in coaxial leads 1 and 5 because they are near the quantum Hall potential, but we ignore this problem.

We will present the complete list of exact and approximate current solutions in the full paper; only three shunt currents of particular interest are given here:

$$I_{C_{s}} \approx \{ [\omega^{2}C_{s}C_{5}R_{H}r_{s} + \omega^{2}C_{5}C_{5}R_{H}r_{5}] + j[\omega C_{5}R_{H}] \} I_{Ot}$$
(3a)

$$I_{C_3} \approx \{-[\omega^2 C_3 C_5 R_H R_H] + j[\omega C_3 R_H]\}I_{Ot}$$
 (3b)

$$I_{C_1} \approx \{-[\omega^2 C_1 (C_3 + C_5) R_H R_H] + j[\omega C_1 R_H]\} I_{Ot}.$$
 (3c)

AC QHE experimental values (with cardinal numbers) can be assigned to the circuit elements to estimate the shunt currents. For example, both the i = 2 (12 906.4 Ω) and i = 4 (6 453.2 Ω) plateaus have been measured in ac experiments, so let $R_{\rm H} = 10\ 000\ \Omega$.

$$R_{\rm H} = 10^4 \,\Omega \tag{4a}$$

$$r_{\rm S} = r_1 = r_2 = r_3 = r_4 = r_5 = r_6 = r_{\rm D} = 1 \ \Omega$$
 (4b)

$$r_{\rm a} = r_{\rm b} = r_{\rm c} = r_{\rm d} = 10^{-3} \,\Omega \tag{4c}$$

$$r_{\rm K_A} = r_{\rm K_B} = 10^{12} \,\Omega \tag{4d}$$

$$C_{\rm S} = C_1 = C_2 = C_3 = C_4 = C_5 = C_6 = C_{\rm D} = C_{\rm A} = C_{\rm B} = 10^{-10} \,\mathrm{F} \,(4\mathrm{e})$$

$$\omega = 10^4 \text{ rad/s} . \tag{4f}$$

Note that, with care, the leakage resistances r_{K_A} and r_{K_B} can be at least 10^{14} , but dirty contacts or poor insulation can make them worse than the $10^{12} \Omega$ assumed in Eq. (4d). The capacitances C_A and C_B are difficult to estimate. They may have been rather large in some experiments because grounded disks were placed close to the bonding wires on the device holders, so we let them be 10^{-10} F in Eq. (4e).

Figure 2 shows an enlargement of Fig. 1 in the vicinity of the Potential coaxial port, plus the approximate numerical values of I_{C_5} , I_{C_3} , and I_{C_1} calculated from Eqs. (3a) to (3c). Note that a 1 % out-of-phase current passes through each of the coaxial cable capacitances C_5 , C_3 , and C_1 in this example (and also through C_D and C_A). That is not necessarily a problem if the bridge drive can provide this extra 5 % of 90° out-of-phase current to I_{Dr} . It is a problem, however, if these out-of-phase currents generate unwanted signals in the quantized Hall voltage $V_{\rm H}$.

 $V_{\rm H} = V_{\rm Pt}$ is obtained by summing the voltages between the inner conductors of the Detector coaxial port and the Potential coaxial port; taking the path through arm 4, voltage generators $V_{\rm c4}$ and $V_{\rm c3}$, and arm 3 we find that

$$V_{\rm Pt} = R_{\rm H} I_{\rm c} - r_3 I_{\rm C_3} \,, \tag{5a}$$

or approximately

$$V_{\rm Pt} \approx \{ [1 + \omega^2 C_{\rm S} C_5 R_{\rm H} r_{\rm S} + \omega^2 C_3 C_5 R_{\rm H} r_3 + \omega^2 C_5 C_5 R_{\rm H} r_5] \}$$

+
$$j[\omega C_5 R_H] R_H I_{Ot}$$
 (5b)

$$V_{\rm Pt} = [1 + \Delta] R_{\rm H}(i) I_{\rm Ot}$$
 (5c)

The numerical value of V_{Pt} in Eq. (5b) for this typical experimental example is

$$V_{\rm Pt} \approx \{1 + [3 \times 10^{-8}] + j[1 \times 10^{-2}]\} R_{\rm H} I_{\rm Ot}$$
. (6)

The 3×10^{-8} correction factor in the real part of the $V_{\rm Pt}$ signal is quite large compared with the 2.4×10^{-8} relative combined standard uncertainty of the complete measurement sequence at NIST [21] between the quantized Hall resistance and the calculable capacitor. This measurement chain is used to assign an SI value to $R_{\rm H}$ and to realize the ohm. The ac QHE would replace or

arms A and B of the equivalent circuit. The voltages are defined as

$$V_{\rm AB} \equiv \frac{R_{\rm H}(i)}{2} \left| I_{\rm A} \pm I_{\rm B} \right| \,, \tag{2}$$

where I_A and I_B are the magnitudes of the current flowing in arms A and B. The currents I_A and I_B within the absolute quantity sign of Eq. (2) are added if they both enter or both leave the voltage generator, and are subtracted if one current enters and the other current leaves the generator. For example $V_{1D} = [R_H(i)/2]I_a - I_{C_1}I$. The voltages generated are functions of $R_H(i)$; therefore their values can vary with temperature and frequency.

Diamond-shaped voltage generator arrays of Ricketts and Kemeny [18] are employed in the equivalent circuit, rather than the ring-shaped arrays used first by Delahaye [19] and then by Jeffery, Elmquist, and Cage [20]. The calculations are much simpler with the diamond arrays when longitudinal resistances are included in the circuits [13]. For clarity, the voltage generators are indicated in the figure as batteries, with positive terminals oriented to give the correct potentials along each arm. The ac currents alternate direction, so the voltage generators reverse sign each half cycle. Thus, for the part of the period in which the currents flow in the directions indicated in Fig. 1, the voltage generators have the polarities shown. Half a period later the currents change direction, and all the voltage generators reverse polarities.

The QHE device is mounted at the bottom of a sample probe. Coaxial leads extend from the device contact pads S', 1' through 6', and D' to connection points S, 1 through 6, and D located outside the cryostat. Each arm of the equivalent circuit has a resistance r_s , r_1 through r_6 , or $r_{\rm D}$. This resistance includes the contact resistance to the 2DEG, the wire resistance connecting a contact pad on the device to a coaxial sample probe lead, and the inner conductor resistance of that coaxial sample probe lead. The inner conductor lead resistances vary with the liquid helium level in the sample probe. They can be measured pair-wise (using access points S, 1 through 6, and D) as a function of liquid helium level via two-terminal resistance measurements by temporarily replacing the quantum Hall device with electrical shorts at positions S', 1' through 6', and D'. The inner conductor coaxial lead resistances are typically each about 1 Ω in ac quantized Hall resistance experiments.

The symbols r_a , r_b , r_c , and r_d in Fig. 1 represent real (in-phase) longitudinal resistances. Sample probes used in dc QHE measurements have a pair of leads to the source contact pad S' and another pair to the drain contact pad D'. Only one lead of each pair carries the current, so all four dc resistances r_a , r_b , r_c , and r_d can be obtained using four-terminal measurements. In order to reduce heat loss, sample probes for the ac QHE have a single coaxial lead to each of the contact pads. Therefore only r_b and r_c can be determined directly via ac measurements. Values for r_a and r_d could be estimated from their dc r_a/r_b and r_d/r_c ratios if the r_b/r_c ratio is the same for both ac and dc measurements. Typical ac r_b and r_c values are reported to be about 1 m Ω at 1592 Hz.

The coaxial leads each have an inner and an outer conductor. The outer conductor coaxial lead resistances are also typically each about 1 Ω in ac quantized Hall resistance experiments. The outer conductors of the coaxial leads are connected together outside the cryostat. They act as electrical shields, and are represented schematically as thick lines in Fig. 1. Large capacitances-to-shield, labeled as C_S , C_1 through C_6 , and $C_{\rm D}$, exist between the inner and outer conductors of these coaxial leads. The open-circuit capacitances can be measured at points S, 1 through 6, and at D as a function of liquid helium level by temporarily disconnecting the coaxial leads at the device contact pads S', 1' through 6', and D'. The capacitance-to-shield of each coaxial lead in typical ac QHE sample probes is at least 100 pF (1×10^{-10} F). A predominately 90° out-ofphase current I_{C_5} , I_{C_1} through I_{C_6} , or I_{C_D} flows through each coaxial lead. These currents have the correct signs in Fig. 1 for this half-cycle.

The outer conductors of the coaxial leads are not the only components of the thick-lined shield in Fig. 1. The grounded outer shell of the sample probe, and variously shaped surfaces placed near the QHE devices have contributed additional inner conductor capacitances-toshield, $C_A + C_B$, in experiments reported to date. The additional capacitances-to-shield are labeled $C_{\rm A}$ and $C_{\rm B}$, and are placed at either end of the QHE device in the figure. (Note that rather than explicitly using C_A and C_B , one-eighth of the additional capacitances $C_{\rm A} + C_{\rm B}$ could instead be added to each of the eight coaxial lead capacitances C_S , C_1 through C_6 , and C_D , but that would make the coaxial lead capacitance notation very confusing.) The additional capacitances $C_A + C_B$ can be determined by connecting all eight coaxial leads to the device and then measuring the total capacitance-to-shield C_{T} at one of the points S, 1 through 6, or D. The total capacitanceto-shield is then $C_{\rm T} = C_{\rm S} + C_1 + C_2 + C_3 + C_4 + C_5 + C_4 + C_5 + C_$ $C_6 + C_D + C_A + C_B$, where $C_A = C_B$ if the device holder and the wire bonds are symmetrically arranged.

The equivalent circuit accounts for leakage currents between the QHE standard's inner conductors and the shields via resistances r_{K_A} and r_{K_B} located on either side of the QHE device. The sample probes should be constructed so these leakage resistances are very large. It would be safest to temporarily replace the device with shorts when measuring the total open-circuit leakage resistance r_{Lk} at point S, 1 through 6, or D. If all the contacts are clean, and if the leakage resistances are symmetrically distributed, then $r_{K_A} \approx r_{K_B} \approx 2r_{Lk}$ because they are connected in parallel within the circuit.





Volume 103, Number 6, November-December 1998 Journal of Research of the National Institute of Standards and Technology

599



Volume 103, Number 6, November-December 1998 Journal of Research of the National Institute of Standards and Technology

Fig. 2. Enlargement of Fig. 1 in the vicinity of the Potential coaxial port, plus the approximate numerical values of I_{C_5} , I_{C_1} , I_{C_1} , and V_{Pt} calculated from Eqs. (3a) to (3c) and Eq. (5b).

verify parts of this measurement chain only if the uncertainty is small enough. However, a correction having a small uncertainty could probably be made to $V_{\rm Pt}$ with careful measurements of the circuit elements in Fig. 1 and by using Eq. (5b) to calculate the correction factor.

A much more serious problem is the 1 % contribution to V_{Pt} in the out-of-phase signal. High precision ac bridges at NIST are not capable of providing accurate measurements of V_{Pt} if the out-of-phase signal is larger than 10⁻⁵ times the in-phase signal. Even at the 1×10^{-5} level, accurate measurements can be done only with great difficulty because care must be taken to correct for the in-phase (phase defect) contributions of the bridge components used to null the out-of-phase $V_{\rm Pt}$ signal. The in-phase (phase defect) signals of these components can be unintentionally added to the real, in-phase components of $V_{\rm Pt}$. Preliminary tests with a NIST bridge suggest that the in-phase (phase defect) signals due to the out-of-phase components could vary linearly with ω (as observed in many ac QHE experiments). These phase defect contributions are in addition to the normal second-order terms in the in-phase part of the signal that vary with ω^2 , such as the $\omega^2 C_S C_S R_H r_S$, $\omega^2 C_3 C_5 R_H r_3$, and $\omega^2 C_5 C_5 R_H r_5$ terms in Eq. (5b).

Equation (5b) predicts that the out-of-phase term $j[\omega C_5 R_{\rm H}]$ in the expression for Δ can be reduced by disconnecting coaxial cable 5 at position 5', where 5' is either located at the potential contact pad on the QHE device, or at an intermediate contact point on the sample holder. Several problems occur with this approach, however. There is a capacitance $C_{5'}$ between the QHE device and the shield. (This capacitance replaces capacitance C_5 in Fig. 1, while r_5 remains unchanged because the shield resistance is typically also about 1 Ω .) $C_{5'}$ is certainly much smaller than C_5 , so the out-of-phase term is much smaller, but does it reduce the out-ofphase term enough, and how can the capacitance $C_{5'}$ be measured since the coaxial lead is disconnected? Disconnecting cable 5 also precludes measuring the quantized Hall resistance with reversed magnetic field direction. All ac QHE experiments have found that these problems make the single-series wiring configuration unacceptable.

3.2 Double-Series Connections

Figure 3 shows an ac QHE standard with two doubleseries connections to the device. These connections numerical values of I_{C_5} , I_{C_3} , $I_{3'}$, and I_{C_1} calculated from Eqs. (8a) to (8d). A 1 % out-of-phase current still passes through each of the coaxial cable capacitances C_5 , C_3 , and C_1 in this example, and also through resistances $R_{\rm H}/2$ and r_3 of arm 3. The current I_{C_1} enters the Drive, goes to point Y, to point 3, through arm 3, through r_b , through arm 1, and then exits through C_1 . The current I_{C_5} enters the Drive, goes to point Y, to point D', through r_a , r_b , and r_c through arm 5, and then exits through C_5 . The current I_{C_3} enters the Drive, goes to point Y, to point 3, and then exits through C_3 .

 $V_{\rm H} = V_{\rm Pt}$ is obtained by summing the voltages between the inner conductors of the Detector coaxial port and the Potential coaxial port. Taking the path through arms 3 and 4 of Fig. 3 we find

$$V_{\rm Pt} = R_{\rm H}I_{\rm c} + R_{\rm H}I_4 + r_4I_4 + r_3I_{3'} - r_{\rm Pt}I_{\rm C_{\rm Pt}}, \qquad (9a)$$

or approximately

$$V_{\rm Pt} \approx \left\{ \left[1 + \frac{r_3 r_{\rm D}}{R_{\rm H} R_{\rm H}} + \frac{r_4 r_{\rm S}}{R_{\rm H} R_{\rm H}} \right] + j[\omega C_5 R_{\rm H}] \right\} R_{\rm H} I_{\rm Ot} \qquad (9b)$$

$$V_{\rm Pt} = [1 + \Delta] R_{\rm H}(i) I_{\rm Ot}$$
 (9c)

The numerical value of V_{Pt} in Eq. (9b) for this typical experimental example is

$$V_{\rm Pt} \approx \{1 + [2 \times 10^{-8}] + j[1 \times 10^{-2}]\}R_{\rm H}I_{\rm Ot}$$
. (10)

The real part of V_{Pt} is just the dc double-series prediction [13]. Unlike Eq. (5b) for the single-series solution, it has no significant second-order ω^2 terms. However, there is still the 1 % out-of-phase problem in the j[$\omega C_5 R_H$] term of Δ that would make it very difficult to perform accurate measurements, due to large phase defects of the components used to null the out-of-phase signal. Also, the same problems discussed at the end of Sec. 3.1 occur if coaxial leads 5 and 1 are disconnected at points 5' and 1'.

We have not yet analyzed a triple-series circuit with short coaxial leads added between points Y and 1, and between Z and 6. However, it seems clear that the large out-of-phase $j[\omega C_5 R_H]$ term would still appear in the V_{Pt} equation, giving the same measurement problems.

3.3 Quadruple-Series Connections

Figure 5 shows an ac QHE standard with two quadruple-series connections to the device. The circuit elements have the same typical values given by Eqs. (4a) to (4f) and Eqs. (7a) to (7b).

Six currents are of particular interest. Their approximate solutions are

$$I_{C_5} \approx \{-[\omega^2 C_5 C_B R_H r_S] + j[\omega C_5 R_H]\}I_{Ot}$$
(11a)

$$I_{5'} \approx \left\{ \left[\frac{r_{\rm c}}{R_{\rm H}} \right] + j \left[\frac{r_{\rm c}}{R_{\rm H}} \omega C_{\rm B} r_{\rm S} \right] \right\} I_{\rm Ot}$$
 (11b)

$$I_{C_3} \approx \{-[\omega^2 C_3 C_B R_H r_S] + j[\omega C_3 R_H]\}I_{Ot}$$
 (11c)

$$I_{3'} \approx \left\{ \left[\frac{r_{\rm b}}{R_{\rm H}} + \frac{r_{\rm D} r_{\rm 1}}{R_{\rm H} R_{\rm H}} \right] \right\}$$

$$+ j \left[\frac{r_{\rm D}}{R_{\rm H}} \frac{r_{\rm 1}}{R_{\rm H}} \omega C_{\rm A} R_{\rm H} + \frac{r_{\rm b}}{R_{\rm H}} \omega C_{\rm B} r_{\rm S} \right] \right\} I_{\rm Ot}$$
(11d)

$$I_{C_1} \approx \{-[\omega^2 C_1 C_B R_H r_S] + j[\omega C_1 R_H]\}I_{Ot}$$
 (11e)

$$I_{\mathrm{I}'} \approx \left\{ \left[\frac{r_{\mathrm{D}}}{R_{\mathrm{H}}} \right] + j \left[\frac{r_{\mathrm{D}}}{R_{\mathrm{H}}} \omega C_{\mathrm{A}} R_{\mathrm{H}} \right] \right\} I_{\mathrm{Ot}} \,. \tag{11f}$$

Figure 6 shows an enlargement of Fig. 5 in the vicinity of the Potential port, plus the approximate numerical values of I_{C_5} , $I_{5'}$, I_{C_3} , $I_{3'}$, and I_{C_1} calculated from Eqs. (11a) to (11e). A 1 % out-of-phase current still passes through each of the coaxial cable capacitances C_5 , C_3 , and C_1 in this example. However, these currents all enter the Drive, go to point Y, to point 3, and then out through the capacitors, bypassing the arms 5, 3, and 1.

 $V_{\rm H} = V_{\rm Pt}$ is obtained by summing the voltages between the inner conductors of the Detector coaxial port and the Potential coaxial port. Taking the path through arms 3 and 4 of Fig. 5 we find

$$V_{\rm Pt} = R_{\rm H}I_{\rm c} + (R_{\rm H} + r_4)I_4 + r_3I_{3'} - r_{\rm Pt}I_{\rm C_{\rm Pt}},$$
 (12a)

or approximately

$$V_{\rm Pt} \approx \left\{ \left[1 - \frac{(r_{\rm b} + r_{\rm c})}{R_{\rm H}} + \left(\frac{r_3 r_{\rm b}}{R_{\rm H} R_{\rm H}} + \frac{r_4 r_{\rm c}}{R_{\rm H} R_{\rm H}} \right) + \left(\frac{r_5 r_4 r_6}{R_{\rm H} R_{\rm H} R_{\rm H}} + \frac{r_{\rm D} r_1 r_3}{R_{\rm H} R_{\rm H} R_{\rm H}} \right) \right] + j[\omega C_{\rm B} r_{\rm S}] \right\} R_{\rm H} I_{\rm Ot} .$$
(12b)

The numerical value of V_{Pt} in Eq. (12b) for this typical experimental example is

$$V_{\text{Pt}} \approx \{1 - [1.999 \ 78 \times 10^{-7}] + j[1 \times 10^{-6}]\}R_{\text{H}}I_{\text{Ot}}.$$
 (13)

The real part of V_{Pi} appears to be an order of magnitude larger than the double-series prediction, but again with no significant second-order ω^2 terms. However, V_{Pi} actually measures the quantized Hall voltage V_H across the device *minus* the longitudinal voltage $V_x(2,6)$ along the device between points 2 and 6; i.e.,

$$V_{\rm Pt} = [1 + \Delta] [R_{\rm H}(i) - R_x(2,6)] I_{\rm Ot}, \qquad (14)$$

were first used by Delahaye [4] to overcome the above problems. All subsequent experiments have used doubleseries or triple-series connections.

Short coaxial leads outside the cryostat connect room temperature access points 3 and D at point Y. Two other short coaxial leads connect access points 4 and S at point Z. Short coaxial leads connect point Y with the Drive and Potential ports, and point Z with the Inner/ Outer and Detector ports. Typical values for the extra circuit elements are

$$r_{\rm Ot} = r_{\rm Dt} = r_{\rm Pt} = r_{\rm Dr} = 10^{-3} \,\Omega$$
 (7a)

$$C_{\text{Ot}} = C_{\text{Dt}} = C_{\text{Pt}} = C_{\text{Dr}} = 10^{-12} \,\text{F} \,.$$
 (7b)

The other circuit elements of Fig. 3 have the typical values listed in Eqs. (4a) to (4f).

Four currents are of particular interest. Their approximate solutions are

$$I_{C_5} \approx \{-[\omega^2 C_5 C_B R_H r_S - \omega^2 C_5 C_5 R_H r_5 + \omega^2 C_5 C_6 R_H r_5] + j[\omega C_5 R_H]\}I_{Ot}$$
(8a)

$$I_{C_3} \approx \{-[\omega^2 C_3 C_5 R_H R_H] + j[\omega C_3 R_H]\}I_{Ot}$$
 (8b)

$$I_{C_1} \approx \{-[-\omega^2 C_1 (C_1 - C_5) R_H R_H + \omega^2 C_1 C_6 R_H r_S]$$

+
$$j[\omega C_1 R_H] I_{Ot}$$
 (8c)

$$I_{3'} \approx \left\{ \left[\frac{r_{\rm D}}{R_{\rm H}} \right] \right\} I_{\rm Ot} + I_{\rm C_1} .$$
 (8d)

Figure 4 shows an enlargement of Fig. 3 in the vicinity of the Potential coaxial port, plus the approximate



Fig. 4. Enlargement of Fig. 3 in the vicinity of the Potential coaxial port, plus the approximate numerical values of I_{C_3} , I_{C_3} , $I_{3'}$, I_{C_1} , and V_{P_1} calculated from Eqs. (8a) to (8d) and Eq. (9b).



Fig. 6. Enlargement of Fig. 5 in the vicinity of the Potential coaxial port, plus the approximate numerical values of I_{C_5} , I_5' , I_{C_3} , $I_{3'}$, I_{C_1} , and V_{Pet} calculated from Eqs. (11a) to (11e) and Eq. (12b).

where $R_x(2,6)$ is the longitudinal resistance $(r_b + r_c)$. The quantity $[R_{\rm H}(i) - R_x(2,6)]$ has a correction factor Δ that is only 2.2×10^{-11} for this example.

There is only a 1×10^{-6} out-of-phase component in the V_{Pt} signal for the numerical example given in Eq. (13). The size of this $j[\omega C_B r_S] R_H I_{Ot}$ component is much smaller than the $j[\omega C_5 R_H] R_H I_{Ot}$ component in Eqs. (5b) and (9b) because the device-side of capacitor C_B is near zero potential, rather than near the quantum Hall potential of the device-side of capacitor C_5 . This small out-ofphase component can be handled, with care, by the NIST ac bridges; it can be further reduced by minimizing the capacitance-to-shield C_B . We suggest that future experiments surround the QHE device with a continuous conducting shield in the shape of a cylindrical pillbox to form reproducible capacitances C_A and C_B , and then support the device in the center of the largest possible pillbox to minimize these capacitances.

4. Summary

We are in the process of analyzing the effects of large capacitances-to-shields in sample probes on measurements of the quantized Hall resistance $R_{\rm H}$ using equivalent electrical circuits with capacitances and leakage resistances to the sample probe shields, longitudinal resistances within the quantized Hall effect devices, and multiple connections to the devices. The exact algebraic solutions for the $R_{\rm H}$ values in these circuits reveal large 90° out-of-phase contributions to the quantized Hall voltages $V_{\rm H}$ that would make it very difficult to make accurate measurements with high precision ac bridges for single-series, double-series, and triple-series connections to the QHE devices.

We predict, however, that quadruple-series connections to the devices yield only small out-of-phase contributions to $V_{\rm H}$ that should allow accurate determinations



Fig. 5. An equivalent electrical circuit representation of a QHE resistance standard with two quadruple-series connections while the standard is being measured in an ac ratio bridge using four-terminal-pair measurement techniques. See Sec. 3.3 for the circuit analysis.

602

with small uncertainties of the quantity $[R_{\rm H} - R_x(2,6)]$, where $R_x(2,6)$ is the longitudinal resistance along the device. It is unfortunate that the quantity $[R_{\rm H} - R_x(2,6)]$ is obtained, rather than $R_{\rm H}$, but the fact that all eight coaxial leads remain connected to the device contact pads means that the values of all the circuit elements in the equivalent circuit representation of the QHE standard could be determined. It also means that both ac and dc measurements of $V_{\rm H}$ and V_x could be performed on the same cool-down, for both magnetic field directions.

Acknowledgments

We thank Drs. Kevin C. Lee, Edwin R. Williams, Alan F. Clark, and Martin Misakian of the Electricity Division, and Dr. Mark D. Stiles of the Electron and Optical Physics Division of NIST for their discussions and comments.

5. References

- K. von Klitzing, G. Dorda, and M. Pepper, New Method for High-Accuracy Determination of the Fine-Structure Constant Based on Quantized Hall Resistance, Phys. Rev. Lett. 45, 494– 497 (1980).
- [2] The Quantum Hall Effect, R. E. Prange and S. M. Girvin, eds., Springer-Verlag, New York (1987) pp. 1–419.
- [3] The Integral and Fractional Quantum Hall Effects, C. T. Van Degrift, M. E. Cage, and S. M. Girvin, eds., American Association of Physics Teachers, College Park, Maryland (1991) pp. 1–116.
- [4] F. Delahaye, Accurate AC Measurements of the Quantized Hall Resistance from 1 Hz to 1.6 kHz, Metrologia 31, 367–373 (1995).
- [5] B. M. Wood, A. D. Inglis, and M. Cote, Evaluation of the AC Quantized Hall Resistance, IEEE Trans. Instrum. Meas. 46, 269–272 (1997).
- [6] J. Melcher, P. Warnecke, and R. Hanke, Comparison of Precision AC and DC Measurements with the Quantized Hall Resistance, IEEE Trans. Instrum. Meas. 42, 292–294 (1993).
- [7] A. Hartland, B. P. Kibble, P. J. Rodgers, and J. Bohacek, AC Measurements of the Quantized Hall Resistance, IEEE Trans. Instrum. Meas. 44, 245–248 (1995).
- [8] J. Bohacek, P. Svoboda, and P. Vasek, AC QHE-Based Calibration of Resistance Standards, IEEE Trans. Instrum. Meas. 46, 273–275 (1997).
- [9] A. Hartland, B. P. Kibble, and S. W. Chua, AC Quantized Hall Resistance Experiments at the National Physical Laboratory, BEMC '97 Conference Digest, National Physical Laboratory, Teddington, U. K. (1997) pp. 18-1 to 18-4.
- [10] F. Piquemal, G. Trapon, and G. Geneves, AC Measurements of the Minimum Longitudinal Resistance of a QHE Sample from 10 Hz to 10 kHz, IEEE Trans. Instrum. Meas. 45, 918–922 (1996).
- [11] F. Cabiati, L. Callegaro, C. Cassiago, V. D'Elia, and G. M. Reedtz, Effect of the Frequency on the Longitudinal Resistance of a GaAs-AlGaAs Heterostructure, CCE/97-32, 1-5 (1997).

- [12] M. E. Cage and A. Jeffery, Intrinsic Capacitances and Inductances of Quantum Hall Effect Devices, J. Res. Natl. Inst. Stand. Technol. 101, 733–744 (1996).
- [13] M. E. Cage, A. Jeffery, R. E. Elmquist, and K. C. Lee, Calculating the Effects of Longitudinal Resistance in Multi-Series-Connected Quantum Hall Effect Devices, J. Res. Natl. Inst. Stand. Technol. 103, 561–592 (1998).
- [14] R. D. Cutkosky, Four-Terminal-Pair Networks as Precision Admittance and Impedance Standards, Communication Electronics 70, 19–22 (1964).
- [15] R. D. Cutkosky, Techniques for Comparing Four-Terminal-Pair Admittance Standards, J. Res. Natl. Bur. Stand. (U.S.) 74C, 63–78 (1970).
- [16] D. N. Homan, Applications of Coaxial Chokes to AC Bridge Circuits, J. Res. Natl. Bureau Stand. (U.S.) 72C, 161–165 (1968).
- [17] M. E. Cage, A. Jeffery, and R. E. Elmquist, Suggested Triple-Series Connection Measurement Tests of the Quantized Hall Resistance and the AC Longitudinal Resistance, IEEE CPEM Digest 98CH36254, 341–342 (1998).
- [18] B. W. Ricketts and P. C. Kemeny, Quantum Hall Effect Devices as Circuit Elements, J. Phys. D: Applied Phys. 21, 483–487 (1988).
- [19] F. Delahaye, Series and Parallel Connection of Multiterminal Quantum Hall Effect Devices, J. Appl. Phys. 73, 7915–7920 (1993).
- [20] A. Jeffery, R. E. Elmquist, and M. E. Cage, Precision Tests of a Quantum Hall Effect Device DC Equivalent Circuit using Double-Series and Triple-Series Connections, J. Res. Natl. Inst. Stand. Technol. 100, 677–685 (1995).
- [21] A. Jeffery, R. E. Elmquist, L. H. Lee, J. Q. Shields, and R. F. Dziuba, NIST Comparison of the Quantized Hall Resistance and the Realization of the SI OHM Through the Calculable Capacitor, IEEE Trans. Instrum. Meas. 46, 264–268 (1997).

About the authors: M. E. Cage and A. Jeffery are physicists in the Electricity Division, Electronics and Electrical Engineering Laboratory, National Institute of Standards and Technology. The National Institute of Standards and Technology is an agency of the Technology Administration, U.S. Department of Commerce.

fourth of Sectoration at some instants for sectors and Technology

Although a strain a strain and the second strain and the second

We walk the one of

(2) A state of the set of the

Contraction and the

- Doute and Y and Y and A long solution of the Growth of the Goldsmann Math. Rev. Letters of the State of the Action and the Quantum Math. Rev. Letters of the Park at the Action of the Action of the State of the State of the State of the State of the Action of the State of the
- [1] Margaran C. Kither and K. Karana and K. Kither and A. Karana and A. Ka Karana and A. Karana and Karana and A. Karana and

- (a) A. C. Salakov, "Color: J. K. A. S. M. Market, J. Concentration and the state of the state of the second state of the state of the state state of the state state of the state state of the state
- the sequence is a second restriction of the second se
- and a second second firms of an and a second se

- [17] M. E. Cape and A. Jefferg, Journal Capacityneau and Inductunes of Contemporal Infl Effort Devices, J. Ken. Mat. Soc. Stand. Technol. 161, 202–244 (1996).
- [12] M. E. Crey, a. J. Tary, R. E. Bragnin, and K. C. Lee, Calendy, log pre-Effects, of Longitudinal Realitrates in Multi-Series Conneural Quantum 1920 Effect Devices, J. Fee, Nucl. Social Total of Adv. 392 (1993).
- (c) R. D. Consula. Four-Tepring Park Introduction Protocom Aulumination and Interactional Standards. Communication Electronics 70, 49–42. (1995).
- [15] R. Y. Carlowky, "Semi-upped int Comparing rout: Internation-time Monitories Symposius, 3, Phys. Nucl. Soc. Soc. Soc. Math. (U.S.) 34C, 805-26, 112031.
- [10] D. E. Human, Appletoilees of Special Checks to AC Boups Checkson, J. Fox, Soli, Barran Sumd. (M.S.) 62C, 161–163 (1996).
- [12] M. L. Cage, A. Scher, and E. E. Ehnquist, Suggested Dispe-Series Configuration (Communical Tests of the Quantized Hull Society and the AC Longitudical Regionsney, IEEE CFEM Dispers SuCIDENSE, 2011–202 (1998).
- (1) K. W. Marker and P. S. Arrive D. Applied Phys. 21, 485–484 (1994).
- [17] F. Elemand, Science and Control of Control 101 Automatication Computer Stati Minute Science, J. Appl. Phys. 73, 7915–7946 (1997).
- (2) A. Delling, R. L. Elinguist, and M. E. Carre, Produces Tests of S. Q. Propiet Test: Refer. Device DC Equivalent Croud using London-Leffercand Trate Script Connections, J. Res. Mat. Ref. Long. Science, 801, 877–833 (1985).
- [11] A. Leffery, A. E. Sumaplat, L. H. Lan, J. Q. Sevenal, and S. P. Brunew, 1982. Comparison of the Quantimed Hall Represence and the Residuation of the St Office Through the Computable Company of HERE Threat. Internal. Mass., 96, 264–266 (1997).

About the authors: M. E. Gage and A. Jeffery are pleatered in the Electricity Division, Electronics and Chartered Environmentary, National Institute of a montaneous Technology, The National Institute of Sumperclosured Technology is an agency of the Technolan algorithms of Danarmient of Commerce.