

NIST Special Publication 260-226

Certification of Standard Reference Material® 114r Portland Cement Fineness

Paul Stutzman
Eduarda Votri
Blaza Toman

This publication is available free of charge from:
<https://doi.org/10.6028/NIST.SP.260-226>



NIST Special Publication 260-226

Certification of Standard Reference Material® 114r Portland Cement Fineness

Paul Stutzman
Eduarda Votri
*Materials and Structural Systems Division
Engineering Laboratory*

Blaza Toman
*Statistical Engineering Division
Information Technology Laboratory*

This publication is available free of charge from:
<https://doi.org/10.6028/NIST.SP.260-226>

June 2022



U.S. Department of Commerce
Gina M. Raimondo Secretary

National Institute of Standards and Technology
Laurie E. Locascio, NIST Director and Under Secretary of Commerce for Standards and Technology

Certain commercial entities, equipment, or materials may be identified in this document in order to describe an experimental procedure or concept adequately. Such identification is not intended to imply recommendation or endorsement by the National Institute of Standards and Technology, nor is it intended to imply that the entities, materials, or equipment are necessarily the best available for the purpose.

National Institute of Standards and Technology Special Publication 260-226
Natl. Inst. Stand. Technol. Spec. Publ. 260-226, 40 pages (June 2021)
CODEN: NSPUE2

This publication is available free of charge from:
<https://doi.org/10.6028/NIST.SP.260-226>

Abstract

The SRM 114 series for portland cement fineness has been provided by NIST since 1934 for use in calibrating instruments for measuring fineness characteristics of portland cement and other powdered construction materials. SRM 114r is the latest of this series providing certified values for cement fineness by Blaine apparatus, + 45 μm (No. 325) Sieve, and informational values for particle size distribution by light scattering. ASTM C01.25 Committee on Fineness provided a range of desired materials properties, and the assistance of the Cement and Concrete Reference Lab provided several candidate plants with materials conforming to the Committee request. Approximately 1000 kg of cement was procured, homogenized, and packaged. An inter-laboratory study was initiated to develop the data set used to establish certified values. The certified values represent the consensus means and uncertainties for the Blaine and + 45 μm residue. Informational values on particle size distribution are provided as the mean of the combined wet and dry laser diffraction data.

Key words

Blaine Fineness, Cement, Particle Size Distribution, Specification, SRM.

Table of Contents

List of Figures.....	iii
1. Introduction	1
2. Characteristic of a SRM Candidate.....	1
2.1. Materials	1
2.1.1. Density.....	1
2.2. Inter-Laboratory Study	2
2.3. Sample Randomization.....	2
2.3.1. Blaine Fineness, ASTM C204.....	2
2.3.2. Particle Size Distribution by Laser Diffraction, AASHTO T353	3
2.3.3. Percent Retained on #325 Mesh Screen, ASTM C430	3
2.3.4. Cement Wet Sieving Procedure	4
3. Blaine Test Results.....	10
3.1. The Data	11
3.2. Analysis	13
4. Particle Size Distribution	16
4.1. The data	16
4.2. Statistical Analysis	17
5. Acknowledgments.....	19
References.....	20
Appendix A: Wet Sieve Test Results.....	21
Appendix B. Blaine fineness test results.....	22
Appendix C: OpenBUGS code for the sieve residue analysis	23
Appendix D. Consensus Means and Uncertainty Calculations	25
Appendix E. Dataplot code for consensus means.....	29
Appendix F Blaine Fineness Test Results formatted for Dataplot.....	33

List of Tables

Table 1. Density Measurements for 114r and SRM 46h using helium pycnometry.....	2
Table 2. Cement Sieve Hole and Wire Diameter Size and Uncertainties following ASTM 11 [2].....	3
Table 3. Sieve residue measurements for the three calibrated sieves	6
Table 4. Parameter estimates with uncertainty.	8
Table 5. Blaine measurements in m^2/kg	11
Table 6. Estimated box means with uncertainties for SRM 114r.	14
Table 7. Cumulative % by bin size in microns for wet and dry measurements.....	16
Table 8. Consensus values with uncertainty for the cumulative particle size distribution....	18

List of Figures

Figure 1. Percent retained on 38 um, 45 um, and 53 um micron sieves do not appear to show any trend based upon sample packing sequence.....	5
Figure 2. The 38 μm sieve shows an increase in percent retained as testing progressed resulting from trapped particles removed upon. This trend was not found for the 45 μm and 53 μm sieves.	5
Figure 3. 95% prediction intervals with observed data for the 38 μm sieve.	8
Figure 4. 95% prediction intervals with observed data for the 45 μm sieve.	9
Figure 5. 95% prediction intervals with observed data for the 53 μm sieve.	9
Figure 6. Residuals with 95% uncertainty intervals.	10
Figure 7. Blaine measurements in m^2/kg by lab.	12
Figure 8. Estimated box means with 95% uncertainty intervals, reference value with 95% uncertainty in green, 95% prediction interval in red.....	15
Figure 9. Bar graph of the cumulative frequencies given in Table 7 separated by method....	17
Figure 10. Estimated cumulative particle size distribution with 95% uncertainty bounds....	19

1. Introduction

The SRM 114 series for portland cement fineness has been provided by NIST since 1934 for calibrating instruments for measuring fineness characteristics of portland cement and other powdered construction materials [1]. Certified values are provided for cement fineness by ASTM C204 [2] Standard Test Methods for Fineness of Hydraulic Cement by Air-Permeability Apparatus (commonly called the Blaine method), ASTM C430, Standard Test Method for Fineness of Hydraulic Cement by the 45- μm (No. 325) Sieve [3], and informational values for particle size distribution by light scattering following AASHTO T 353-14 [4].

The Blaine apparatus measures total surface area relative to the SRM by drawing a fixed quantity of air through a bed of cement having a specific porosity where the pore volume and size is related to the particle size of the cement. The Blaine fineness is used as a material test requirement or an optional requested material test. ASTM C430 measures the material residue greater than 45 μm using a No. 325 mesh sieve (45 μm opening). SRM 114 is used to calibrate the sieve by comparing the amount retained on the lab sieve to the certified value and is also part of a material specification or an optional reported material property. The particle size distribution is a newer measure of cement fineness and is standardized in the AASHTO T 353 test method [4].

2. Characteristic of a SRM Candidate

A poll was conducted with members of ASTM C01.25 on Cement Fineness on the materials characteristics for the new 114 cement. The Committee consensus was for a Blaine fineness around 400 m^2/kg and +45 μm residue between 4 % and 6 %. Reviewing the CCRL proficiency test data (CCRL.US) shows that a small number of cements have +the desired characteristics as modern grinding systems are more effective at producing a finer, more uniform particle size. The larger retention is helpful to calibrate laboratory #325 sieves for analysis of cement and other materials used with cements, such as fly ash. Low retention values make sieve calibration more difficult.

2.1. Materials

Once prospective candidate material sources were identified, cement manufacturers were contacted to confirm current production characteristics (Blaine, +45 μm) and a selection was made to procure about 1100 kg of material. The cement was packed in plastic barrels lined with heavy plastic bags, sealed and shipped to the Cement and Concrete Reference Laboratory in Frederick Maryland for homogenization using their large V-blender, capable of mixing the entire 1100 kg in one run. The homogenized material was re-packaged in the barrels and staff from the Standard Reference Materials Program (SRMP) picked up the barrels for delivery to NIST in Gaithersburg for packaging. Packaging in vials of about 5g each, with 20 vials per unit to remain consistent with past SRM 114 production and material use needs. Vials were stored in boxes of 200 as packaging proceeded.

2.1.1. Density

Cement density (Table 1) was measured using helium pycnometry [5]. This method was demonstrated to provide density measurements for SRM 114q consistent with the Le Chatelier method used in ASTM C188 [5]. For some of the CCRL cements tested, He

Pycnometry provided values slightly greater than the Le Chatelier method, likely due to the more complete permeation of the helium into fine fractures and pores within the cement particles. These density values are provided as informational values. Section 13.2 of [3] requires a density difference of no greater than 0.06 g/cm³

Table 1. Density Measurements for 114r and SRM 46h using helium pycnometry

Trial	114r	46h
1	3.1422	3.2118
2	3.1474	3.1971
3	3.1435	3.1975
4	3.1572	
Mean	3.1476	3.2021
σ	0.0068	0.0084

2.2. Inter-Laboratory Study

Participants were recruited through ASTM Committee C01.25, Fineness, and membership of the Portland Cement Association to provide Blaine (ASTM C204) and particle size distribution data (AASHTO T353) from samples of SRM 46h and the candidate SRM 114. This practice of using an inter-laboratory study to establish the data set from which the consensus values provide certified and reference values is consistent with previous editions of the SRM 114 series. The tests and their significance will be briefly described, and references provided which provide greater detail on the procedures and application of the SRM for their calibration procedures.

2.3. Sample Randomization

Once the list of participants and tests were established a randomized assignment of samples, drawn from storage trays was established and recorded. The Participants and Test Methods (Blaine, Laser PSD, +45 µm residue) were matched to the list as the participant list developed. For each test procedure results could be evaluated by lab and by box, representing the packing sequence or by testing sequence to assess for trends.

2.3.1. Blaine Fineness, ASTM C204

The Blaine fineness test used in ASTM C204 is based upon the relationship between the air permeability of a packed bed of cement powder and its fineness and, consequently, its surface area. Cement fineness influences the rheological and strength gain properties and are material specification criteria in ASTM C150 [6] and an optional reporting element in Specifications C595 [7] and C1157 [8]. The current cement fineness SRM 46h served to calibrate the Blaine apparatus so the new SRM 114 measurements would be referenced against a recent measurement of a current fineness SRM. This approach ensured each lab was

consistent in their calibrations if the test procedure was properly followed. Section 13.2 of [3] requires a Blaine fineness difference of no greater than $200 \text{ m}^2/\text{kg}$.

2.3.2. Particle Size Distribution by Laser Diffraction, AASHTO T353

A particle size distribution measurement provides a more complete measurement of a cement, but the test results are not yet part of any cement specifications. AASHTO T353 involves the suspension of cement particles in either air or isopropanol, measuring the scattering of laser light to calculate a particle size distribution. Each participant was instructed to follow the AASHTO procedure and record their results in a spreadsheet provided with the samples. The spread sheet compiled the laser diffraction data into bins consistent with previous SRM 114 samples. Data from SRM 46h were also collected to see if a lab's test protocol produced results consistent with the SRM values and may serve as a means of identifying outliers.

2.3.3. Percent Retained on #325 Mesh Screen, ASTM C430

The residue of cement on a calibrated #325 ($45\text{-}\mu\text{m}$) sieve is a specification property for cement and supplementary cementitious materials. Since a direct certification of sieve openings is impractical for production-scale work, sieves are calibrated by using a certified reference material, SRM 114. A sieve correction factor is calculated after measuring SRM 114 on the selected sieve and calculation of a correction factor.

Three sieves with nominal openings of $38 \mu\text{m}$, $45 \mu\text{m}$, and $54 \mu\text{m}$ were directly calibrated for use in reference material certification. Interpolation was then used to obtain the value at $45 \mu\text{m}$. These sieves consist of three cylindrical bodies about 50 mm in diameter with an individually calibrated screen held in the base. The test involves wet-sieving a 1.000 g sample of cement through each of the three sieves and measurement of the mass of the residue on each sieve.

The $+45\text{-}\mu\text{m}$ data were measured at NIST using a set of three cement wet sieves with calibrated screens following ASTM E11-20 [9] with the certified mean opening value and uncertainty listed in Table 2. The procedure for sieving is provided in Section 2.3.4 below, with results presented graphically in Figures 1 and 2 and test results provided in Appendix A.

Table 2. Cement Sieve Hole and Wire Diameter Size and Uncertainties following ASTM 11 [2].

No. 400	Diameter	Tolerance	Warp Diameter	Shoot Diameter	Uncertainty
Hole	0.038 mm	0.003 mm	0.03834 mm	0.03812 mm	0.00059 mm
Wire	0.030 mm	0.0045 mm	0.2898 mm	0.02759 mm	0.00059 mm
No. 325	Diameter	Tolerance	Warp Diameter	Shoot Diameter	Uncertainty
Hole	0.045 mm	0.003 mm	0.04496 mm	0.04347 mm	0.00059 mm
Wire	0.032 mm	0.0048 mm	0.03198 mm	0.03153 mm	0.00059 mm

No. 270	Diameter	Tolerance	Warp Diameter	Shoot Diameter	Uncertainty
Hole	0.053 mm	0.004 mm	0.05148 mm	0.05412 mm	0.00059 mm
Wire	0.036 mm	0.0054 mm	0.03650 mm	0.03447 mm	0.00059 mm

2.3.4. Cement Wet Sieving Procedure

Materials & Tools:

- 1) Cylindrical sieves made of chromium-plated brass
- 2) Type 304 stainless steel woven-type cloth mounted to the sieve base without distortion or wrinkling
- 3) ASTM C430-compliant spray nozzle and pressure gauge
- 4) Low-temperature oven (105 °C) for drying; leather gloves for heat protection when removing sieve from oven
- 5) Nitrile or latex gloves for hand protection while sieving
- 6) Safety glasses with side shields for eye protection

Method:

- 1) Place a 1.000 g sample of the cement in the clean, dry sieve
- 2) Wet the sample thoroughly with a gentle stream of water
- 3) Wash for 1.0 min, moving the sieve with a circular motion in a horizontal plane at the rate of one motion per second in the spray. The bottom of the spray nozzle should extend below the top of the sieve frame about 12 mm (0.5 in.)
- 4) Immediately upon removing the sieve from the spray, rinse once with about 50 cm³ of distilled water, using caution not to lose any of the residue
- 5) Displace the water using about 25 ml of ethanol
- 6) Dry the sieve and residue in an oven or over a hot plate supporting the sieve in such a manner that air may pass freely beneath it
- 7) Cool the sieve; then carefully brush the residue from the sieve and weigh on an analytical balance capable of reproducing results within 0.0005 g. Alternatively, weigh the sieve and residue after drying, subtract the mass of the sieve to calculate residue mass
- 8) *Acceptable Cleaning Procedures*— Clean sieve using a stream of water in the opposite direction followed by a rinse using ethanol to displace the water and dry in low-temperature oven.




Figure 1. Percent retained on 38 μm , 45 μm , and 53 μm micron sieves do not appear to show any trend based upon sample packing sequence.




Figure 2. The 38 μm sieve shows an increase in percent retained as testing progressed resulting from trapped particles removed upon. This trend was not found for the 45 μm and 53 μm sieves.

To determine the mass fraction of cement that would be retained by a 45 µm sieve, determinations of sieve residue were obtained using calibrated sieves of three sizes, 38.0 µm, 45.0 µm, and 53.0 µm. The measurement data based on two vials of 36 different boxes is given in Table 3.

Table 3. Sieve residue measurements for the three calibrated sieves

Box	Vial	Residue for 38 µm sieve	Residue for 45 µm sieve	Residue for 53 µm sieve	Sequence position
153	1	9.699	5.89	3.397	1
153	2	9.812	5.979	3.43	2
159	1	10.268	6.178	3.55	3
159	2	10.324	6.142	3.586	4
125	1	9.951	6.106	3.479	5
125	2	9.707	6.236	3.387	6
85	1	10.054	6.077	3.43	7
85	2	10.278	6.289	3.479	8
156	1	10.171	6.177	3.422	9
156	2	10.14	6.08	3.33	10
190	1	9.747	6.106	3.475	11
190	2	10.339	6.077	3.571	12
194	1	10.317	6.369	3.616	13
194	2	10.387	6.392	3.661	14
138	1	10.306	6.38	3.625	15
138	2	9.74	6.437	3.668	16
57	1	10.051	6.481	3.592	17
57	2	10.264	6.447	3.693	18
108	1	10.236	6.578	3.509	19
108	2	10.473	6.595	3.788	20
98	1	10.483	6.555	3.644	21
98	2	10.329	6.565	3.678	22
168	1	10.483	NA	3.795	23
168	2	10.091	6.581	3.69	24
143	1	10.354	NA	3.787	25
143	2	10.417	6.53	3.648	26
43	1	10.581	6.818	3.774	27
43	2	10.486	6.885	3.78	28
149	1	10.582	6.872	3.727	29
149	2	10.97	6.866	3.645	30
92	1	10.926	6.813	3.789	31
92	2	10.897	6.83	3.762	32
3	1	10.764	6.816	3.81	33
3	2	10.789	6.912	3.784	34

129	1	10.911	6.961	3.877	35
129	2	10.98	6.906	3.824	36

The sieve sizes were measured with uncertainty of 0.59 μm . Also, it is known that as a sieve is being used over time, small amount of the residue is retained, thus increasing the residue measurements over time. The data in Table 3 is arranged in time sequence. The following statistical model can capture the effects of sieve size on residue in terms of a quadratic function, also the effects of sequence and the uncertainty in the sieve size.

The measurements are y_{38i} , y_{45i} , and y_{53i} .

$$y_{38i} \sim N(\mu_{38i}, \sigma_{38}), i = 1, \dots, 36,$$

$$y_{45i} \sim N(\mu_{45i}, \sigma_{45}), i = 1, \dots, 34,$$

$$y_{53i} \sim N(\mu_{53i}, \sigma_{53}), i = 1, \dots, 36.$$

The means for the i^{th} set of measurements were modeled using a quadratic equation in terms of the sieve size, and the added residue due to sequence, as

$$\mu_{38i} = \beta_1 + \beta_2 (\mu.s38 - \bar{\mu.s}) + \beta_3 (\mu.s38 - \bar{\mu.s})^2 + \beta_4(i - 1), i = 1, \dots, 36,$$

$$\mu_{45i} = \beta_1 + \beta_2 (\mu.s45 - \bar{\mu.s}) + \beta_3 (\mu.s45 - \bar{\mu.s})^2 + \beta_5(i - 1), i = 1, \dots, 34,$$

$$\mu_{53i} = \beta_1 + \beta_2 (\mu.s53 - \bar{\mu.s}) + \beta_3 (\mu.s53 - \bar{\mu.s})^2 + \beta_6(i - 1), i = 1, \dots, 36.$$

The sieve sizes were measured with uncertainty and were modeled as:

$$\mu.s38 \sim N(38.0, 0.59^2),$$

$$\mu.s45 \sim N(45.0, 0.59^2),$$

$$\mu.s53 \sim N(53.0, 0.59^2),$$

$$\bar{\mu.s} = \frac{1}{3}(\mu.s38 + \mu.s45 + \mu.s53).$$

This statistical model was fitted using Bayesian MCMC (code is in the Appendix) to obtain parameter estimates of the regression coefficients with uncertainties. Prior distributions were as non-informative as possible, using half Cauchy for the variances σ_{38}^2 , σ_{45}^2 , and σ_{53}^2 , and uniform distributions for the regression parameters. The results for the parameter estimates, which are given in Table 4, are not sensitive to the prior distribution assumptions.

Table 4. Parameter estimates with uncertainty.

parameter	Posterior mean	Posterior std	Lower 95%	median	Upper 95 %
β_1	5.774	0.341	5.056	5.782	6.433
β_2	-0.443	0.033	-0.519	-0.440	-0.388
β_3	0.017	0.008	0.004	0.017	0.035
β_4	0.028	0.003	0.022	0.028	0.035
β_5	0.031	0.002	0.027	0.031	0.034
β_6	0.012	0.001	0.009	0.012	0.015
σ_{38}	0.213	0.026	0.169	0.211	0.272
σ_{45}	0.097	0.013	0.076	0.096	0.126
σ_{53}	0.080	0.010	0.063	0.079	0.104
μ_{s38}	38.04	0.57	36.92	38.04	39.14
μ_{s45}	45.02	0.53	43.94	45.04	45.95
μ_{s53}	52.94	0.59	51.74	52.94	54.10

The fit of the statistical model was examined using prediction intervals and residuals. The following three plots (Figures 3,4,5) show the measured data values together with the 95% prediction intervals for the three sieves. Except for one or two points the data is included in the prediction intervals.




Figure 3. 95% prediction intervals with observed data for the 38 μm sieve.




Figure 4. 95% prediction intervals with observed data for the 45 μm sieve.




Figure 5. 95% prediction intervals with observed data for the 53 μm sieve.

The residuals from the statistical model with their 95% uncertainty interval are plotted in Figure 6. The residuals appear to be random, not systematic, and the intervals are wider for the smaller sieve sizes as could be anticipated.




Figure 6. Residuals with 95% uncertainty intervals.

The estimate of the residue fraction for 45 μm sieve can be estimated as

$$\mu_{45} = \beta_1 + \beta_2 (45 - \bar{\mu}_{\cdot S}) + \beta_3 (45 - \bar{\mu}_{\cdot S})^2.$$

This is equal to 5.97 % with standard uncertainty of 0.24 %, and with 95% uncertainty interval of (5.44%, 6.39%).

3. Blaine Test Results

Each lab received a single vial of SRM 46h and two vials of SRM 114r for analysis. They were instructed to run the 46h as a calibration and to provide air flow times; the calculated Blaine surface area was performed at NIST using their air flow times. Using SRM 46h for

calibration provided for consistency across labs. For the candidate SRM 114r, each lab was requested to run their analysis in duplicate, with the test result being the mean of the four measurements. This requires the sample be removed and broken up to form a powder. ASTM C 204 cautions against comparing cements that are too different regards density and fineness. As mentioned previously, SRM 46h and candidate 114r are just within the limits for density and fall well within the limits for fineness (+- 200) so these cements may be compared.

3.1. The Data

The Blaine measurements were obtained as described in ASTM C 204 [2]. There were 42 different boxes, each sampled 2 times and each sample measured twice. Table 5 gives the measurements by box and a plot of test results is shown in Figure 7.

Table 5. Blaine measurements in m²/kg.

Sample Box	Blaine 114r #1		Blaine 114r #2	
replicate	1	2	1	2
8	387.6	385.5	384.8	383.1
15	395.4	386.8	386.8	391.1
17	395.5	395.4	389.2	389.0
21	390.1	391.9	390.3	393.3
22	398.0	396.8	394.2	393.0
26	391.5	390.7	391.5	391.1
35	366.6	367.3	370.1	371.1
36	378.8	381.1	381.1	375.1
40	397.3	397.3	397.3	395.2
44	400.4	411.9	417.5	417.5
49	389.2	390.7	394.3	395.9
58	394.9	390.9	395.4	391.4
61	395.9	392.9	391.5	392.1
70	392.3	392.7	392.7	391.3
73	392.7	397.1	393.5	395.6
78	374.8	363.5	405.2	415.3
87	398.5	399.0	392.4	394.3
88	371.8	374.8	368.8	375.4
90	398.2	395.1	395.1	399.2
101	394.1	389.6	394.5	390.0
107	398.3	395.7	394.6	396.1
110	396.4	398.5	396.3	399.0
115	391.4	392.4	391.2	392.0
122	391.7	393.2	389.4	389.5
124	386.6	386.4	387.5	387.6
131	385.3	385.7	395.7	395.8

145	402.4	404.6	400.1	402.4
148	396.8	396.2	390.9	391.7
152	396.9	396.9	394.9	395.9
157	388.9	390.1	390.6	391.8
160	396.4	394.6	395.6	396.1
166	400.4	398.3	398.3	398.3
171	398.6	396.9	396.8	395.4
173	396.2	397.4	397.6	399.0
174	391.1	390.7	391.0	390.6
175	393.6	396.0	388.9	391.3
176	389.6	389.6	387.3	389.6
177	403.1	398.8	396.4	396.9
178	394.3	394.9	393.6	394.9
196	377.4	377.2	377.7	377.2
200	394.8	396.8	395.8	396.8




Figure 7. Blaine measurements in m^2/kg by lab.

3.2. Analysis

Figure 7 shows that the data may be somewhat inhomogeneous. A formal test of between box heterogeneity called Cochran's Q is provided as part of the NIST Decision Tree [11]. It indicates heterogeneity with p-value of < 0.001 . A further test of symmetry (passed) and of Gaussian distribution (rejected with p-value <0.001), both included in the NIST Decision Tree, indicate that a random effects model which includes between box heterogeneity using the Laplace probability distribution is the best choice to obtain a reference value and its uncertainty. The statistical model is as follows:

$y_{ijk} \sim N(\gamma_i, \sigma_i^2)$, is the measurement where i denotes box ($i = 1, \dots, 42$), j denotes sample (1, 2), k denotes replicate (1, 2),

$\gamma_i \sim Laplace(\mu, \tau)$, γ_i is the mean value of box i , μ is the consensus value, τ is the scale parameter of the Laplace distribution.

The estimation of the parameters of this statistical model is accomplished using Bayesian MCMC implemented in OpenBUGS [12, 13]. Table 6 gives the estimates of the box means γ_i with their standard uncertainties and 95% uncertainty intervals.

The estimated μ (reference value) is $393.2 \text{ m}^2/\text{kg}$ with standard uncertainty of $0.9 \text{ m}^2/\text{kg}$ and 95% uncertainty interval of $(391.5, 394.9)$. It may also be of interest, since there is heterogeneity among the boxes, to include a prediction interval for the value of a randomly selected box. This is $(383.3, 403.3)$.

Figure 8 shows the estimated box means with 95% uncertainty intervals, the consensus (reference) value with its 95% uncertainty interval in green, and the 95% prediction interval in red.

Table 6. Estimated box means with uncertainties for SRM 114r.

number	box	Mean m ² /kg	Standard uncertainty	l95	u95
1	88	373.5	1.6	371.0	377.0
2	70	392.3	0.3	391.7	392.8
3	73	394.7	0.9	392.9	396.4
4	173	397.5	0.5	396.4	398.6
5	148	393.8	1.3	391.2	396.4
6	171	396.9	0.6	395.7	398.0
7	110	397.4	1.1	395.1	399.6
8	122	391.0	0.8	389.4	392.5
9	58	393.1	1.3	390.5	395.7
10	15	390.2	1.7	386.9	393.6
11	200	395.9	1.2	393.5	398.2
12	176	389.0	0.5	388.0	390.0
13	21	391.4	0.8	389.8	393.0
14	124	387.0	0.3	386.4	387.7
15	131	390.8	2.3	386.2	395.4
16	61	393.1	1.0	391.1	395.0
17	107	396.1	1.0	394.1	398.0
18	8	385.4	0.9	383.7	387.1
19	26	391.2	0.4	390.5	391.9
20	177	398.6	1.3	395.8	401.0
21	196	377.4	0.3	376.8	378.0
22	40	396.8	0.5	395.7	397.8
23	90	396.8	1.1	394.6	398.9
24	162	396.1	0.5	395.2	397.1
25	35	369.3	1.2	367.4	371.7
26	160	395.7	0.4	394.9	396.4
27	36	379.5	1.3	377.2	382.3
28	44	407.3	4.3	397.1	414.1
29	166	398.8	0.6	397.6	399.9
30	101	392.1	1.1	389.8	394.3
31	115	391.8	0.2	391.3	392.2
32	174	390.9	0.1	390.6	391.1
33	17	392.3	1.9	388.5	396.1
34	22	395.4	1.1	393.3	397.5
35	145	402.3	0.8	400.6	403.8
36	87	395.9	1.4	393.1	398.5
37	78	391.1	5.6	380.4	402.2

38	157	390.4	0.5	389.3	391.4
39	49	392.5	1.3	389.8	395.1
40	178	394.4	0.3	393.8	395.1
41	175	392.4	1.5	389.4	395.4
42		393.1	0.8	391.6	394.7




Figure 8. Estimated box means with 95% uncertainty intervals, reference value with 95% uncertainty in green, 95% prediction interval in red.

4. Particle Size Distribution

Informational values are provided for particle size distribution measured by laser diffraction. The values are means of the combined wet and dry PSD measurements as only nine data sets were returned. Two of these were deemed outliers as they fell distinctly below the cluster of results from other participants and represent the wet and dry data contribution from one participant. The black line represents the mean of the measurements, with the values for each size interval provided in Table 7.

4.1. The data

To determine the particle size distribution (PSD), laser diffraction measurements were performed on samples of the material either as dry powder or in liquid suspension. Table 7 gives the cumulative percent by bin size. The data in Table 7 can be represented by a bar graph [Fig. 9] with separate bars for dry and wet method.

Table 7. Cumulative % by bin size in microns for wet and dry measurements

Bin size in microns	Sample 1 wet	Sample 2 wet	Sample 3 wet	Sample 4 wet	Sample 1 dry	Sample 2 dry	Sample 3 dry	Sample 5 dry	Sample 7 dry
1	3.32	7.13	3.68	2.25	4.33	9.78	7.3	6.02	7.13
1.5	4.92	11.41	5.74	4.42	5.83	13.26	12.56	8.94	12.21
2	6.12	15.26	7.71	6.33	7.73	15.94	16.32	11.79	16.03
3	10.18	22.12	12.08	11.06	10.95	20.76	21.76	17.22	21.84
4	13.78	28.02	16.78	16.28	13.38	25.02	25.83	22.11	26.28
6	19.38	37.5	25.96	25.69	17.2	31.12	32.84	30.21	33.56
8	24.65	44.64	34.2	33.18	20.58	34.78	39.03	36.54	39.61
12	34.09	54.88	47.65	44.69	27.64	40.56	49.67	46.36	49.74
16	41.9	62.88	58.01	53.95	34.23	47.32	58.54	54.4	58.19
24	55.2	76.01	72.91	68.74	46.74	62.53	72.63	67.83	71.79
32	66.92	85.19	82.75	79.41	57.25	76.13	82.69	78.18	81.74
48	85.36	94	93.4	91.67	75.36	93.05	94.31	90.72	93.43
64	92.68	97.18	97.72	96.93	85.09	98.87	98.62	96.31	98.1
96	96.16	99.42	99.87	99.67	90.66	100	99.97	99.55	100
128	97.41	99.93	100	99.89	92.48	100	100	100	100




Figure 9. Bar graph of the cumulative frequencies given in Table 7 separated by method.

4.2. Statistical Analysis

Data in Table 7 can be combined using a Bayesian hierarchical model (estimated using Markov Chain Monte Carlo in OpenBUGS [13]) to obtain a consensus cumulative distribution table with values, their standard uncertainty, and 95% uncertainty interval per each bin. The results are given in Table 8. Consensus values with uncertainty for the cumulative particle size distribution are presented in Table 8 and are displayed as a graph in Figure 10.

Table 8. Consensus values with uncertainty for the cumulative particle size distribution.
The results in Table 8 are shown in Figure 10.

Bin size in microns	Consensus mean of cumulative %	Standard uncertainty of cumulative %	L95 in %	U95 in %
1	5.97	1.12	3.59	7.92
1.5	9.02	1.27	6.41	11.32
2	11.82	1.32	9.06	14.26
3	16.82	1.37	13.97	19.36
4	21.18	1.43	18.25	23.88
6	28.42	1.66	25.12	31.54
8	34.49	1.83	30.94	38.04
12	44.54	2.10	40.42	48.6
16	52.74	2.16	48.55	56.9
24	66.78	2.18	62.52	70.97
32	77.51	2.24	73.11	81.89
48	91.29	2.74	85.94	96.66
64	96.84	2.87	91.23	100
96	99.54	2.91	93.88	100
128	100	2.92	94.42	100




Figure 10. Estimated cumulative particle size distribution with 95% uncertainty bounds.

5. Acknowledgments

Mark Faijt, Buzzi Unicem, Greencastle IN for material acquisition. Sean Makens, Chair of ASTM C01.25, Kent Niedzielski CCRL for assistance in identifying prospective suppliers and for use of their large v-blender for homogenization, Mark Cronise for packaging, Theodore Dorian for providing a sieve calibration, Richard Eason for providing density measurements.

Participants

The assistance of the participants for providing the Blaine and the laser particle size distribution test data is gratefully acknowledged, including: Monika Adamova, Sudjendra Baijnath, Merlita Banez, Sonya Becknel, Nick Nennett, Ray Brison, Peter Cahill, Mary Carpenter, Doug Conroy, Benoit Dagnais, Wesley Dess, Amanda Dorsey, Melissa Eaton, Oscarw Echavarris, Brandon Eckard, Paul Engel, John Falletta, Xiuping Feng, Jane Lemming, Kristen Freeman, Tom Gepford, David Goldblume, Kevin Hall, Jessica

Hardenburg, Joe Howse, April Innes, Karolina Jagiello, Glenn Kosaki, Doug Kraszka, Prasad Kudiapur, Angelica Marclp, Simon McCarthy, Willy Morrison, Scott Nettles, Adam Oliver, Ed Rafacz, Wade Ramsdell, Michael Reylander, Lorraine Richards, Jason Schultz, Michael Stuart, Bob Sylvia, Violetta Vega

References

- [1] C.F. Ferraris, W. Gutherie, M. Pelts, and R. Haupt, “Certification of SRM 46H”, NIST Special Publication 260-169, January 2008, 32 pp.
- [2] ASTM C204, Standard Test Methods for Fineness of Hydraulic Cement by Air - Permeability Apparatus, ASTM Book of Standards, Vol. 4.02, ASTM International, West Conshocken, PA 2020
- [3] ASTM C430, Standard Test Method for Fineness of Hydraulic Cement by the 45- μm (No. 325) Sieve,” ASTM Book of Standards, Vol. 4.02, ASTM International, West Conshocken, PA 2020
- [4] AASHTO T 353-14, Standard Method of Test for Particle Size Analysis of Hydraulic Cement and Related Materials by Light Scattering” AASHTO Designation T353-14, American Association of State Highway and Transportation Officials, Washington, DC.
- [5] M.A. Helsel, C.F. Ferraris, and D. Bentz, “Comparative study of methods to measure the density of Cementious powders,” J. Test Eval. 44, no. 6. (2016) pp. 2147-2154.
<https://doi.org/10.1520/JTE20150148>
- [6] ASTM C188, “Standard Test Method for Density of Hydraulic Cement”.
- [7] ASTM C150, Standard Specification for Portland Cement, ASTM Book of Standards, Vol. 4.02, ASTM International, West Conshocken, PA 2020
- [8] ASTM C595 Standard Specification for Blended Hydraulic Cements, ASTM Book of Standards, ASTM International, West Conshocken, PA 2020
- [9] ASTM C1157, Standard Performance Specification for Hydraulic Cement, ASTM Book of Standards, ASTM International, West Conshocken, PA 2020
- [10] ASTM E11-20, Standard Specification for Woven Wire Test Sieve Cloth and Test Sieves ASTM International, West Conshohockin, PA
- [11] Possolo, A. , Koepke, A. , Newton, D. and Winchester, M. (2021), Decision Tree for Key Comparisons, Journal of Research (NIST JRES), National Institute of Standards and Technology, Gaithersburg, MD, [online], <https://doi.org/10.6028/jres.126.007>
- [12] Gelman A, Carlin J, Stern H, Dunson D, Vehtari A, Rubin D. Bayesian Data Analysis. Chapman & Hall, Boca Raton, 3rd edition, 2013.
- [13] Lunn DJ, Spiegelhalter D, Thomas A, Best N. The BUGS project: Evolution, critique and future directions (with discussion), Statistics in Medicine 2009;28:3049–3082.

Appendix A: Wet Sieve Test Results

Sieve Test Results by box and vial showing mass of cement and mass retained on each of three calibrated sieves.

Box	Vial	Mass	38 um		45 um		53 um	
			Mass	Retained	Mass	Retained	Mass	Retained
153	1	1.00018	0.09701	1.00044	0.05893	1.00061	0.03399	
153	2	1.00062	0.09818	1.00033	0.05981	1.00084	0.03433	
159	1	1.00010	0.10269	1.00072	0.06182	1.00089	0.03553	
159	2	1.00045	0.10329	1.00084	0.06147	1.00045	0.03588	
125	1	1.00095	0.09960	1.00040	0.06108	1.00061	0.03481	
125	2	1.00065	0.09713	1.00067	0.06240	1.00083	0.03390	
85	1	1.00028	0.10057	1.00033	0.06079	1.00021	0.03431	
85	2	1.00015	0.10280	1.00086	0.06294	1.00011	0.03479	
156	1	1.00086	0.10180	1.00043	0.06180	1.00052	0.03424	
156	2	1.00078	0.10148	1.00041	0.06082	1.00022	0.03331	
190	1	1.00090	0.09756	1.00096	0.06112	1.00083	0.03478	
190	2	1.00024	0.10341	1.00030	0.06079	1.00090	0.03574	
194	1	1.00036	0.10321	1.00045	0.06372	1.00085	0.03619	
194	2	1.00085	0.10396	1.00024	0.06394	1.00050	0.03663	
138	1	1.00004	0.10306	1.00014	0.06381	1.00094	0.03628	
138	2	1.00087	0.09748	1.00025	0.06439	1.00082	0.03671	
57	1	1.00027	0.10054	1.00090	0.06487	1.00046	0.03594	
57	2	1.00076	0.10272	1.00011	0.06448	1.00022	0.03694	
108	1	1.00055	0.10242	1.00040	0.06581	1.00016	0.03510	
108	2	1.00093	0.10483	1.00031	0.06597	1.00043	0.03790	
98	1	1.00040	0.10487	1.00063	0.06559	1.00066	0.03646	
98	2	1.00064	0.10336	1.00080	0.06570	1.00078	0.03681	
168	1	1.00088	0.10492	1.00082	0.06221	1.00015	0.03796	
168	2	1.00016	0.10093	1.00024	0.06583	1.00032	0.03691	
143	1	1.00091	0.10363	1.00088	0.06241	1.00048	0.03789	
143	2	1.00038	0.10421	1.00023	0.06532	1.00021	0.03649	
43	1	1.00124	0.10594	1.00001	0.06818	1.00015	0.03775	
43	2	1.00051	0.10491	1.00080	0.06891	1.00013	0.03780	
149	1	1.00083	0.10591	1.00008	0.06873	1.00080	0.03730	
149	2	1.00011	0.10971	1.00004	0.06866	1.00006	0.03645	

92	1	1.00072	0.10934	1.00015	0.06814	1.00076	0.03792
92	2	1.00128	0.10911	1.00065	0.06834	1.00026	0.03763
3	1	1.00021	0.10766	1.00029	0.06818	1.00005	0.03810
3	2	1.00017	0.10791	1.00063	0.06916	1.00071	0.03787
129	1	1.00061	0.10918	1.00079	0.06966	1.00010	0.03877
129	2	1.00011	0.10981	1.00098	0.06913	1.00067	0.03827

. +45 Sieve results for 38 um, 45um and 53 um sieves

Appendix B. Blaine fineness test results.

Sample Box	Blaine 114r #1		Blaine 114r #2		
	replicate	1	2	1	2
8	388	386	385	383	
15	395	387	387	391	
17	396	395	389	389	
21	390	392	390	393	
22	398	397	394	393	
26	391	391	391	391	
35	367	367	370	371	
36	379	381	381	375	
40	397	397	397	395	
44	400	412	418	418	
49	389	391	394	396	
58	395	391	395	391	
61	396	393	392	392	
70	392	393	393	391	
73	393	397	393	396	
78	375	364	405	415	
87	399	399	392	394	
88	372	375	369	375	
90	398	395	395	399	
101	394	390	394	390	
107	398	396	395	396	
110	396	399	396	399	
115	391	392	391	392	
122	392	393	389	390	
124	387	386	387	388	
131	385	386	396	396	
145	402	405	400	402	
148	397	396	391	392	
152	397	397	395	396	
157	389	390	391	392	
160	396	395	396	396	
166	400	398	398	398	
171	399	397	397	395	

173	396	397	398	399
174	391	391	391	391
175	394	396	389	391
176	390	390	387	390
177	403	399	396	397
178	394	395	394	395
196	377	377	378	377
200	395	397	396	397

Appendix C: OpenBUGS code for the sieve residue analysis

```
{  
for(i in 1:3)  
{  
pr.cv.s[i] <- 1/(uc.cv.s[i]*uc.cv.s[i])  
mu.s[i]~dnorm(cv.s[i],pr.cv.s[i])  
mu.r[i] <- b[1]+b[2]*(mu.s[i]-45.237)+b[3]*(mu.s[i]-45.237)  
  
xiNs[i] ~ dnorm(0, 0.0016)|(0.001,)  
chSqNs[i] ~ dgamma(0.5,0.5)  
tau.wv[i] <- xiNs[i]/sqrt(chSqNs[i])  
itau.wv[i]<-1/sqrt(tau.wv[i])  
}  
  
b[1]~dunif(0,20)  
b[2]~dunif(-5,0)  
b[3]~dunif(-1,1)  
b[4]~dunif(-1,1)  
b[5]~dunif(-1,1)  
b[6]~dunif(-1,1)  
for(i in 1:36)  
{  
y38[i]~dt(mu38[i],tau.wv[1],df[1])  
y45[i]~dt(mu45[i],tau.wv[2],df[2])  
y53[i]~dt(mu53[i],tau.wv[3],df[3])  
mu38[i] <- mu.r[1]+b[4]*(i-1)  
mu45[i] <- mu.r[2]+b[5]*(i-1)  
mu53[i] <- mu.r[3]+b[6]*(i-1)  
p.y38[i]~dt(mu38[i],tau.wv[1],df[1])  
p.y45[i]~dt(mu45[i],tau.wv[2],df[2])  
p.y53[i]~dt(mu53[i],tau.wv[3],df[3])  
res[i] <- y38[i]-p.y38[i]  
res[i+36] <- y45[i]-p.y45[i]  
res[i+72] <- y53[i]-p.y53[i]  
pv[i] <- step(res[i])  
pv[i+36] <- step(res[i+36])  
pv[i+72] <- step(res[i+72])  
}
```

```
mr45 <- b[1]+b[2]*(45-45.237)+b[3]*(45-45.237)*(45-45.237)
}
```

Data:

```
list(cv.s=c(38.12,43.47,54.12),uc.cv.s=c(0.59,0.59,0.59),df=c(200,200,200))
```

Box[]	Vial[]	y38[]	y45[]	y53[]	seq[]
153	1	9.699	5.89	3.397	1
153	2	9.812	5.979	3.43	2
159	1	10.268	6.178	3.55	3
159	2	10.324	6.142	3.586	4
125	1	9.951	6.106	3.479	5
125	2	9.707	6.236	3.387	6
85	1	10.054	6.077	3.43	7
85	2	10.278	6.289	3.479	8
156	1	10.171	6.177	3.422	9
156	2	10.14	6.08	3.33	10
190	1	9.747	6.106	3.475	11
190	2	10.339	6.077	3.571	12
194	1	10.317	6.369	3.616	13
194	2	10.387	6.392	3.661	14
138	1	10.306	6.38	3.625	15
138	2	9.74	6.437	3.668	16
57	1	10.051	6.481	3.592	17
57	2	10.264	6.447	3.693	18
108	1	10.236	6.578	3.509	19
108	2	10.473	6.595	3.788	20
98	1	10.483	6.555	3.644	21
98	2	10.329	6.565	3.678	22
168	1	10.483	6.216	3.795	23
168	2	10.091	6.581	3.69	24
143	1	10.354	6.236	3.787	25
143	2	10.417	6.53	3.648	26
43	1	10.581	6.818	3.774	27
43	2	10.486	6.885	3.78	28
149	1	10.582	6.872	3.727	29
149	2	10.97	6.866	3.645	30
92	1	10.926	6.813	3.789	31
92	2	10.897	6.83	3.762	32
3	1	10.764	6.816	3.81	33
3	2	10.789	6.912	3.784	34
129	1	10.911	6.961	3.877	35
129	2	10.98	6.906	3.824	36
END					

Appendix D. Consensus Means and Uncertainty Calculations

SRM 114

Consensus Means Analysis
(Full Sample Case)

Data Summary:

Response Variable: Y

Lab-ID Variable: IX

Number of Observations: 164

Grand Mean: 0.3921951E+03

Grand Standard Deviation: 0.8397705E+01

Total Number of Labs: 41

Minimum Lab Mean: 0.3687750E+03

Maximum Lab Mean: 0.4118250E+03

Minimum Lab SD: 0.2362908E+00

Maximum Lab SD: 0.2452251E+02

Mean of Lab Means: 0.3921951E+03

SD of Lab Means: 0.7464031E+01

SD of Lab Means (wrt to grand mean): 0.7464031E+01

Within Lab (pooled) SD: 0.4580903E+01

Within Lab (pooled) Variance: 0.2098467E+02

Table 1: Summary Statistics by Lab

Lab ID	n(i)	Mean	Standard Variance	Standard Deviation	Deviation of the Mean
1	4	0.3852500E+03	0.3470000E+01	0.1862794E+01	0.9313968E+00
2	4	0.3900250E+03	0.1694917E+02	0.4116937E+01	0.2058468E+01
3	4	0.3922750E+03	0.1344917E+02	0.3667311E+01	0.1833655E+01
4	4	0.3914000E+03	0.2253333E+01	0.1501111E+01	0.7505553E+00
5	4	0.3955000E+03	0.5293333E+01	0.2300725E+01	0.1150362E+01
6	4	0.3912000E+03	0.1466667E+00	0.3829708E+00	0.1914854E+00
7	4	0.3687750E+03	0.4689167E+01	0.2165448E+01	0.1082724E+01
8	4	0.3790250E+03	0.8022500E+01	0.2832402E+01	0.1416201E+01
9	4	0.3967750E+03	0.1102500E+01	0.1050000E+01	0.5250000E+00
10	4	0.4118250E+03	0.6498250E+02	0.8061172E+01	0.4030586E+01
11	4	0.3925250E+03	0.9642500E+01	0.3105238E+01	0.1552619E+01
12	4	0.3931500E+03	0.5416667E+01	0.2327373E+01	0.1163687E+01

13	4	0.3931000E+03	0.3813333E+01	0.1952776E+01	0.9763879E+00
14	4	0.3922500E+03	0.4366667E+00	0.6608076E+00	0.3304038E+00
15	4	0.3947250E+03	0.4002500E+01	0.2000625E+01	0.1000312E+01
16	4	0.3897000E+03	0.6013533E+03	0.2452251E+02	0.1226125E+02
17	4	0.3960500E+03	0.1036333E+02	0.3219213E+01	0.1609607E+01
18	4	0.3727000E+03	0.9240000E+01	0.3039737E+01	0.1519868E+01
19	4	0.3969000E+03	0.4486667E+01	0.2118175E+01	0.1059088E+01
20	4	0.3920500E+03	0.6803333E+01	0.2608320E+01	0.1304160E+01
21	4	0.3961750E+03	0.2409167E+01	0.1552149E+01	0.7760745E+00
22	4	0.3975500E+03	0.1963333E+01	0.1401190E+01	0.7005950E+00
23	4	0.3917500E+03	0.3033333E+00	0.5507571E+00	0.2753785E+00
24	4	0.3909500E+03	0.3376667E+01	0.1837571E+01	0.9187854E+00
25	4	0.3870250E+03	0.3758333E+00	0.6130525E+00	0.3065262E+00
26	4	0.3906250E+03	0.3504917E+02	0.5920234E+01	0.2960117E+01
27	4	0.4023750E+03	0.3375833E+01	0.1837344E+01	0.9186720E+00
28	4	0.3939000E+03	0.9180000E+01	0.3029851E+01	0.1514926E+01
29	4	0.3961500E+03	0.9166667E+00	0.9574271E+00	0.4787136E+00
30	4	0.3903500E+03	0.1443333E+01	0.1201388E+01	0.6006940E+00
31	4	0.3956750E+03	0.6225000E+00	0.7889867E+00	0.3944933E+00
32	4	0.3988250E+03	0.1102500E+01	0.1050000E+01	0.5250000E+00
33	4	0.3969250E+03	0.1715833E+01	0.1309898E+01	0.6549491E+00
34	4	0.3975500E+03	0.1316667E+01	0.1147461E+01	0.5737305E+00
35	4	0.3908500E+03	0.5666667E-01	0.2380476E+00	0.1190238E+00
36	4	0.3924500E+03	0.9283333E+01	0.3046856E+01	0.1523428E+01
37	4	0.3890250E+03	0.1322500E+01	0.1150000E+01	0.5750000E+00
38	4	0.3988000E+03	0.9286667E+01	0.3047403E+01	0.1523702E+01
39	4	0.3944250E+03	0.3825000E+00	0.6184658E+00	0.3092329E+00
40	4	0.3773750E+03	0.5583333E-01	0.2362908E+00	0.1181454E+00
41	4	0.3960500E+03	0.9166667E+00	0.9574271E+00	0.4787136E+00

1. Method: Mandel-Paule

Estimate of (unscaled) Consensus Mean:	0.3921526E+03
Estimate of (scaled) Consensus Mean:	0.5430326E+00
Between Lab Variance (unscaled):	0.5223863E+02
Between Lab SD (unscaled):	0.7227630E+01
Between Lab Variance (scaled):	0.2818678E-01
Standard Deviation of Consensus Mean:	0.1135894E+01
Standard Uncertainty ($k = 1$):	0.1135894E+01
Expanded Uncertainty ($k = 2$):	0.2271788E+01
Expanded Uncertainty ($k = 1.9599640$):	0.2226311E+01
Normal PPF of 0.975:	0.1959964E+01
Lower 95% (normal) Confidence Limit:	0.3899262E+03
Upper 95% (normal) Confidence Limit:	0.3943789E+03

Note: Mandel-Paule Best Usage:

6 or More Labs:

3. Method: Vangel-Rukhin Maximum Likelihood

Estimate of (unscaled) Consensus Mean:	0.3920690E+03
Estimate of (scaled) Consensus Mean:	0.5410923E+00
Between Lab Variance (unscaled):	0.4807536E+02
Between Lab SD (unscaled):	0.6933640E+01
Between Lab Variance (scaled):	0.2594037E-01

Standard Deviation of Consensus Mean:	0.1111312E+01
Standard Uncertainty (k = 1):	0.1111312E+01
Expanded Uncertainty (k = 2):	0.2222624E+01
Expanded Uncertainty (k = 1.9599640):	0.2178131E+01
Normal PPF of 0.975:	0.1959964E+01
Lower 95% (normal) Confidence Limit:	0.3898909E+03
Upper 95% (normal) Confidence Limit:	0.3942472E+03

Note: Vangel-Rukhin Maximum Likelihood
Best Usage: 6 or More Labs

4a. Method: DerSimonian Laird (original variance)

Estimate of Consensus Mean:	0.3921593E+03
Estimate of Variance of Consensus Mean:	0.1518269E+01
Estimate of Between Lab Variance:	0.5966491E+02
Standard Uncertainty (k = 1):	0.1232181E+01
Expanded Uncertainty (k = 2):	0.2464361E+01
Degrees of Freedom:	40
t Percent Point Value:	0.2021075E+01
Lower 95% (t-value) Confidence Limit:	0.3896689E+03
Upper 95% (t-value) Confidence Limit:	0.3946496E+03

Note: DerSimonian-Laird Best Usage:
Any Number of Labs:

4b. Method: DerSimonian Laird - Horn-Horn-Duncan Variance

Estimate of Consensus Mean:	0.3921593E+03
Estimate of Variance of Consensus Mean:	0.1328099E+01
Estimate of Between Lab Variance:	0.5966491E+02
Standard Uncertainty (k = 1):	0.1152432E+01
Expanded Uncertainty (k = 2):	0.2304863E+01
Degrees of Freedom:	40
t Percent Point Value:	0.2021075E+01
Lower 95% (t-value) Confidence Limit:	0.3898301E+03
Upper 95% (t-value) Confidence Limit:	0.3944884E+03

Note: DerSimonian-Laird Best Usage:
Any Number of Labs:

4d. Method: DerSimonian Laird - Bootstrap Variance

Number of Bootstrap Samples	100000
Estimate of Consensus Mean:	0.3921593E+03
Estimate of Variance of Consensus Mean:	0.1522168E+01
Standard Uncertainty (k = 1):	0.1233762E+01
Expanded Uncertainty (k = 2):	0.2467524E+01
Lower 95% (percentile bootstrap) Confidence Limit:	0.3897386E+03
Upper 95% (percentile bootstrap) Confidence Limit:	0.3945844E+03
Lower 95% (symmetric bootstrap) Confidence Limit:	0.3897341E+03
Upper 95% (symmetric bootstrap) Confidence Limit:	0.3945844E+03
K (symmetric bootstrap) Coverage Factor:	0.1965654E+01
Lower 95% (kernel bootstrap) Confidence Limit:	0.3897266E+03
Upper 95% (kernel bootstrap) Confidence Limit:	0.3945951E+03
K (kernel bootstrap) Coverage Factor:	0.1974298E+01

Note: DerSimonian-Laird Best Usage:

Any Number of Labs:

Table 2: 95% Confidence Limits

Method	Consensus Mean	Lower Limit	Upper Limit	Uncertainty (k*SE)
1. Mandel-Paule	0.3921526E+03	0.3899262E+03	0.3943789E+03	0.2226311E+01
3a. Vangel-Rukhin ML	0.3920690E+03	0.3898909E+03	0.3942472E+03	0.2178131E+01
4a. DerSimonian-Laird (original)	0.3921593E+03	0.3896689E+03	0.3946496E+03	0.2490330E+01
4b. DerSimonian-Laird (H-H-D)	0.3921593E+03	0.3898301E+03	0.3944884E+03	0.2329151E+01
4d. DerSimonian-Laird (perc. bootstrap)	0.3921593E+03	0.3897386E+03	0.3945844E+03	0.2425148E+01
4d. DerSimonian-Laird (symm. bootstrap)	0.3921593E+03	0.3897341E+03	0.3945844E+03	0.2425148E+01
4d. DerSimonian-Laird (kern bootstrap)	0.3921593E+03	0.3897266E+03	0.3945951E+03	0.2435813E+01

Table 3: Standard Uncertainties (k = 1)

Method	Standard Consensus Mean	Relative Uncertainty (k = 1)	Standard Uncertainty (%)
1. Mandel-Paule	0.3921526E+03	0.1135894E+01	0.2896562E+00
3a. Vangel-Rukhin ML	0.3920690E+03	0.1111312E+01	0.2834480E+00
4a. DerSimonian-Laird (original)	0.3921593E+03	0.1232181E+01	0.3142042E+00
4b. DerSimonian-Laird (H-H-D)	0.3921593E+03	0.1152432E+01	0.2938683E+00
4d. DerSimonian-Laird (bootstrap)	0.3921593E+03	0.1233762E+01	0.3146073E+00

Table 4: Expanded Uncertainties (k = 2)

Method	Consensus Mean	Expanded Uncertainty (k = 2)	Relative Expanded Uncertainty (%)
1. Mandel-Paule	0.3921526E+03	0.2271788E+01	0.5793123E+00
3a. Vangel-Rukhin ML	0.3920690E+03	0.2222624E+01	0.5668960E+00
4a. DerSimonian-Laird (original)	0.3921593E+03	0.2464361E+01	0.6284083E+00
4b. DerSimonian-Laird (H-H-D)	0.3921593E+03	0.2304863E+01	0.5877366E+00
4d. DerSimonian-Laird (bootstrap)	0.3921593E+03	0.2467524E+01	0.6292147E+00

Appendix E. Dataplot code for consensus means

```
. Name: blaine.dp
. Purpose: Run a consensus means analysis for blaine.dat

. Step 0: Define the output devices

device 2 close
let string fplot = blaine.ps
set ipl1na ^fplot
call checksys.dp

. Step 1: Read the data

skip 25
read blaine.dat x y1 y2 y3 y4
skip 0

let x = sortc x y1 y2 y3 y4
let nlab = size x
loop for k = 1 1 nlab
    let aval = x(k)
    let string s^k = ^aval
end of loop

let ix = code x
let ix = combine ix ix ix ix
let y = combine y1 y2 y3 y4

. Step 2: Initializations for consensus means analysis

set write decimals -7
set modified mandel paule off
set vangel rukhin on
set vangel rukhin bootstrap off
set dersimonian laird minmax off
set dersimonian laird bootstrap on
set schiller eberhardt off
set mean of means off
set grand mean off
set graybill deal off
set generalized confidence interval off
set fairweather off
set bayesian consensus procedure off
set bob off
bootstrap samples 100000
set random number generator fibonacci congruential
```

```
seed 66783
.
let string sx1 = 1
let string sx2 = 2
let string sx3 = 3
let string sx4 = 4
let string sx5 = 5
let nmeth = 5
.
let string srmid = SRM 114
date
.
. Step 3: Set plot control features
.
frame corner coordinates 15 20 90 90
line solid all
character blank all
character fill on
. character hw 1 0.75
character hw 0.5 0.375
character circle
line blank
.
case asis
title asis
tic mark label case asis
title offset 2
.
. Step 3b: Generate strings for lab data
.
loop for k = 1 1 nlab
    let icnt = nmeth + k
    . let ival = mod(k,2)
    . if ival = 1
        let string sx^icnt = ^s^k
    . else
        . let string sx^icnt = sp()cr()^s^k
    . end of if
end of loop
let ntot = nmeth + nlab
.
xlimits 1 ntot
major xtic mark number ntot
minor xtic mark number 0
tic mark offset units data
xtic mark offset 0.5 0.5
```

```
x1tic mark label size 1.7
x1tic mark label direction vertical
title case asis
title offset 2
.
. iflagm = 1 => 95% Confidence Interval
.     = 2 => Two Standard Errors Confidence Interval
.     = 3 => One Standard Errors Confidence Interval
.
let iflagm = 2
. set consensus mean plot sorted on
if iflagm = 1
    set consensus mean plot error confidence intervals
else if iflagm = 2
    set consensus mean plot error two standard errors
else if iflagm = 3
    set consensus mean plot error one standard errors
end of if
.
feedback off
capture screen on
capture cons_114.out
capture suspend
.
capture resume
print "
"
print "SRM: ^srmid"
print "Date: ^currdate"
print " "
print " "
.
loop for l = 1 1 ntot
    let string sxnew^l = ^sx^l
end of loop
.
title ^srmid
x1tic mark label format group label
let igx = group label sxnew1 to sxnew^ntot
x1tic mark label content igx
.
consensus mean plot y ix
.
capture suspend
.
line dotted
```

```

let xcoor = nmeth + 0.5
drawdsds xcoor 20 xcoor 90
justification center
moveds 2.5 8
if iflagm = 1
    text Consensus Method: 95% Confidence Limits
else if iflagm = 2
    text Consensus Method: Two Standard Errors
else if iflagm = 3
    text Consensus Method: One Standard Error
end of if
moveds 25 8
text Method/Laboratorycr()meansp() +/- 2*sd/sqrt(n)
line blank
just right
move 98 4
height 1.5
text Date: ^currdate

height 1.4
justification left
move 91 88
text Method:
move 91 86
text 1: MP
move 91 84
text 2: VR
move 91 82
text 3: DSL
move 93 80
text Original
move 91 78
text 4: DSL
move 93 76
text HHD
move 91 74
text 5: DSL
move 93 72
text Bootstrap

end of capture
feedback on

Step 4: Device close

device 2 close

```

```
system ps2pdf ^fplot  
.quit
```

Appendix F Blaine Fineness Test Results formatted for Dataplot

Name: blaine.dat
Source: Paul Stutzman

Description: Column 1: Lab ID

Column 2: Vial 114r #1 Replicate 1
Column 3: Vial 114r #1 Replicate 2
Column 4: Vial 114r #2 Replicate 1
Column 5: Vial 114r #2 Replicate 2

X	Y1	Y2	Y3	Y4
8	387.6	385.5	384.8	383.1
15	395.4	386.8	386.8	391.1
17	395.5	395.4	389.2	389.0
21	390.1	391.9	390.3	393.3
22	398.0	396.8	394.2	393.0
26	391.5	390.7	391.5	391.1
35	366.6	367.3	370.1	371.1
36	378.8	381.1	381.1	375.1
40	397.3	397.3	397.3	395.2
44	400.4	411.9	417.5	417.5
49	389.2	390.7	394.3	395.9
58	394.9	390.9	395.4	391.4
61	395.9	392.9	391.5	392.1
70	392.3	392.7	392.7	391.3
73	392.7	397.1	393.5	395.6
78	374.8	363.5	405.2	415.3
87	398.5	399.0	392.4	394.3
88	371.8	374.8	368.8	375.4
90	398.2	395.1	395.1	399.2
101	394.1	389.6	394.5	390.0
107	398.3	395.7	394.6	396.1
110	396.4	398.5	396.3	399.0
115	391.4	392.4	391.2	392.0
122	391.7	393.2	389.4	389.5
124	386.6	386.4	387.5	387.6
131	385.3	385.7	395.7	395.8
145	402.4	404.6	400.1	402.4
148	396.8	396.2	390.9	391.7

152 396.9 396.9 394.9 395.9
157 388.9 390.1 390.6 391.8
160 396.4 394.6 395.6 396.1
166 400.4 398.3 398.3 398.3
171 398.6 396.9 396.8 395.4
173 396.2 397.4 397.6 399.0
174 391.1 390.7 391.0 390.6
175 393.6 396.0 388.9 391.3
176 389.6 389.6 387.3 389.6
177 403.1 398.8 396.4 396.9
178 394.3 394.9 393.6 394.9
196 377.4 377.2 377.7 377.2
200 394.8 396.8 395.8 396.8