

Standard Reference Material[®] 1570a

Trace Elements in Spinach Leaves

CERTIFICATE OF ANALYSIS

Purpose: The certified values delivered by this Standard Reference Material (SRM) are intended for use in validating analytical methods for the determination of major, minor, and trace elements in botanical materials, agricultural food products, and materials of similar matrix.

Description: A unit of SRM 1570a consists of 60 g of powdered dried spinach leaves.

Certified Values: Certified mass fraction values for selected constituent elements, reported on a dry-mass basis, are provided in Table 1. A NIST certified value is a value for which NIST has the highest confidence in its accuracy in that all known or suspected sources of bias have been investigated or taken into account [1]. Certified mass fraction values are weighted means of results from two or more different analytical methods combined using the DerSimonian-Laird procedure [2]. The uncertainty in the certified mass fraction values was calculated according to the methods in Supplement 1 to the ISO/JCGM Guide [3] and the results are consistent with the methods given in the ISO/JCGM Guide [4]. The uncertainty of each certified value is expressed as $U = ku_c$. The quantity u_c is the combined standard uncertainty, which accounts for the combined effect of within-method uncertainty from all potential sources and any bias between methods at the level of one standard deviation. The coverage factor, k , is determined from the Student's t -distribution corresponding to the appropriate associated degrees of freedom and a 95 % level of confidence for each analyte. NIST certified values are traceable to the International System of Units (SI) derived unit of mass fraction, expressed as percent.

Table 1. Certified Mass Fraction Values (Dry-Mass Basis) of Constituent Elements

Element	Mass Fraction (%)	Coverage Factor (k)
Calcium (Ca)	1.526 ± 0.066	2.0299
Phosphorus (P)	0.5187 ± 0.0067	1.9772
Potassium (K)	2.900 ± 0.026	2.3226
Sodium (Na)	1.821 ± 0.023	1.9943

Element	Mass Fraction (mg/kg)	Coverage Factor (k)	Element	Mass Fraction (mg/kg)	Coverage Factor (k)
Aluminum (Al)	310 ± 15	2.0102	Mercury (Hg)	0.0297 ± 0.0021	2.0492
Arsenic (As)	0.068 ± 0.012	2.0561	Nickel (Ni)	2.142 ± 0.058	1.9709
Boron (B)	37.7 ± 1.2	2.0127	Selenium (Se)	0.1152 ± 0.0043	2.0371
Cadmium (Cd)	2.876 ± 0.058	2.0464	Strontium (Sr)	55.54 ± 0.50	2.5029
Cobalt (Co)	0.393 ± 0.030	2.0320	Thorium (Th)	0.0480 ± 0.0017	2.0053
Copper (Cu)	12.22 ± 0.86	2.0488	Vanadium (V)	0.568 ± 0.017	2.4938
Manganese (Mn)	76.0 ± 1.2	1.9855	Zinc (Zn)	82.3 ± 3.9	2.0136

Period of Validity: The certified values delivered by **SRM 1570a** are valid within the measurement uncertainty specified until **31 August 2027**. The certified values are nullified if the material is stored or used improperly, damaged, contaminated, or otherwise modified.

Carlos A. Gonzalez, Chief
Chemical Sciences Division
Certificate History on Page 3

Steven J. Choquette, Director
Office of Reference Materials

Non-Certified Values: Non-certified values are provided in Appendix A.

Additional Information: Values of potential interest to users and additional information are provided in Appendix B.

Maintenance of Certified Values: NIST will monitor this SRM over the period of its validity. If substantive technical changes occur that affect the certification, NIST will issue an amended certificate through the NIST SRM website (<https://www.nist.gov/srm>) and notify registered users. SRM users can register online from a link available on the NIST SRM website or fill out the user registration form that is supplied with the SRM. Registration will facilitate notification. Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (<https://www.nist.gov/srm>).

Safety: Consult the Safety Data Sheet (SDS) for hazard information.

Storage: The material should be kept in its tightly closed original bottle and stored in the dark at a temperature between 10 °C and 30 °C. It should not be exposed to intense sources of electromagnetic radiation. Ideally, the bottle should be kept in a desiccator under the conditions indicated above. Spinach leaves tend to rapidly bleach and turn a tan or light brown color in the presence of visible light. By monitoring SRM 1570, the previous version of this SRM, it was determined that there is no evidence of any change in elemental mass fractions as a result of the color change.

Use: The contents of a bottle should be thoroughly mixed by rotating and/or rolling before each use. Allow the contents to settle for 1 minute prior to opening to minimize the loss of fine dust particles. A minimum sample mass of 150 mg of the material, dried as described in the section (see “Instructions for Drying”), should be used to relate analytical determinations to the certified values on this certificate. In some cases, especially for volatile elements such as mercury, it is preferable to analyze samples from the bottle without drying, determine the moisture content on a separate sample from the same bottle taken at the same time, and convert the analytical results to a dry-mass basis.

Digestion procedures should be designed to avoid loss of volatile elements, such as arsenic and mercury. Digestion of the SRM in nitric and perchloric acids was found to be incomplete, with a small residue of siliceous material remaining. This residue must be considered an integral part of this SRM and should be dissolved with a small amount of hydrofluoric acid to obtain total dissolution. All certified values are based on the total dissolution.

Instructions for Drying: Samples of this SRM must be dried by one of the following two procedures in order for certified values to be valid:

1. Drying in a desiccator at room temperature for 120 h over fresh anhydrous magnesium perchlorate. The sample depth should not exceed 1 cm.
2. Freeze-drying for 24 h at a pressure of 13.3 Pa or lower and a shelf temperature of -5 °C or lower after having frozen the sample (not to exceed 1 cm in depth) at -40 °C or lower for at least 1 h. At the end of the 24 h period, samples should be placed immediately in a desiccator with fresh anhydrous magnesium perchlorate. Samples should be weighed after allowing a minimum of 4 h to establish temperature equilibrium.

Note: An approximate mass loss on drying of 3.5 % was observed for the measurements reported here. Vacuum drying at room temperature and oven drying at elevated temperatures have resulted in excessive mass losses and therefore are **NOT** recommended.

REFERENCES

- [1] Beauchamp, C.R.; Camara, J.E.; Carney, J.; Choquette, S.J.; Cole, K.D.; DeRose, P.C.; Diewer, D.L.; Epstein, M.S.; Kline, M.C.; Lippa, K.A.; Lucon, E.; Molloy, J.; Nelson, M.A.; Phinney, K.W.; Polakoski, M.; Possolo, A.; Sander, L.C.; Schiel, J.E.; Sharpless, K.E.; Toman, B.; Winchester, M.R.; Windover, D.; *Metrological Tools for the Reference Materials and Reference Instruments of the NIST Material Measurement Laboratory*; NIST Special Publication (NIST SP) 260-136, 2021 edition; National Institute of Standards and Technology, Gaithersburg, MD (2021); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-136-2021.pdf> (accessed May 2023).
- [2] DerSimonian, R; Laird, N.; *Meta-Analysis in Clinical Trials*; Controlled Clin. Trials, Vol. 7, pp. 177–188 (1986).
- [3] JCGM 101:2008; *Evaluation of Measurement Data – Supplement 1 to the “Guide to the Expression of Uncertainty in Measurement” - Propagation of Distributions using a Monte Carlo Method*; JCGM (2008); available at <https://www.bipm.org/en/committees/jc/jcgm/publications> (accessed May 2023).
- [4] JCGM 100:2008; *Evaluation of Measurement Data - Guide to the Expression of Uncertainty in Measurement*; (GUM 1995 with Minor Corrections), Joint Committee for Guides in Metrology (2008); available at <https://www.bipm.org/en/committees/jc/jcgm/publications> (accessed May 2023); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at <https://www.nist.gov/pml/nist-technical-note-1297> (accessed May 2023).

Certificate Revision History: **08 May 2023** (Change of period of validity; updated format; editorial changes) **25 February 2014** (Extension of certification period; updated certified values and uncertainties; removed reference and information values for proximates, calories, total dietary fiber, fatty acids, and nitrogen (organic and protein) due to instability of organic constituents; editorial changes); **08 October 2008** (Update of expiration date; editorial changes); **31 August 2001** (This technical revision reports the addition of reference and information values for proximates, calories, total dietary fiber, and fatty acids and a change from non-certified to reference and information values for several inorganic constituents); **15 July 1996** (Editorial changes); **20 October 1994** (Original certificate date).

Certain commercial equipment, instruments, or materials may be identified in this Certificate of Analysis to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.

Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the Office of Reference Materials 100 Bureau Drive, Stop 2300, Gaithersburg, MD 20899-2300; telephone (301) 975-2200; e-mail srminfo@nist.gov; or the Internet at <https://www.nist.gov/srm>.

* * * * * End of Certificate of Analysis * * * * *

APPENDIX A

Non-Certified Mass Fraction Values: Non-certified mass fraction values of constituent elements, reported on a dry-mass basis, are provided in Table A1. Non-certified values are values that are the best estimates of the true values based on available data; however, the values do not meet NIST criteria for certification [1] and are provided with associated uncertainties that may reflect only measurement reproducibility, may not include all sources of uncertainty, or may reflect a lack of sufficient statistical agreement among multiple analytical methods.

Each non-certified value, expressed as a mass fraction on a dry-mass basis, is an equally weighted mean of results provided by NIST and/or collaborating laboratories. The provided uncertainties represent a 95 % level of confidence for each analyte.

Table A1. Non-Certified Mass Fraction Values (Dry-Mass Basis) of Constituent Elements

Element	Mass Fraction (%)
Nitrogen (N) (Total)	6.06 ± 0.20

Element	Mass Fraction (mg/kg)	Element	Mass Fraction (mg/kg)
Europium (Eu)	0.0055 ± 0.0010	Rubidium (Rb)	12.7 ± 1.6
Scandium (Sc)	0.0055 ± 0.0006	Uranium (U)	0.155 ± 0.023

Period of Validity: The non-certified values are valid within the measurement uncertainty specified until **31 August 2027**. The value assignments are nullified if the material is stored or used improperly, damaged, contaminated, or otherwise modified.

Maintenance of Non-Certified Values: NIST will monitor this material to the end of its period of validity. If substantive technical changes occur that affect the non-certified values during this period, NIST will update this Appendix and notify registered users. SRM users can register online from a link available on the NIST SRM website or fill out the user registration form that is supplied with the SRM. Registration will facilitate notification. Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (<https://www.nist.gov/srm>).

***** End of Appendix A *****

APPENDIX B

Values of Potential Interest: Values of potential interest for mass fraction for additional constituent elements are provided in Table B1. A NIST value of potential interest is a value that may be of interest to the SRM user, but insufficient information is available to assess the uncertainty associated with the value, therefore no uncertainty is provided [1]. Values are reported on a dry-mass basis.

Each value of potential interest, expressed as a mass fraction on a dry-mass basis, is an equally weighted mean of results provided by NIST and/or collaborating laboratories. Insufficient information is available to assess the uncertainty associated with the value, therefore no uncertainty is provided.

Table B1. Values of Potential Interest Mass Fraction (Dry-Mass Basis) of Constituent Elements

Element	Mass Fraction (%)
Magnesium (Mg)	0.9
Sulfur (S)	0.5

Element	Mass Fraction (mg/kg)
Lead (Pb)	0.2

Source and Preparation of Material: The material (approximately 2270 kg) for this SRM was obtained from commercial supplier Oregon Freeze-Drying Corp. (Albany, OR). It consists of U.S. Grade A chopped frozen spinach. The material was thawed, placed in a ribbon mixer, thoroughly mixed, and blended. After mixing, the spinach was freeze-dried. The freeze-dried material was then ground in a stainless steel grinder and shipped to NIST. At NIST, the freeze-dried material was sieved through a polypropylene sieve having openings of 0.25 mm (equivalent to a U.S. Series 60 standard sieve). The sieved material was then jet milled and air classified to a particle size of approximately 75 μm (200 mesh). After mixing in a large blender, the spinach was irradiated with cobalt-60 radiation to a minimum absorbed dose of approximately 27.8 kGy for microbiological control and was bottled.

Elemental Analysis: Value assignment of the certified mass fractions of the elements in SRM 1570a was based on the combination of measurements from two or more different analytical methods at NIST and collaborating laboratories. NIST and collaborating laboratories provided measurements by using colorimetry (COLOR), cold-vapor atomic absorption spectrometry (CVAAS), flow injection hydride generation atomic absorption spectrometry (FI-HGAAS), inductively coupled plasma optical emission spectrometry (ICPOES), isotope dilution inductively coupled plasma mass spectrometry (IDICPMS), isotope dilution thermal ionization mass spectrometry (IDTIMS), instrumental neutron activation analysis (INAA), laser-excited atomic fluorescence spectrometry (LEAFS), prompt gamma activation analysis (PGAA), and radiochemical neutron activation analysis (RNAA). Data from an IAEA interlaboratory comparison exercise were also used where available. A list of analytical methods used for measurement of each element is provided in Appendix B, Table B2.

Homogeneity Assessment: Samples from randomly selected bottles of SRM 1570a were tested for homogeneity. No evidence of statistically significant inhomogeneity was observed.

Table B2. Methods Used in Elemental Determinations

Element	Method Code	Element	Method Code
Aluminum (Al)	ICPOES INAA	Nitrogen (N)	PGAA
Arsenic (As)	FI-HGAAS RNAA	Phosphorus (P)	COLOR ICPOES
Boron (B)	IDICPMS PGAA	Potassium (K)	IDTIMS INAA
Cadmium (Cd)	IDICPMS PGAA RNAA	Rubidium (Rb)	IAEA INAA
Calcium (Ca)	IDTIMS INAA	Scandium (Sc)	IAEA INAA
Cobalt (Co)	INAA RNAA	Selenium (Se)	FI-HGAAS INAA RNAA
Copper (Cu)	ICPOES RNAA	Sodium (Na)	PGAA INAA
Europium (Eu)	IAEA INAA	Strontium (Sr)	IDTIMS INAA
Lead (Pb)	IAEA IDICPMS	Sulfur (S)	PGAA IAEA
Magnesium (Mg)	IDICPMS	Thorium (Th)	INAA RNAA
Manganese (Mn)	INAA LEAFS	Uranium (U)	RNAA
Mercury (Hg)	CVAAS RNAA	Vanadium (V)	IDTIMS INAA
Nickel (Ni)	IDICPMS RNAA	Zinc (Zn)	ICPOES INAA

Key:

- COLOR: Colorimetry
- CVAAS: Cold-vapor atomic absorption spectrometry
- FI-HGAAS: Flow injection hydride generation atomic absorption spectrometry
- IAEA: Various methods from an IAEA interlaboratory comparison exercise.
- ICPOES: Inductively coupled plasma optical emission spectrometry
- IDICPMS: Isotope dilution inductively coupled plasma mass spectrometry
- IDTIMS: Isotope dilution thermal ionization mass spectrometry
- INAA: Instrumental neutron activation analysis
- LEAFS: Laser-excited atomic fluorescence spectrometry
- PGAA: Prompt gamma activation analysis
- RNAA: Radiochemical neutron activation analysis

* * * * * End of Appendix B * * * * *