

# Standard Reference Material® 1662a

## Sulfur Dioxide in Nitrogen

### (Nominal Amount-of-Substance Fraction 1000 $\mu\text{mol/mol}$ )

### Lot 93-I-XX

## CERTIFICATE OF ANALYSIS

**Purpose:** The certified value delivered by this Standard Reference Material (SRM) is intended for use in producing metrologically traceable secondary standards for the calibration of instruments used for sulfur dioxide determinations.

**Description:** This SRM is a primary gas mixture supplied in a DOT 3AL-specification aluminum (6061 alloy) cylinder with a water volume of 6 L. Mixtures are shipped with a nominal pressure exceeding 12.4 MPa (1800 psig), which provides the user with 0.73 m<sup>3</sup> (25.8 ft<sup>3</sup>) of useable mixture. The cylinder is the property of the purchaser and is equipped with a CGA-660 stainless-steel valve, which is the recommended outlet for this sulfur dioxide mixture.

**Certified Value:** A NIST certified value is a value for which NIST has the highest confidence in that all known or suspected sources of bias and imprecision have been accounted [1]. This SRM mixture has been certified for sulfur dioxide amount-of-substance fraction (mole fraction, sometimes termed “molar concentration”). This certified value is traceable to the International System of Units (SI) through the gravimetric primary standards and procedures used in the preparation of this mixture. The certified value given below applies to all cylinders in this lot.

Sulfur Dioxide Mole Fraction: 985.9  $\mu\text{mol/mol}$   $\pm$  3.7  $\mu\text{mol/mol}$ <sup>(a)</sup>

<sup>(a)</sup> The certified value is expressed as  $x \pm U_{95\%}(x)$ , where  $x$  is the value and  $U_{95\%}(x)$  is the expanded uncertainty of the value. The true value of the analyte lies within the interval  $x \pm U_{95\%}(x)$  with 95 % confidence. For guidance in propagating this uncertainty, see reference 2.

**Additional Information:** Appendix A contains a table of all cylinder Sample Numbers and their associated information for this lot. Appendix B lists information on other components in the mixture.

**Period of Validity:** The certified value delivered by **SRM 1662a, Lot No. 93-I-XX** is valid within the measurement uncertainty specified until **27 January 2028**, provided the SRM is handled and stored in accordance with the instructions given in this certificate (see “Storage” and “Cylinder and Gas Handling Information”). The certification is nullified if the SRM is damaged, contaminated or otherwise modified, or if the internal pressure drops below 0.7 MPa (100 psig).

**Maintenance of Certified Value:** NIST will monitor this SRM over the period of its validity. If substantive technical changes occur that affect the certification, NIST will issue an amended certificate through the NIST SRM website (<https://www.nist.gov/srm>). Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (<https://www.nist.gov/srm>).

Carlos A. Gonzalez, Chief  
Chemical Sciences Division  
*Certificate of Analysis Revision History on Page 2*

Steven J. Choquette, Director  
Office of Reference Materials

**Safety:** Consult the Safety Data Sheet (SDS) for hazard information.

**Storage:** This SRM should be stored under normal laboratory conditions within the temperature range of 15 °C to 30 °C.

**Cylinder and Gas Handling Information:** NIST recommends the use of a high purity, two stage pressure regulator with a stainless-steel diaphragm and CGA-660 outlet to safely reduce the pressure and to deliver this mixture to the instrument. The regulator should be purged to prevent accidental contamination of the SRM by repeatedly (minimum three times) opening the valve and pressurizing the regulator, then closing the valve and releasing the pressure safely into a vent line.

**Mixture Preparation:** The gas mixtures comprising this SRM lot were prepared in accordance with NIST technical specifications by a commercial specialty gas vendor under contract to NIST. The specifications stipulate that each SRM mixture be identical in sulfur dioxide mole fraction and stable with time.

**Analytical Method:** Analyses of the sulfur dioxide mole fraction for this lot of cylinders were conducted by comparison to a representative cylinder chosen from the lot, designated as the SRM lot standard (LS). The LS was compared to NIST primary gravimetric standards using Nondispersive Ultraviolet analyzer (NDUV). Each of the sulfur dioxide mixtures that comprise this SRM lot was then compared to the LS using NDUV. Within the precision of the NIST measurements, all the cylinders comprising this SRM lot have identical sulfur dioxide mole fractions.

**CAS Registry Numbers:** This SRM is certified for sulfur dioxide in nitrogen. The relevant CAS Registry numbers for these components are: sulfur dioxide CAS Registry 7446-09-5; nitrogen (balance gas) CAS Registry 7727-37-9.

## NOTICE TO USERS

NIST strives to maintain the SRM inventory supply, but NIST cannot guarantee the continued or continuous supply of any specific SRM. Accordingly, NIST encourages the use of this SRM as a primary benchmark for the quality and accuracy of the user's in house reference materials and working standards. As such, the SRM should be used to validate the more routinely used reference materials in a laboratory. Comparisons between the SRM and in house reference materials or working measurement standards should take place at intervals appropriate to the conservation of the SRM and the stability of relevant in house materials. For further guidance on how this approach can be implemented, contact NIST by email at [srms@nist.gov](mailto:srms@nist.gov).

## REFERENCES

- [1] Beauchamp, C.R.; Camara, J.E.; Carney, J.; Choquette, S.J.; Cole, K.D.; DeRose, P.C.; Duewer, D.L.; Epstein, M.S.; Kline, M.C.; Lippa, K.A.; Lucon, E.; Molloy, J.; Nelson, M.A.; Phinney, K.W.; Polakoski, M.; Possolo, A.; Sander, L.C.; Schiel, J.E.; Sharpless, K.E.; Toman, B.; Winchester, M.R.; Windover, D.; *Metrological Tools for the Reference Materials and Reference Instruments of the NIST Material Measurement Laboratory*; NIST Special Publication (NIST SP) 260-136, 2021 edition; U.S. Government Printing Office: Washington, DC (2021); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-136-2021.pdf> (accessed Sep 2025).
- [2] Possolo, A.M; *Evaluating, Expressing, and Propagating Measurement Uncertainty for NIST Reference Materials*; NIST Special Publication (NIST SP) 260-202; U.S. Government Printing Office: Washington, DC (2020); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-202.pdf> (accessed Sep 2025).

### If you use this SRM in published work, please reference:

Cecelski C, Harris K, Goodman C, Kimes W, Liu Q, Miller W, Carney J (2021) Certification of NIST Gas Mixture Standard Reference Materials. (National Institute of Standards and Technology, Gaithersburg, MD), NIST Special Publication (SP) 260-222. <https://doi.org/10.6028/NIST.SP.260-222>

**Certificate of Analysis Revision History:** 24 September 2025 (Editorial changes; format update); 12 January 2022 (Original certification date).

*Certain commercial equipment, instruments, or materials may be identified in this Certificate of Analysis to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.*

*Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the Office of Reference Materials 100 Bureau Drive, Stop 2300, Gaithersburg, MD 20899-2300; telephone (301) 975-2200; e-mail [srminfo@nist.gov](mailto:srminfo@nist.gov); or the Internet at <https://www.nist.gov/srm>.*

\* \* \* \* \* End of Certificate of Analysis \* \* \* \* \*

## APPENDIX A

The table contains all cylinder Sample Numbers and their associated information for this lot.

Sample Number	Cylinder Number	Hydrostat Date	Fill Date
93-I-04	FF60031	April 2019	April 2019
93-I-05	FF60261	April 2019	April 2019
93-I-06	FF60242	April 2019	April 2019
93-I-07	FF60256	April 2019	April 2019
93-I-09	FF60038	April 2019	April 2019
93-I-10	FF60226	April 2019	April 2019
93-I-11	FF60042	April 2019	April 2019
93-I-12	FF60253	April 2019	April 2019
93-I-13	FF60131	April 2019	April 2019
93-I-14	FF60248	April 2019	April 2019
93-I-16	FF60039	April 2019	April 2019
93-I-17	FF60292	April 2019	April 2019
93-I-18	FF60262	April 2019	April 2019
93-I-19	FF60288	April 2019	April 2019
93-I-20	FF60282	April 2019	April 2019
93-I-21	FF60293	April 2019	April 2019
93-I-22	FF60053	April 2019	April 2019
93-I-23	FF60268	April 2019	April 2019
93-I-25	FF60034	April 2019	April 2019
93-I-26	FF60035	April 2019	April 2019
93-I-27	FF60283	April 2019	April 2019
93-I-28	FF60290	April 2019	April 2019
93-I-29	FF60249	April 2019	April 2019
93-I-30	FF60047	April 2019	April 2019
93-I-31	FF60281	April 2019	April 2019
93-I-32	FF60250	April 2019	April 2019
93-I-33	FF60094	April 2019	April 2019
93-I-34	FF60272	April 2019	April 2019
93-I-35	FF60277	April 2019	April 2019
93-I-36	FF60246	April 2019	April 2019

\* \* \* \* \* End of Appendix A \* \* \* \* \*

## APPENDIX B

**Balance Gas:** The balance gas for SRM 1662a Lot 93-I-XX is nitrogen.

**Nitrogen:** Nitrogen CAS Registry 7727-37-9.

### Trace Compounds:

**Water:** The mole fraction of water in this mixture is expressed as a maximum allowable purity level of  $< 1 \mu\text{mol/mol}$ . While the best estimate of the water mole fraction lies within the interval [0 to 1]  $\mu\text{mol/mol}$ , this interval may not include the true value. Water was detected using an electrolytic analyzer. Water CAS Registry 7732-18-5.

**Total Hydrocarbons expressed as Methane:** The mole fraction of total hydrocarbons in this mixture is expressed as a maximum allowable purity level of  $< 10 \text{ nmol/mol}$ . While the best estimate of the total hydrocarbons mole fraction lies within the interval [0 to 10]  $\text{nmol/mol}$ , this interval may not include the true value. Total hydrocarbons expressed as methane was detected using a total hydrocarbon analyzer. Methane CAS Registry 74-82-8.

\* \* \* \* \* \* \* End of Appendix B \* \* \* \* \*