NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY
U.S. DEPARTMENT OF COMMERCE

Date of Issue:
03 December 2025

Standard Reference Material® 189c
Potassium Tetroxalate Dihydrate
(pH Standard)
CERTIFICATE OF ANALYSIS

Purpose: The certified values delivered by this Standard Reference Material (SRM) are intended for calibrating electrodes for pH measuring systems. SRM 189c Potassium Tetroxalate Dihydrate, $KH_3(C_2O_4)_2\ 2H_2O$, was prepared to ensure high purity and uniformity. However, this SRM is certified ONLY as a pH standard, pH(S), not as a pure substance.

Description: A unit of SRM 189c consists of 65 g of potassium tetroxalate dihydrate.

Certified Values: A NIST certified value is a value for which NIST has the highest confidence in its accuracy in that all known or suspected sources of bias have been investigated or taken into account [1]. The certified pH(S) values provided in Table 1 correspond to $\log(1/a_H)$, where a_H is the conventional activity of the hydrogen (hydronium) ion referred to the standard state ($p^\circ = 1$ atm = 101 325 Pa) on the scale of molality. The values were derived from electromotive force (emf) measurements of cells without liquid junction by the primary measurement (Harned cell) method [2]. The measurand is the pH of the specified buffer solution.

The certified values are metrologically traceable to the International System of Units (SI) of amount-of-substance and mass and to the convention [3] used to define ionic activity, including its uncertainty [2,4]. A solution of molality 0.05 mol/kg is recommended for the calibration of pH measuring systems. Values of pH(S), $u_c(y)$, and U of this solution as a function of Celsius temperature are given below. Values of U are given at the approximate 95 % level of confidence. The coverage factor, k, at each temperature is 1.96. The corresponding degrees of freedom, ν_{eff}, of $u_c(y)$ at each temperature is >1000.

The expanded uncertainty in the certified value, U, is calculated as $U = ku_c(y)$, where $u_c(y)$ is the combined standard uncertainty calculated according to the ISO/JCGM Guide [5]. The value of $u_c(y)$ is intended to represent the combined effect of the following uncertainty components associated with the primary measurement method and material homogeneity: the standard deviation of the pH(S) values after smoothing with respect to temperature as described in reference 6 standard electrode potentials, E^o; molality of HCl, b_{HCl}, used for determining E^o; measured cell potentials; correction to the standard pressure for H_2 gas; mean activity coefficient of HCl at b_{HCl}; gas constant; temperature; Faraday constant; the molality of added NaCl; the stability of the primary measurement; material homogeneity; and the uncertainty of the conventional calculation of $\log \gamma_{Cl}$ (Bates-Guggenheim convention [3]). Current expert opinion [2,4] has assessed the uncertainty attributable to the Bates-Guggenheim convention as 0.010 pH (95 % confidence level). The value of $u_c(y)$ has been multiplied by a coverage factor, k, obtained by the Student's t-distribution for effective degrees of freedom at the given temperature and a 95 % confidence level.
Non-Certified Values: Non-certified values are provided in Appendix A.

Period of Validity: The certified values delivered by **SRM 189c** are valid within the measurement uncertainty specified until **30 May 2033**. The certified values are nullified if the material is stored or used improperly, damaged, contaminated, or otherwise modified.

Maintenance of Certified Values: NIST will monitor this SRM over the period of its validity. If substantive technical changes occur that affect the certification, NIST will issue an amended certificate through the NIST SRM website (https://www.nist.gov/srm). Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (https://www.nist.gov/srm).

Carlos A. Gonzalez, Chief
Chemical Sciences Division

Steven J. Choquette, Director
Office of Reference Materials

Certificate Revision History on Page 3

Table 1. Certified pH(S) Values, Combined Standard Uncertainties, and Expanded Uncertainties for SRM 189c

Temperature[a] (°C)	pH(S)	$u_c(y)$[b]	U
5	1.666	0.0051	0.010
10	1.667	0.0051	0.010
15	1.669	0.0051	0.010
20	1.672	0.0051	0.010
25	1.677	0.0051	0.010
30	1.682	0.0051	0.010
37	1.690	0.0051	0.010
40	1.694	0.0051	0.010
45	1.700	0.0051	0.010
50	1.707	0.0051	0.010

[a] No value is certified at 0 °C owing to solubility restrictions of this compound [7].
[b] $u_c(y)$ is the combined standard uncertainty, which includes u_c for the measurement and the standard uncertainty of the Bates-Guggenheim Convention (0.0050) [2,4].

Safety: Consult the Safety Data Sheet (SDS) for hazard information.

Storage: SRM 189c is stable when stored in its original container, with the cap tightly closed, in a dry environment and under normal laboratory temperatures.

Use: Use as received. The salt must **not** be dried in an oven, nor in a desiccator, before use.

Minimum Sample: A minimum of 4 g of SRM 189c should be used to prepare SRM 189c pH(S) buffer solutions as described below, yielding approximately 314 g of solution. Use of a smaller mass of SRM 189c may increase the uncertainty of pH(S) of the prepared solution if the material is heterogeneous at that level.

Preparation of the 0.05 mol/kg Solution: Quantities denoted by m_W and associated numerical factors in this paragraph include the effect of air buoyancy, i.e., they correspond to the balance indication in units of mass obtained in the laboratory (the "balance reading"). Weigh by difference approximately 12.1 g of SRM 189c, $m_{W,189c}$, to an accuracy of 1 mg, into a clean, dry, 1 L polyethylene bottle. Add a quantity of deionized or distilled water (resistivity greater than 17 MΩ·cm at delivery; electrical conductivity less than 0.06 µS/cm), equal to 78.4962 multiplied by $m_{W,189c}$, to an accuracy of 0.1 g. Shake until the solid has completely dissolved. Gravimetric preparation in this manner eliminates the need to weigh exactly predetermined masses of solid samples. The SRM material dissolves slowly. Shaking for 12 h or longer may be necessary to effect complete dissolution.

The corresponding apparent mass in air yielding a molality equal to 0.050 000 mol kg^{-1} is 12.7395 g $KH_3(C_2O_4)_2{\cdot}2H_2O$ per 1000.0 g H_2O. As an alternative to the above procedure, solid SRM 189c may be added in this proportion to water (resistivity greater than 17 MΩ·cm at delivery; electrical conductivity less than 0.06 µS/cm), yielding an identical solution to that obtained above.

Calibrations with SRM 189c in pH Cells with Liquid Junction: The liquid junction potential of the common pH cell displays a greater variability in solutions of pH less than 2.5 than in solutions of pH between 2.5 and 11.5 [8]. Hence, the experimental pH in such cells for SRM 189c may differ by 0.02 to 0.05 from the values of pH(S) given in this Certificate of Analysis, owing to variations in liquid junction potential. This SRM is recommended for purposes of calibration and for confirmatory purposes in cells with liquid junction with this understanding.

Stability of Prepared Solution: Solutions are stable for at least one month. For the highest accuracy, prepare fresh solutions on a weekly basis.

Source of Material: The potassium tetroxalate dihydrate was obtained from a commercial company. The homogeneity of the material was assessed in the course of the certification of this SRM.

REFERENCES

[1] Beauchamp, C.R.; Camara, J.E.; Carney, J.; Choquette, S.J.; Cole, K.D.; DeRose, P.C.; Duewer, D.L.; Epstein, M.S.; Kline, M.C.; Lippa, K.A.; Lucon, E.; Molloy, J.; Nelson, M.A.; Phinney, K.W.; Polakoski, M.; Possolo, A.; Sander, L.C.; Schiel, J.E.; Sharpless, K.E.; Toman, B.; Winchester, M.R.; Windover, D.; *Metrological Tools for the Reference Materials and Reference Instruments of the NIST Material Measurement Laboratory*; NIST Special Publication 260-136, 2021 edition; National Institute of Standards and Technology, Gaithersburg, MD (2021); available at https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-136-2021.pdf (accessed Dec 2025).

[2] Buck, R.P.; Rondinini, S., Covington, A.K.; Baucke, F.G.K.; Brett, C.M.A; Camões, M.F.; Milton, M.J.T.; Mussini; T.; Naumann, R.; Pratt, K.W; Spitzer, P.; Wilson, G.S; *Measurement of pH. Definition, Standards, and Procedures*; IUPAC Recommendations 2002; Pure Appl. Chem, Vol. 74, No 11, pp. 2169–2200 (2002); available at https://www.iupac.org/publications/pac/2002/pdf/7411x2169.pdf (accessed Dec 2025).

[3] Bates, R.G.; Guggenheim, E.A.; *A Report on the Standardization of pH and Related Terminology*; Pure Appl. Chem., Vol. 1, pp. 163–168 (1960).

[4] Baucke, F.G.K.; *New IUPAC Recommendations on the Measurement of pH – Background and Essentials*; Anal. Bioanal. Chem., Vol. 734, pp. 772–777 (2002).

[5] JCGM 100:2008; *Evaluation of Measurement Data — Guide to the Expression of Uncertainty in Measurement* (GUM 1995 with Minor Corrections); Joint Committee for Guides in Metrology (JCGM) (2008); available at https://www.bipm.org/en/committees/jc/jcgm/publications (accessed Dec 2025); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at https://www.nist.gov/pml/nist-technical-note-1297 (accessed Dec 2025).

[6] Wu, Y.C.; Koch, W.F.; Durst, R.A.; *Standard Reference Materials: Standardization of pH Measurements*; NBS Spec. Publ. 260-53; U.S. Government Printing Office: Washington, DC (1988) available at https://www.nist.gov/srm/upload/SP260-53.PDF (accessed Dec 2025).

[7] Baucke, F.G.K.; *Lower Temperature Limit of NBS (DIN) pH Standard Buffer Solution Potassium Tetroxalate*; Electrochim. Acta, Vol. 24, pp. 95–97 (1979).

[8] Bates, R.G.; Pinching, G.D.; Smith, E.R.; *pH Standards of High Acidity and High Alkalinity and the Practical Scale of pH*; J. Res. Nat. Bur. Stand., Vol. 45, pp. 418–429 (1950).

Certificate Revision History: 03 December 2025 (Change of period of validity; updated format; editorial changes); 19 July 2018 (Change of expiration date; editorial changes); 04 September 2013 (Editorial changes); 26 January 2009 (Original certificate date).

Certain commercial equipment, instruments, or materials may be identified in this Certificate of Analysis to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.

Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the Office of Reference Materials 100 Bureau Drive, Stop 2300, Gaithersburg, MD 20899-2300; telephone (301) 975-2200; e-mail srminfo@nist.gov; or the Internet at https://www.nist.gov/srm.

* * * * * * * End of Certificate of Analysis * * * * * * *

APPENDIX A

Non-Certified Values: Non-certified values are suitable for use in method development, method harmonization, and process control but do not provide metrological traceability to the International System of Units (SI) or other higher-order reference system. To attain traceability to the NIST non-certified pH(S) values for SRM 189c when traceability to the SI is not necessary, the uncertainty of the Bates-Guggenheim convention is excluded from the uncertainty calculation. The non-certified pH(S) values are identical to the certified ones, but they are listed to the number of decimal places corresponding to two significant figures for the corresponding expanded uncertainty, U_R:

$$U_R = k_R u_c \text{ (measurement)}$$

where k_R is the coverage factor corresponding to a level of confidence of 95 % for U_R.

Table A1. Non-Certified pH(S) Values and Non-Certified Uncertainty Information[(a)] for SRM 189c

Temperature[(b)] (°C)	pH(S)	u_c(measurement)	v_{eff}	k_R	U_R
5	1.6657	0.00083	40	2.02	0.0017
10	1.6668	0.00079	35	2.03	0.0016
15	1.6691	0.00079	36	2.03	0.0016
20	1.6724	0.00080	36	2.03	0.0016
25	1.6766	0.00081	38	2.03	0.0016
30	1.6816	0.0011	137	1.98	0.0022
37	1.6897	0.0011	137	1.98	0.0022
40	1.6936	0.0011	138	1.98	0.0022
45	1.7003	0.0011	135	1.98	0.0022
50	1.7074	0.0011	138	1.98	0.0022

[(a)] u_c(measurement) and U_R each include all components associated with the measurement method and assessment of material homogeneity, but do not include the uncertainty of the Bates-Guggenheim Convention (0.0050) [2,4].
[(b)] No value provided at 0 °C owing to solubility restrictions of this compound [7].

Period of Validity: The non-certified values are valid within the measurement uncertainty specified until **30 May 2033.** The value assignments are nullified if the material is stored or used improperly, damaged, contaminated, or otherwise modified.

Maintenance of Non-Certified Values: NIST will monitor this material to the end of its period of validity. If substantive technical changes occur that affect the non-certified values during this period, NIST will update this Reference Material Information Sheet. Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (https://www.nist.gov/srm).

* * * * * * * * * * End of Appendix A * * * * * * * * * *