

# Standard Reference Material<sup>®</sup> 2372a

## Human DNA Quantitation Standard

### CERTIFICATE OF ANALYSIS

**Purpose:** This Standard Reference Material (SRM) is intended for use in the value assignment of human genomic deoxyribonucleic acid (DNA) quantitation materials, primarily those used for quantitative polymerase chain reaction (qPCR). **NOTE:** See page 2 for “Usage and Privacy Agreement” regarding identifiable private information.

**Description:** A unit of SRM 2372a consists of three well-characterized human genomic DNA materials in pH 8.0 aqueous buffer. The components are derived from human buffy coat samples and labeled A, B, and C. Component A consists of genomic DNA from a single male donor. Component B consists of genomic DNA from a single female donor. Component C consists of a gravimetric mixture of genomic DNA (1 part male donor to 3 parts female donor). A unit of the SRM consists of one sterile 0.5 mL vial of each component, each vial containing approximately 50  $\mu$ L of DNA solution. Each of these vials is labeled and sealed with a color-coded screw cap.

**Certified Values:** Certified values are provided in the table below. These certified values were determined based on droplet digital polymerase chain reaction (ddPCR) assay counts of ten unique targets on eight chromosomes, dilution factors, and droplet volume measurements. A NIST certified value is a value for which NIST has the highest confidence in that all known or suspected sources of bias have been taken into account [1]. The copy number values are metrologically traceable to the natural units count 1 and ratio 1 and International System of Units (SI) derived units of volume. The DNA mass concentration values are metrologically traceable to the natural units count and ratio 1 and SI derived units of mass and volume.

**Certified Values of Number and Mass Concentration for SRM 2372a<sup>(a)</sup>**

Component	Copy Number <sup>(b)</sup> (per nL)	DNA <sup>(c)</sup> (ng/ $\mu$ L)
A (red cap)	15.1 $\pm$ 1.5	49.8 $\pm$ 5.0
B (white cap)	17.5 $\pm$ 1.8	57.8 $\pm$ 5.8
C (blue cap)	14.5 $\pm$ 1.5	47.9 $\pm$ 4.8

<sup>(a)</sup> Values are expressed as  $x \pm U_{95\%}(x)$ , where  $x$  is the certified value and  $U_{95\%}(x)$  is the expanded uncertainty of the certified value. The true value of the analyte is believed to lie within the interval  $x \pm U_{95\%}(x)$  with 95 % confidence. To propagate this uncertainty, treat the certified value as a normally distributed random variable with mean  $x$  and standard deviation  $U_{95\%}(x)/2$ .

<sup>(b)</sup> Copy number concentration values, human haploid genome equivalents (HHGE) per nanoliter, are based on ddPCR assay counts of ten unique targets on eight chromosomes, dilution factors, and droplet volume measurements.

<sup>(c)</sup> Mass concentration values, nanograms of human genomic DNA per microliter, are calculated from the copy number concentration results, the number of nucleotide base pairs (bp) in the reference HHGE, and the mean molecular weight of the sodium salts of the nucleotide monomers that comprise the DNA polymer.

**Non-Certified Values:** Non-certified values are provided in Appendix A.

**Period of Validity:** The certified values delivered by **SRM 2372a** are valid within the measurement uncertainty specified until **01 October 2033**. The certified values are nullified if the material is stored or used improperly, damaged, contaminated, or otherwise modified.

**Maintenance of Certified Values:** NIST will monitor this SRM over the period of its validity. If substantive technical changes occur that affect the certification, NIST will issue an amended certificate through the NIST SRM website (<https://www.nist.gov/srm>) and notify registered users. SRM users can register online from a link available on the NIST SRM website or fill out the user registration form that is supplied with the SRM. Registration will facilitate notification. Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (<https://www.nist.gov/srm>).

**Safety:** SRM 2372a IS INTENDED FOR RESEARCH USE. This is a human-source material. SRM 2372a is a Biosafety Level 1 material and should be handled according to applicable federal, state, and/or local regulations and according to policies and procedures of recipient's organization. The supplier has reported that each donor unit of plasma used in the preparation of this product was tested by FDA-licensed tests and found to be negative for human immunodeficiency virus (HIV), HIV-1 antigen, hepatitis B, surface antigen, hepatitis C, West Nile virus (WNV RNA) and syphilis. Additionally, each donor was tested according to FDA guidelines for *Trypanosoma cruzi* (Chagas) and found to be negative/non-reactive.

**Storage:** Until required for use, SRM 2372a should be stored in the dark between 2 °C and 8 °C. Component vials should be mixed briefly and centrifuged (without opening the vial cap) prior to removing sample aliquots for analysis. For the certified values to be applicable, materials should be withdrawn immediately after opening the vials and processed without delay. Certified values do not apply to any material remaining in recapped vials. The certification only applies to the initial use and the same results are not guaranteed if the remaining material is used at a later date.

**Usage and Privacy Agreement:** NIST does not possess any identifiable private information or any code to identify any identifiable private information related to the Material and will not provide the Purchaser with any identifiable private information of any living individual or any code to identify any identifiable private information related to the Material. Purchaser agrees not to attempt to decipher any identifiable private information from the Material or to identify the individual who is the subject of the Material. Purchaser agrees not to upload or search the genetic data into any public, private, commercial, or genealogy databases. In the event that Purchaser determines the identity of any individual related to the Material, Purchaser shall not make use of such knowledge, shall safeguard, or destroy such information, shall not disclose such information to any other party, and shall notify NIST of such discovery in accordance with the applicable laws and regulations in effect at the time of the discovery.

**Use:** For qPCR assays, value assign secondary standard DNA solutions relative to the certified mass concentrations. Calibrate the assays to SRM 2372a using one or more dilution series prepared from an SRM 2372a component. For non-gender-specific qPCR assays, all three components should be used as calibrants to elucidate potential material-assay specific interactions. Users should keep in mind that a potential for pipetting error exists; therefore, suitable care should be exercised in preparing calibration solutions.

**Source:** This SRM was developed after an appropriate human subjects research determination by NIST.

**Additional Information:** Full details on the production, analysis, and statistical evaluation of SRM 2372a are provided in NIST Special Publication 260-189, *Certification of Standard Reference Material® 2372a Human DNA Quantitation Standard*. This publication is available free of charge at <https://doi.org/10.6028/NIST.SP.260-189>.

## REFERENCES

- [1] Beauchamp, C.R.; Camara, J.E.; Carney, J.; Choquette, S.J.; Cole, K.D.; DeRose, P.C.; Diewer, D.L.; Epstein, M.S.; Kline, M.C.; Lippa, K.A.; Lucon, E.; Molloy, J.; Nelson, M.A.; Phinney, K.W.; Polakoski, M.; Possolo, A.; Sander, L.C.; Schiel, J.E.; Sharpless, K.E.; Toman, B.; Winchester, M.R.; Windover, D.; *Metrological Tools for the Reference Materials and Reference Instruments of the NIST Material Measurement Laboratory*; NIST Special Publication (NIST SP) 260-136, 2021 edition; U.S. Government Printing Office: Washington, DC (2021); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-136-2021.pdf> (accessed Dec 2024).

### **If you use this SRM in published work, please reference:**

Romsos EL, Kline MC, Diewer DL, Toman B, Farkas N (2018) Certification of Standard Reference Materials® 2372a Human DNA Quantitation. (National Institute of Standards and Technology, Gaithersburg, MD), NIST Special Publication (SP) 260-189. <https://doi.org/10.6028/NIST.SP.260-189>

<b>Certificate Revision History:</b> <b>13 December 2024</b> (Included “Usage and Privacy Agreement” regarding identifiable private information; editorial changes); <b>23 September 2024</b> (Change of period of validity; updated unit size; editorial changes); <b>18 August 2022</b> (Change of period of validity; editorial changes); <b>13 March 2018</b> (Original certificate date).
--

*Certain commercial equipment, instruments, or materials may be identified in this Certificate of Analysis to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.*

*Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the Office of Reference Materials 100 Bureau Drive, Stop 2300, Gaithersburg, MD 20899-2300; telephone (301) 975-2200; e-mail [srminfo@nist.gov](mailto:srminfo@nist.gov); or the Internet at <https://www.nist.gov/srm>.*

**\* \* \* \* \* End of Certificate of Analysis \* \* \* \* \***

# APPENDIX A

**Non-Certified Values:** The table below lists values for the number of mitochondrial genomes (mtDNA) relative to the number of nuclear haploid genomes (nDNA) present in the SRM 2372a components. Because they are based on a limited number of independent ddPCR assays, these values do not meet NIST's criteria for certification but are the best currently available estimates.

## Assay-Specific mtDNA/nDNA Ratios<sup>(a)</sup>

These values are metrologically traceable to the ddPCR measurement method and assays used.

Component	mtDNA/nDNA
A (red cap)	174 ± 4
B (white cap)	206 ± 5
C (blue cap)	279 ± 7

<sup>(a)</sup> Values are expressed as  $x \pm U_{95\%}(x)$ , where  $x$  is the best-estimate value and  $U_{95\%}(x)$  is its expanded uncertainty. The true value of the ratio is believed to lie within the interval  $x \pm U_{95\%}(x)$  with 95 % confidence. To propagate this uncertainty, treat the value as a normally distributed random variable with mean  $x$  and standard deviation  $U_{95\%}(x)/2$ .

**Maintenance of Non-Certified Values:** NIST will monitor this material to the end of its period of validity. If substantive technical changes occur that affect the non-certified values during this period, NIST will update this Appendix and notify registered users. SRM users can register online from a link available on the NIST SRM website or fill out the user registration form that is supplied with the SRM. Registration will facilitate notification. Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (<https://www.nist.gov/srm>).

\*\*\*\*\* End of Appendix A \*\*\*\*\*