



# National Institute of Standards & Technology

## Certificate

### Standard Reference Material<sup>®</sup> 4322d

#### Americium-241 Radioactivity Standard

This Standard Reference Material (SRM) consists of a solution of a standardized and certified quantity of radioactive Americium-241 in a suitably stable and homogeneous matrix. It is intended primarily for the calibration of instruments that are used to measure radioactivity and for the monitoring of radiochemical procedures. A unit of SRM 4322d consists of approximately 5 mL of a solution, whose composition is specified in Tables 1 and 2, contained in a flame-sealed borosilicate-glass ampoule [1].

The certified **Americium-241** massic activity, at a **Reference Time of 1200 EST, 01 March 2019**, is:

$$(133.7 \pm 0.4) \text{ Bq}\cdot\text{g}^{-1}.$$

A NIST certified value, as used within the context of this certificate, is a value for which NIST has the highest confidence in its uncertainty assessment. It is a “measurement result” [2] obtained directly or indirectly from a “primary reference measurement procedure” [3]. The certified value is traceable to the derived SI unit, becquerel (Bq).

Additional physical, chemical, and radiological properties for this SRM, as well as details on the standardization method, are given in Tables 1 and 2. Uncertainties for the certified quantities are expanded ( $k = 2$ ). The uncertainties are calculated according to the ISO/JCGM and NIST Guides [4,5]. Table 3 contains a specification of the components that comprise the uncertainty analysis.

**Expiration of Certification:** The certification of **SRM 4322d** is valid indefinitely, within the measurement uncertainty specified, provided that the SRM is handled and stored properly and that no evaporation or change in composition has occurred. The solution matrix, in an unopened ampoule, is homogeneous and stable within its half-life-dependent useful lifetime provided the SRM is handled in accordance with instructions given in this certificate (see “Instructions for Use and Handling”). Periodic recertification of this SRM is not required. The certification is nullified if the SRM is damaged, contaminated, or otherwise modified.

**Maintenance of Certification:** NIST will monitor this SRM over the period of its certification. If substantive technical changes occur that affect the certification before the expiration of this certificate, NIST will notify the purchaser. Registration (see attached sheet or register online) will facilitate notification.

**Radiological and chemical hazard:** Consult the Safety Data Sheet (SDS), enclosed with the SRM shipment, for radiological and chemical hazard information.

This SRM was prepared by the NIST Physical Measurement Laboratory, Radiation Physics Division, under the direction of B.E. Zimmerman, Group Leader of the Radioactivity Group. Overall technical direction and physical measurement leading to certification were provided by R. Collé and L. Laureano-Pérez of the NIST Radiation Physics Division, Radioactivity Group. Photon-emitting-impurity analyses were provided by L. Pibida of the NIST Radiation Physics Division, Radioactivity Group.

Support aspects involved in the issuance of this SRM were coordinated through the NIST Office of Reference Materials.

James M. Adams, Chief  
Radiation Physics Division

Gaithersburg, Maryland 20899  
Certificate Issue Date: 30 August 2019

Steven J. Choquette, Director  
Office of Reference Materials

Table 1. Certified Massic Activity of SRM 4322d

<b>Radionuclide</b>	<b>Americium-241</b>
<b>Reference time</b>	<b>1200 EST, 01 March 2019</b>
<b>Massic activity of the solution<sup>(a)</sup></b>	<b>133.7 Bq•g<sup>-1</sup></b>
<b>Relative expanded uncertainty (<math>k = 2</math>)<sup>(b)</sup></b>	<b>0.32 %</b>

<sup>(a)</sup> This solution is gravimetrically linked to a <sup>241</sup>Am stock solution used to prepare several previous NIST <sup>241</sup>Am standards, including a primary live-timed anticoincidence standardization in 2007, two proficiency testing standards issued for a nuclear power plant measurement assurance program and for the Integrated Consortium of Laboratory Networks, as well as to SRM 4322c. A remeasurement of SRM 4322c at the time of this work in 2019 agreed with the 2007 certified value to within 0.05 %

<sup>(b)</sup> The uncertainties on certified values are expanded uncertainties,  $U = ku_c$ . The quantity  $u_c$  is the combined standard uncertainty calculated according to the ISO/JCGM and NIST Guides [4,5]. The combined standard uncertainty is multiplied by a coverage factor of  $k = 2$  and was chosen to obtain an approximate 95 % level of confidence.

Table 2. Uncertified Information of SRM 4322d

Source description	Liquid in a flame-sealed 5 mL borosilicate-glass ampoule [1]
Solution composition	$(1.03 \pm 0.06) \text{ mol} \cdot \text{L}^{-1} \text{ HNO}_3$
Solution density	$(1.032 \pm 0.002) \text{ g} \cdot \text{mL}^{-1}$ at 22.4 °C <sup>(a)</sup>
Solution mass	$(5.14 \pm 0.01) \text{ g}^{(a)}$
Alpha-particle-emitting impurities	None detected <sup>(b)</sup>
Photon-emitting impurities	None detected <sup>(c)</sup>
Half-lives used [6]	<sup>241</sup> Am: $(432.6 \pm 0.6) \text{ a}^{(d)}$
Calibration methods (and instruments)	The certified massic activity for <sup>241</sup> Am was obtained by $4\pi\alpha$ liquid scintillation (LS) counting using two different LS measurement systems.

<sup>(a)</sup> The stated uncertainty is two times the standard uncertainty [5].

<sup>(b)</sup> The estimated limits of detection for alpha-emitting impurities were:

0.006 s<sup>-1</sup>•g<sup>-1</sup> for energies between 3.5 MeV and 5.0 MeV, and

0.002 s<sup>-1</sup>•g<sup>-1</sup> for energies between 5.5 MeV and 12 MeV.

<sup>(c)</sup> The estimated limits of detection for photon-emitting impurities, expressed as massic photon emission rates (numbers of photons per second per gram), are:

0.02 s<sup>-1</sup>•g<sup>-1</sup> for energies between 30 keV and 40 keV,

0.05 s<sup>-1</sup>•g<sup>-1</sup> for energies between 45 keV and 60 keV,

0.01 s<sup>-1</sup>•g<sup>-1</sup> for energies between 65 keV and 500 keV, and

0.02 s<sup>-1</sup>•g<sup>-1</sup> for energies between 510 keV and 2000 keV,

provided that the photons are separated in energy by 4 keV or more from photons emitted in the decay of <sup>241</sup>Am.

<sup>(d)</sup> The stated uncertainty is the standard uncertainty. See reference [6].

Table 2. Uncertainty Evaluation for the Massic Activity of SRM 4322d

Uncertainty component		Assessment Type <sup>(a)</sup>	Relative standard uncertainty contribution on massic activity of <sup>241</sup> Am (%)
1	LS measurement precision: Relative standard deviation of the mean on the great-grand mean for 4 LS measurement trials, considering all within-trial and between-trial components of variance. Each of the four grand mean values was based on 6 replicate measurements on each of either 5 or 4 LS counting sources (prepared with two different LS cocktail compositions) in two different LS counters. The typical within-trial relative standard deviation of the mean (considering the variations for the between 6 measurements and the between 4 or 5 sources) for each trial was 0.13 %. The between-trial relative standard deviation across the 4 trials was 0.035 %. The 108 individual data values fit a Normal distribution. The relative standard deviation of the mean on the set of 108 values was 0.021 %, with a 95 % confidence interval about the mean of $\pm 0.041$ %.	A	0.081
2	Background; LS measurement variability and cocktail composition stability effects; wholly embodied in component 1.	A	--
3	LS counters dependencies; wholly embodied in components 1 & 2.	A	--
4	Live time determinations for LS counting time intervals.	B	0.07
5	Aliquant mass determinations by gravimetric measurements for preparation of counting sources; includes mass measurement precision partially embodied in component 1.	B	0.05
6	LS detection inefficiency: includes wall effect; partially embodied in component 1.	B	0.01
7	<sup>241</sup> Am decay corrections for half-life uncertainty of 0.14 %.	B	$< 10^{-7}$
8	Potential undetected alpha- and photon-emitting impurities.	B	0.1
<b>Relative combined standard uncertainty</b>			<b>0.16</b>
<b>Relative expanded uncertainty (<math>k = 2</math>)</b>			<b>0.32</b>

<sup>(a)</sup> Letter A denotes evaluation by statistical methods; Letter B denotes evaluation by other methods.

## INSTRUCTIONS FOR USE AND HANDLING

**Storage:** SRM 4322d should be stored and used at a temperature between 5 °C and 65 °C. The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material.

**Handling:** If the ampoule is transported, it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) because of both the radioactivity and the strong acid. The ampoule should be opened only by persons qualified to handle both radioactive material and alkaline and/or acidic solutions. Appropriate shielding and/or distance should be used to minimize personnel exposure. Refer to the SDS for further information.

## REFERENCES

- [1] NIST Physical Measurement Laboratory; *Storage and Handling of Radioactive Standard Reference Materials, Ampoule Specifications and Opening Procedure*; available at <https://www.nist.gov/pml/radiation-physics/ampoule-specifications-and-opening-procedure> (accessed Aug 2019). Note: This SRM is contained in a generic borosilicate-glass ampoule and not in the standard NIST ampoule. This is **not** a pre-scored ampoule, the gold color band is only for identification.
- [2] JCGM 200:2012; *International Vocabulary of Metrology - Basic and General Concepts and Associated Terms (VIM)*; (2008 version with Minor Corrections), 3rd edition; Joint Committee for Guides in Metrology (JCGM): BIPM, Sevres Cedex, France; p. 19 (2012); available at [https://www.bipm.org/utis/common/documents/jcgm/JCGM\\_200\\_2012.pdf](https://www.bipm.org/utis/common/documents/jcgm/JCGM_200_2012.pdf) (accessed Aug 2019).
- [3] JCGM 200:2012; *International Vocabulary of Metrology - Basic and General Concepts and Associated Terms (VIM)*; (2008 version with Minor Corrections), 3rd edition; JCGM: BIPM, Sevres Cedex, France; p. 18 (2012); available at [https://www.bipm.org/utis/common/documents/jcgm/JCGM\\_200\\_2012.pdf](https://www.bipm.org/utis/common/documents/jcgm/JCGM_200_2012.pdf) (accessed Aug 2019).
- [4] JCGM 100:2008; *Guide to the Expression of Uncertainty in Measurement*; (GUM 1995 with Minor Corrections), JCGM: BIPM, Sevres Cedex, France (2008); available at [https://www.bipm.org/utis/common/documents/jcgm/JCGM\\_100\\_2008\\_E.pdf](https://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008_E.pdf) (accessed Aug 2019).
- [5] Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297, U.S. Government Printing Office: Washington, DC (1994); available at <https://www.nist.gov/pml/nist-technical-note-1297> (accessed Aug 2019).
- [6] Chechev, V.P. and N.K. Kuzmenko ; *LNE-LNHB/CEA Table of Radionuclides*, <sup>241</sup>Am; (July 2010); available at [http://www.lnhb.fr/nuclides/Am-241\\_tables.pdf](http://www.lnhb.fr/nuclides/Am-241_tables.pdf) (accessed Aug 2019).

*Users of this SRM should ensure that the certificate in their possession is current. This can be accomplished by contacting the SRM Program at: telephone (301) 975-2200; fax (301) 948-3730; e-mail [srminfo@nist.gov](mailto:srminfo@nist.gov); or via the Internet at <https://www.nist.gov/srm>.*