National Institute of Standards and Technology
U.S. Department of Commerce

Date of Issue:
10 January 2023

Reference Material 8547
IAEA-N-1
(Nitrogen Isotopes in Ammonium Sulfate)
REFERENCE MATERIAL INFORMATION SHEET

Purpose: This Reference Material (RM) is intended for use in developing and validating methods for measuring relative differences in nitrogen (N) stable isotope ratios [1,2]. While atmospheric nitrogen officially defines the zero point for the nitrogen δ-scale, RM 8547 is recommended as the scale anchor point for samples with a preparation requiring combustion. The equivalent name used by the International Atomic Energy Agency (IAEA) and the U.S. Geological Survey (USGS) for this RM is IAEA-N-1.

Description: A unit of RM 8547 consists of one bottle containing approximately 0.5 g of ammonium sulfate salt ($[NH_4]_2SO_4$).

Non-Certified Value: Although not certified, the assigned isotope-delta value for this RM, provided in Table 1 below, is at present the best estimate of the true value.

Table 1. Non-Certified Value for Nitrogen Stable Isotope of RM 8547 (IAEA-N-1)

NIST RM Number	Name	Non-Certified Value $\delta^{15}N_{AIR}$	Combined Uncertainty $\delta^{15}N_{AIR}$[(a)]	Expanded Uncertainty $\delta^{15}N_{AIR}$[(a)]
8547	IAEA-N-1	+0.43 ‰	0.07 ‰	0.14 ‰

[(a)] RM 8547 is given with a combined standard uncertainty in addition to an expanded uncertainty value, $k = 2$, for the assigned value. The expanded uncertainty is equal to $U = ku_c$, where u_c is the combined standard uncertainty and k is the coverage factor, as defined in the ISO/JCGM Guide [3]. The non-certified value and uncertainty are given in units of per mil (‰), which is equivalent to per thousand.

Metrological Traceability: RM 8547 is a secondary reference material with a known nitrogen stable isotope value. The nitrogen stable isotope value assigned for RM 8547 is traceable to atmospheric nitrogen (AIR). While atmospheric nitrogen officially defines the zero delta point for the nitrogen δ-scale [4,5], RM 8547 is recommended as the scale anchor point for samples with a preparation requiring combustion [2].

Stable isotope values for nitrogen are not traceable to the International System of Units (SI) or other higher-order reference system [3,6]. A *Traceability Exception* has been approved by the Bureau International des Poids et Mesures (BIPM) International Committee for Weights and Measures (CIPM), which states non-SI traceable isotope values "should be made traceable to materials recognized as International Standards" [6,7].

Period of Validity: The non-certified value is valid within the measurement uncertainty specified until **31 December 2032.** The value assignments are nullified if the material is stored or used improperly, damaged, contaminated, or otherwise modified.

Maintenance of Non-Certified Value: NIST will monitor this material to the end of its period of validity. If substantive technical changes occur that affect the non-certified value during this period, NIST will update this Reference Material Information Sheet and notify registered users. RM users can register online from a link available on the NIST SRM website or fill out the user registration form that is supplied with the RM. Registration will facilitate notification. Before making use the value delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (https://www.nist.gov/srm).

Carlos A. Gonzalez, Chief
Chemical Sciences Division

Steven J. Choquette, Director
Office of Reference Materials

Information Sheet Revision History on Page 2

Safety: Consult the Safety Data Sheet (SDS) for hazard information.

Storage: This RM is stable at normal room temperatures and should be stored in the original container at ambient temperature (20 ºC to 30 ºC) to minimize the potential for contamination.

Additional Information: The distribution of RM 8547 is limited to one unit per customer per year. Users are encouraged to prepare their own standards for daily use and calibrate those standards against international reference materials. Preparation, analysis, and reporting information can be found in Appendix A.

REFERENCES

[1] ISO/IEC, International Vocabulary of Metrology (VIM); *Basic and General Concepts and Associated Terms;* Guide 99-12:2007; Joint Committee for Guides in Metrology, International Organization for Standardization, Geneva, pp. 1–127 (2008); available at https://www.bipm.org/en/committees/jc/jcgm/publications (accessed Jan 2023).

[2] Brand, W.A.; Coplen, T.B.; Vogl, J.; Rosner, M.; Prohaska, T.; *Assessment of International Reference Materials for Isotope Ratio Analysis (IUPAC Technical Report)*; Pure and Appl. Chem., Vol. 86, pp. 425–467 (2014).

[3] JCGM 100:2008; *Evaluation of Measurement Data — Guide to the Expression of Uncertainty in Measurement* (GUM 1995 with Minor Corrections); Joint Committee for Guides in Metrology (2008); available at https://www.bipm.org/en/committees/jc/jcgm/publications (accessed Jan 2023).

[4] Mariotti, A.; *Atmospheric Nitrogen is a Reliable Standard for Natural ^{15}N Abundance Measurements;* Nature, Vol. 303, pp. 685–687 (1983).

[5] Junk, G.; Svec, H.J.; *The Absolute Abundance of the Nitrogen Isotopes in the Atmosphere and Compressed Gas from Various Sources;* Geochimica Cosmochimica Acta, Vol. 14, pp. 234–243 (1958).

[6] Beauchamp, C.R.; Camara, J.E.; Carney, J.; Choquette, S.J.; Cole, K.D.; DeRose, P.C.; Duewer, D.L.; Epstein, M.S.; Kline, M.C.; Lippa, K.A.; Lucon, E.; Molloy, J.; Nelson, M.A.; Phinney, K.W.; Polakoski, M.; Possolo, A.; Sander, L.C.; Schiel, J.E.; Sharpless, K.E.; Toman, B.; Winchester, M.R.; Windover, D.; *Metrological Tools for the Reference Materials and Reference Instruments of the NIST Material Measurement Laboratory*; NIST Special Publication (NIST SP) 260-136, 2021 edition; U.S. Government Printing Office: Washington, DC (2021); available at https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-136-2021.pdf (accessed Jan 2023).

[7] BIPM Traceability Exception: Delta Value Isotope Ratio Measurements (2015); available at https://www.bipm.org/documents/20126/50116808/%5BQM%5D+Delta+value+isotope+ratio+measurements.pdf/4fe4f00a-7f3c-9683-59f2-e49a11db074a?version=1.4&download=true (accessed Jan 2023Note that this document is a summary of Decision CIPM/104-26 from the International Committee for Weights and Measures (CIPM); *Proceedings of Session 1 of the 104th meeting: Executive Summary*; 9-10 March 2015, p. 34; available at https://www.bipm.org/utils/en/pdf/CIPM/CIPM2015-I-EN.pdf (accessed Jan 2023).

[8] Gonfiantini, R.; *Standards for Stable Isotope Measurements in Natural Compounds;* Nature, Vol. 271, pp. 534–536 (1978).

[9] Gonfiantini, R.; *Report on an Advisory Group Meeting on Stable Isotope Reference Samples for Geochemical and Hydrochemical Investigations*; In Proceedings of a meeting: Vienna, Sept. 1983; IAEA (1984).

[10] Böhlke, J.K.; Coplen, T.B.; *Interlaboratory Comparison of Reference Materials for Nitrogen-Isotope-Ratio Measurements*, *in Reference and Intercomparison Materials for Stable Isotopes of Light Elements;* In Proceedings of a consultants meeting: Vienna, Dec. 1993; IAEA-TECDOC-825*;* IAEA , pp. 51-66 (1995) available at http://www-pub.iaea.org/MTCD/publications/PDF/te_825_prn.pdf (accessed Jan 2023).

[11] Böhlke, J.K.; Mroczkowski, S.J.; Coplen, T.B.; *Oxygen Isotopes in Nitrate: New Reference Materials for ^{18}O:^{17}O:^{16}O Measurements and Observations on Nitrate-Water Equilibration;* Rapid Communications in Mass Spectrometry, Vol. 17, pp. 1835–1846 (2003).

[12] Böhlke, J.K.; Gwinn, C.J.; Coplen, T.B.; *New Reference Materials for Nitrogen-Isotope-Ratio Measurements;* Geostands Newsletter, Vol. 17, pp. 159–164 (1993).

[13] Coplen, T. B.; *Guidelines and Recommended Terms for Expression of Stable-Isotope-Ratio and Gas-Ratio Measurement Results*, Rapid Communications in Mass Spectrometry, Vol. 25, pp. 2538–2560 (2011).

[14] Coplen, T.B.; Krouse, H.R.; Böhlke, J.K.; *Reporting of Nitrogen-Isotope Abundances;* Pure and Applied Chemistry, Vol. 64. pp. 907–908 (1992).

[15] DeBievre, P.; Valkiers, S.; Peiser, H.S.; Taylor, P.D.P.; Hansen, P.; *Mass Spectrometric Methods for Determining Isotopic Composition and Molar Mass Traceable to the SI, Exemplified by Improved Values for Nitrogen;* Metrologia, Vol. 33, pp. 447–455 (1996).

Information Sheet Revision History: 10 January 2023 (Revision of uncertainty values; expiration date extension; corrected amount contained in each unit; updated format; editorial changes); 30 January 2013 (Uncertainty updated to an expanded uncertainty for $\delta^{15}N_{AIR}$; expiration date assigned; editorial changes); 03 February 1993 (Updated value for RM 8549; added RM 8558 to report); 22 June 1992 (Original report issue date).

Certain commercial equipment, instruments, or materials may be identified in this Reference Material Information Sheet to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.

Users of this RM should ensure that the Reference Material Information Sheet in their possession is current. This can be accomplished by contacting the Office of Reference Materials 100 Bureau Drive, Stop 2300, Gaithersburg, MD 20899-2300; telephone (301) 975-2200; e-mail srminfo@nist.gov; or the Internet at https://www.nist.gov/srm.

* * * * * * * * * End of Reference Material Information Sheet * * * * * * * * *

APPENDIX A

PREPARATION AND ANALYSIS

Technical aspects involved in the issuance of this RM were coordinated through the NIST Chemical Sciences Division by R.A. Kraft.

Support aspects involved in the issuance of this RM were coordinated through the NIST Office of Reference Materials.

Sample Preparation: RM 8547 (IAEA-N-1) was prepared as a relatively coarse dry salt by E. Salati (CENA, Piracicaba, Sao Paulo, Brazil) between 1978 and 1983 [8,9].

Analytical Methods: The non-certified value for δ^{15}N in RM 8547 (IAEA-N-1) was derived from an inter-laboratory comparison [10]. The δ^{15}N values were measured by mass spectrometry on N_2 gas that was quantitatively produced off-line using variants of a buffered sample combustion method coupled with additional purification steps. The measured results were then normalized to using a consensus value of +180 ‰ for RM 8558 (USGS32) [10]. Working laboratory standards with a range of nitrogen stable isotope-ratios can be produced by the methods described by Böhlke *et al.* [11,12].

Homogeneity: There is no evidence of isotopic heterogeneity in this RM for sample sizes in the range of 10 µmol to 100 µmol of nitrogen [10].

REPORTING

Terminology: The terminology used here is based on the guidance given by IUPAC for isotope terminology, where stable isotope-number ratio refers to the number of atoms of one isotope relative to the number of atoms of a second isotope in the same system [2]. This is often abbreviated to stable isotope ratio. Isotope-delta value refers to the stable isotope-number ratio of a measured sample relative to the stable isotope-number ratio of a reference material (see example below). Isotope-amount ratio is numerically the same as isotope-number ratio but refers specifically to the amount (moles) of an isotope relative to the amount (moles) of another isotope in the same system [13].

Isotope-delta Values: The nitrogen stable isotope-delta value of a measured sample reported on the AIR scale are defined as the difference in measured isotope-number ratio of nitrogen in a sample relative to the isotope-number ratio of nitrogen in AIR:

$$\delta^{15}N = \frac{\left[\frac{N_{sample}(^{15}N)}{N_{sample}(^{14}N)}\right] - \left[\frac{N_{AIR}(^{15}N)}{N_{AIR}(^{14}N)}\right]}{\left[\frac{N_{AIR}(^{15}N)}{N_{AIR}(^{14}N)}\right]}$$

Normalization: By convention AIR is the zero point of the nitrogen stable isotope δ-scale. AIR refers to N_2 of tropospheric air [4,5], for which $R(^{15}N/^{14}N)_{AIR} = 0.003677$ [14,15]. The δ-value for RM 8558 (USGS32) is also defined by convention and has a δ^{15}N value of +180 ‰. A formula for normalizing nitrogen isotope measurement results using two laboratory standards LS1 (AIR) and LS2 (USGS32) can be expressed as:

$$\delta^{15}N_{sample,cal} = \delta^{15}N_{LS1,cal} + \left(\delta^{15}N_{sample,WS} - \delta^{15}N_{LS1,WS}\right) \times f$$

where the normalization factor f is:

$$f = \frac{\left(\delta^{15}N_{LS2,cal} - \delta^{15}N_{LS1,cal}\right)}{\left(\delta^{15}N_{LS2,WS} - \delta^{15}N_{LS1,WS}\right)}$$

where *WS* denotes measurements made versus a transfer gas (working standard),
cal denotes calibrated measurements made versus the AIR scale, and
$\delta^{15}N_{LS1,cal}$ and $\delta^{15}N_{LS2,cal}$ are the conventionally fixed δ^{15}N values for AIR and RM 8558 (USGS32), or those of calibrated laboratory working standards.

The δ-definition above assumes $f = 1$, and does not account for scale compression.

The following recommendations are provided for reporting the relative difference of nitrogen stable isotope-number ratios using the δ-notation modified from Coplen [13], it is recommended that:

- $\delta^{15}N$ values should be reported with respect to air (atmospheric nitrogen gas) and normalized to RM 8558 (USGS32) [14].
- Authors should report δ values of internationally distributed (secondary) isotopic reference materials that were assumed for normalization of data for samples of similar chemical composition, as appropriate for the measurement method. In this manner, measurement results can be adjusted in the future as analytical methods improve and consensus values of internationally distributed isotopic reference materials change.

* * * * * * * * * * * End of Appendix A * * * * * * * * * * *