

# Reference Material 8690

## Per- and Polyfluoroalkyl Substances (PFAS) in Aqueous Film-Forming Foams (AFFF) Formulation I

### REFERENCE MATERIAL INFORMATION SHEET

**Purpose:** This Reference Material (RM) is intended for the identification and quantification of per- and polyfluoroalkyl substances (PFAS) in aqueous film-forming foams (AFFFs).

**Description:** A unit of RM 8690 consists of four 2 mL ampoules. Each ampoule contains approximately 1.2 mL of methanol solution.

**Non-Certified Values:** Non-certified values are suitable for use in method development, method harmonization, and process control but do not provide metrological traceability to the International System of Units (SI) or other higher-order reference system [1]. Non-certified mass fraction values are provided in Table 1 below.

Table 1. Non-Certified Values for RM 8690

Measurand	Common Acronym	Mass Fraction (as-received) <sup>(a)</sup> ( $\mu\text{g/g}$ )	
Perfluorobutanoic Acid	PFBA	0.104	$\pm 0.016$
Perfluorohexanoic Acid	PFHxA	0.295	$\pm 0.026$
Perfluoroheptanoic Acid	PFHpA	0.0918	$\pm 0.0370$
Perfluorooctanoic Acid	PFOA	0.317	$\pm 0.078$
Perfluoropropanesulfonic Acid	PFPrS	0.302	$\pm 0.016$
Perfluorobutanesulfonic Acid	PFBS	0.473	$\pm 0.032$
Perfluoropentanesulfonic Acid	PFPeS	0.409	$\pm 0.085$
Perfluorohexanesulfonic Acid	PFHxS	2.64	$\pm 0.12$
Perfluoroheptanesulfonic Acid	PFHpS	0.315	$\pm 0.225$
Perfluorooctanesulfonic Acid	PFOS	19.4	$\pm 0.4$
Perfluorobutane Sulfonamide	PFBSA	0.0356	$\pm 0.0065$
Perfluorohexane Sulfonamide	PFHxSA	0.035	$\pm 0.013$
6:2 Fluorotelomer Sulfonic Acid	6:2 FTS	0.222	$\pm 0.066$
Perfluorohexane Sulfonamido Amine	N-AP-FHxSA	9.12	$\pm 1.71$
N-[3-(perfluoro-1-hexanesulfonamido)propan-1-yl]- N,N,N-trimethylammonium	N-TAmP-FHxSA	0.293	$\pm 0.023$
6:2 Fluorotelomer Sulfonamide Propyl Betaine	6:2 FTAB	2.50	$\pm 0.5$

<sup>(a)</sup> Values are expressed as  $x \pm U(x)$ , where  $x$  is the value and  $U(x)$  is the expanded uncertainty of the value. The expanded uncertainty is calculated as  $U(x) = k u_c$  where  $u_c$  is the standard uncertainty and  $k$  is the coverage factor. For values shown,  $k = 2$ . While the best estimate of the mass fraction for the measurand lies within the interval  $x \pm U(x)$ , this interval may not include the true value.

**Period of Validity:** The non-certified values are valid within the measurement uncertainty specified until **31 January 2033**. The value assignments are nullified if the material is stored or used improperly, damaged, contaminated, or otherwise modified.

Carlos A. Gonzalez, Chief  
Chemical Sciences Division

Steven J. Choquette, Director  
Office of Reference Materials

**Maintenance of Non-Certified Values:** NIST will monitor this material to the end of its period of validity. If substantive technical changes occur that affect the non-certified values during this period, NIST will update this Reference Material Information Sheet and notify registered users. RM users can register online from a link available on the NIST SRM website or fill out the user registration form that is supplied with the RM. Registration will facilitate notification. Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (<https://www.nist.gov/srm>).

**Safety:** This material contains per- and polyfluoroalkyl substances, many of which have been reported to have toxic and/or carcinogenic properties and should be handled with care. Use proper disposal methods. Please consult the Safety Data Sheet provided with this material.

**Storage:** The original unopened ampoules of RM 8690 should be stored at room temperature ( $20\text{ }^{\circ}\text{C} \pm 5\text{ }^{\circ}\text{C}$ ). The value assignment does not apply to contents of previously opened ampoules as the stability of the measurands in opened ampoules has not been investigated.

**Use:** Open ampoules carefully to prevent contamination or injury. The ampoules are pre-scored and should **NOT** be opened using a file. Sample aliquots for analysis should be withdrawn at  $20\text{ }^{\circ}\text{C}$  to  $25\text{ }^{\circ}\text{C}$  **immediately** after opening and should be processed without delay for the values in Table 1 to be valid within the stated uncertainties.

**Additional Information:** The non-certified value is a weighted mean of average mass fractions, with one average from gravimetric preparation and two averages from chromatographic measures [2,3]. The expanded uncertainty was determined using the Horn-Horn-Duncan (HHD) method [4] for variances and with the bootstrap method [5], which is consistent with the ISO/JCGM Guides [6,7]. The effective coverage factor,  $k$ , is 2.

Full details on the production and evaluation of RM 8690 are provided free of charge in reference 8.

## REFERENCES

- [1] Beauchamp, C.R.; Camara, J.E.; Carney, J.; Choquette, S.J.; Cole, K.D.; DeRose, P.C.; Duewer, D.L.; Epstein, M.S.; Kline, M.C.; Lippa, K.A.; Lucon, E.; Phinney, K.W.; Polakoski, M.; Possolo, A.; Sharpless, K.E.; Sieber, J.R.; Toman, B.; Winchester, M.R.; Windover, D.; *Metrological Tools for the Reference Materials and Reference Instruments of the NIST Material Measurement Laboratory*; NIST Special Publication 260-136; U.S. Government Printing Office: Washington, DC (2020); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-136-2020.pdf> (accessed Mar 2023).
- [2] Dersimonian, R.; Laird, N.; *Meta-Analysis in Clinical Trials*; Control Clin. Trials, Vol. 7, pp. 177–188 (1986).
- [3] Rukhin, A.L.; *Weighted Means Statistics in Interlaboratory Studies*; Metrologia, Vol. 46, pp. 323–331 (2009).
- [4] Horn, R.A.; Horn, S.A.; Duncan, D.B.; *Estimating Heteroscedastic Variance in Linear Models*; J. American Statistical Association, Vol. 70, pp. 380–385 (1975).
- [5] Efron, B.; Tibshirani, R.J.; *An Introduction to the Bootstrap*; Chapman & Hall (1993).
- [6] JCGM 100:2008; Evaluation of Measurement Data — Guide to the Expression of Uncertainty in Measurement (GUM 1995 with Minor Corrections); Joint Committee for Guides in Metrology (2008); available at <https://www.bipm.org/en/committees/jc/jcgm/publications> (accessed Mar 2023); see also Taylor, B.N.; Kuyatt, C.E.; Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results; NIST Technical Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at <https://www.nist.gov/pml/nist-technical-note-1297> (accessed Mar 2023).
- [7] JCGM 101:2008; Evaluation of Measurement Data – Supplement 1 to the Guide to Expression of Uncertainty in Measurement; Propagation of Distributions Using a Monte Carlo Method; Joint Committee for Guides in Metrology (BIPM, IEC, IFCC, ILAC, ISO, IUPAC, IUPAP and OIML), International Bureau of Weights and Measures (BIPM), Sèvres, France (2008); available at <https://www.bipm.org/en/committees/jc/jcgm/publications> (accessed Mar 2023).
- [8] Reiner J.L.; Place B.J.; Heckert N.A.; Peter K.T.; Rodowa A.E.; *Characterization of Reference Materials® 8690 to 8693 Per- and Polyfluoroalkyl Substances (PFAS) in Four Formulations of Aqueous Film-Forming Foams (AFFF)*. (National Institute of Standards and Technology, Gaithersburg, MD), NIST Special Publication (SP) 260-234. (2023); available at <https://doi.org/10.6028/NIST.SP.260-234> (accessed Mar 2023).

**If you use this RM in published work, please reference:**

Reiner JL, Place BJ, Heckert NA, Peter KT, Rodowa AE (2023) Characterization of Reference Materials 8690 to 8693 Per- and Polyfluoroalkyl Substances (PFAS) in Four Formulations of Aqueous Film-Forming Foams (AFFF). (National Institute of Standards and Technology, Gaithersburg, MD), NIST Special Publication (SP) 260-234. <https://doi.org/10.6028/NIST.SP.260-234>

*Certain commercial equipment, instruments, or materials may be identified in this Reference Material Information Sheet to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.*

*Users of this RM should ensure that the Reference Material Information Sheet in their possession is current. This can be accomplished by contacting the Office of Reference Materials 100 Bureau Drive, Stop 2300, Gaithersburg, MD 20899-2300; telephone (301) 975-2200; e-mail [srminfo@nist.gov](mailto:srminfo@nist.gov); or the Internet at <https://www.nist.gov/srm>.*

\* \* \* \* \* End of Reference Material Information Sheet \* \* \* \* \*