

Standard Reference Material[®] 1684b
Nitric Oxide in Nitrogen
(Nominal Amount-of-Substance Fraction 100 $\mu\text{mol/mol}$)
Lot 44-T-XX

CERTIFICATE OF ANALYSIS

Purpose: The certified value delivered by this Standard Reference Material (SRM) is intended for use in producing metrologically traceable secondary standards for the calibration of instruments used for nitric oxide determinations.

Description: This SRM is a primary gas mixture supplied in a DOT 3AL-specification aluminum (6061 alloy) cylinder with a water volume of 6 L. Mixtures are shipped with a nominal pressure exceeding 12.4 MPa (1800 psig), which provides the user with 0.73 m³ (25.8 ft³) of usable mixture. The cylinder is the property of the purchaser and is equipped with a CGA-660 stainless-steel valve, which is the recommended outlet for this nitric oxide mixture.

Certified Value: A NIST certified value is a value for which NIST has the highest confidence in that all known or suspected sources of bias and imprecision have been accounted [1]. This SRM mixture has been certified for nitric oxide amount-of-substance fraction (mole fraction, sometimes termed “molar concentration”) [2]. This certified value is traceable to the International System of Units (SI) through the gravimetric primary standards and procedures used in the preparation of this mixture. The certified value given below applies to all cylinders in this lot.

Nitric Oxide Mole Fraction: 99.75 $\mu\text{mol/mol}$ \pm 1.20 $\mu\text{mol/mol}$ ^(a)

^(a) The certified value is expressed as $x \pm U_{95\%}(x)$, where x is the value and $U_{95\%}(x)$ is the expanded uncertainty of the value. The true value of the analyte lies within the interval $x \pm U_{95\%}(x)$ with 95 % confidence. The uncertainty is expressed as an expanded uncertainty $U = k u_c$ with u_c determined by experiment and a coverage factor $k = 2$. For guidance in propagating this uncertainty, see reference 3.

Additional Information: Appendix A contains a table of all cylinder Sample Numbers and their associated information for this lot.

Period of Validity: The certified value delivered by **SRM 1684b, Lot No. 44-T-XX** is valid within the measurement uncertainty specified until **21 March 2029**, provided the SRM is handled and stored in accordance with the instructions given in this certificate (see “Storage” and “Cylinder and Gas Handling Information”). The certification is nullified if the SRM is damaged, contaminated or otherwise modified, or if the internal pressure drops below 0.7 MPa (100 psig).

Maintenance of Certified Value: NIST will monitor this SRM over the period of its validity. If substantive technical changes occur that affect the certification, NIST will issue an amended certificate through the NIST SRM website (<https://www.nist.gov/srm>). Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (<https://www.nist.gov/srm>).

Marcela Najarro, Acting Chief
Chemical Sciences Division
Certificate Revision History on Page 2

Steven J. Choquette, Director
Office of Reference Materials

Safety: Consult the Safety Data Sheet (SDS) for hazard information.

Storage: This SRM should be stored under normal laboratory conditions within the temperature range of 15 °C to 30 °C.

Cylinder and Gas Handling Information: NIST recommends the use of a high purity, two stage pressure regulator with a stainless-steel diaphragm and CGA-660 outlet to safely reduce the pressure and to deliver this mixture to the instrument. The regulator should be purged to prevent accidental contamination of the SRM by repeatedly (minimum three times) opening the valve and pressurizing the regulator, then closing the valve and releasing the pressure safely into a vent line.

Mixture Preparation: The gas mixtures comprising this SRM lot were prepared in accordance with NIST technical specifications by a commercial specialty gas vendor under contract to NIST. The specifications stipulate that each SRM mixture be identical within the lot in nitric oxide mole fraction and stable with time.

Analytical Method: Analyses of the nitric oxide mole fraction for this lot of cylinders were conducted by comparison to a representative cylinder chosen from the lot, designated as the SRM lot standard (LS). The LS was compared to NIST primary gravimetric standards using a non-dispersive ultraviolet (NDUV) spectrometry. Each of the nitric oxide mixtures that comprise this SRM lot was then compared to the LS using NDUV. Within the precision of the NIST measurements, all the cylinders comprising this SRM lot have identical nitric oxide mole fractions.

CAS Registry Numbers: This SRM is certified for nitric oxide in nitrogen. The relevant CAS Registry numbers for these components are: nitric oxide CAS Registry 10102-43-9; nitrogen (balance gas) CAS Registry 7727-37-9.

NOTICE TO USERS

NIST strives to maintain the SRM inventory supply, but NIST cannot guarantee the continued or continuous supply of any specific SRM. Accordingly, NIST encourages the use of this SRM as a primary benchmark for the quality and accuracy of the user's in house reference materials and working standards. As such, the SRM should be used to validate the more routinely used reference materials in a laboratory. Comparisons between the SRM and in house reference materials or working measurement standards should take place at intervals appropriate to the conservation of the SRM and the stability of relevant in house materials. For further guidance on how this approach can be implemented, contact NIST by email at srms@nist.gov.

REFERENCES

- [1] Beauchamp, C.R.; Camara, J.E.; Carney, J.; Choquette, S.J.; Cole, K.D.; DeRose, P.C.; Duerwer, D.L.; Epstein, M.S.; Kline, M.C.; Lippa, K.A.; Lucon, E.; Molloy, J.; Nelson, M.A.; Phinney, K.W.; Polakoski, M.; Possolo, A.; Sander, L.C.; Schiel, J.E.; Sharpless, K.E.; Toman, B.; Winchester, M.R.; Windover, D.; *Metrological Tools for the Reference Materials and Reference Instruments of the NIST Material Measurement Laboratory*; NIST Special Publication (NIST SP) 260-136, 2021 edition; U.S. Government Printing Office: Washington, DC (2021); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-136-2021.pdf> (accessed May 2026).
- [2] Thompson, A.; Taylor, B.N.; *Guide for the Use of the International System of Unit (SI)*; NIST Special Publication (NIST SP) 811; U.S. Government Printing Office: Washington, DC (2008); available at <https://www.nist.gov/pml/special-publication-811> (accessed May 2026).
- [3] Possolo, A.M.; *Evaluating, Expressing, and Propagating Measurement Uncertainty for NIST Reference Materials*; NIST Special Publication (NIST SP) 260-202; U.S. Government Printing Office: Washington, DC (2020); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-202.pdf> (accessed May 2026).

If you use this SRM in published work, please reference:

Cecelski C, Harris K, Goodman C, Kimes W, Liu Q, Miller W, Carney J (2021) Certification of NIST Gas Mixture Standard Reference Materials. (National Institute of Standards and Technology, Gaithersburg, MD), NIST Special Publication (SP) 260-222. <https://doi.org/10.6028/NIST.SP.260-222>

Certificate Revision History: 18 May 2026 (Uncertainty value update; removal of Total Oxides of Nitrogen (NO_x) value; change of period of validity; editorial changes; format update); 31 August 2012 (Original certification date).

Certain commercial equipment, instruments, or materials may be identified in this Certificate of Analysis to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.

Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the Office of Reference Materials 100 Bureau Drive, Stop 2300, Gaithersburg, MD 20899-2300; telephone (301) 975-2200; e-mail srminfo@nist.gov; or the Internet at <https://www.nist.gov/srm>.

*** * * * * End of Certificate of Analysis * * * * ***

APPENDIX A

The table contains all cylinder Sample Numbers and their associated information for this lot.

Sample Number	Cylinder Number	Hydrotest Date	Fill Date
44-U-05	FF55546	November 2015	September 2010
44-U-06	FF55489	November 2015	September 2010
44-U-07	FF55557	November 2015	September 2010
44-U-08	FF55469	November 2015	September 2010
44-U-09	FF55522	November 2015	September 2010
44-U-10	FF55482	November 2015	September 2010
44-U-11	FF55508	November 2015	September 2010
44-U-12	FF55573	November 2015	September 2010
44-U-13	FF55470	November 2015	September 2010
44-U-15	FF55533	November 2015	September 2010
44-U-16	FF55631	November 2015	September 2010
44-U-17	FF55477	November 2015	September 2010
44-U-18	FF55497	November 2015	September 2010
44-U-19	FF55626	November 2015	September 2010
44-U-20	FF55524	November 2015	September 2010
44-U-21	FF55495	November 2015	September 2010
44-U-22	FF55503	November 2015	September 2010
44-U-23	FF55525	November 2015	September 2010
44-U-24	FF55479	November 2015	September 2010
44-U-25	FF55458	November 2015	September 2010
44-U-27	FF55585	November 2015	September 2010
44-U-28	FF55548	November 2015	September 2010
44-U-29	FF55510	November 2015	September 2010
44-U-31	FF55519	November 2015	September 2010
44-U-32	FF55493	November 2015	September 2010
44-U-33	FF55516	November 2015	September 2010
44-U-34	FF55504	November 2015	September 2010
44-U-35	FF55473	November 2015	September 2010
44-U-36	FF55513	November 2015	September 2010
44-U-38	FF55530	November 2015	September 2010
44-U-39	FF55578	November 2015	September 2010
44-U-40	FF55465	November 2015	September 2010
44-U-41	FF55520	November 2015	September 2010
44-U-42	FF55452	November 2015	September 2010
44-U-43	FF55476	November 2015	September 2010
44-U-44	FF55480	November 2015	September 2010
44-U-45	FF55483	November 2015	September 2010
44-U-46	FF55491	November 2015	September 2010
44-U-47	FF55518	November 2015	September 2010
44-U-48	FF55560	November 2015	September 2010
44-U-49	FF55509	November 2015	September 2010
44-U-50	FF55511	November 2015	September 2010
44-U-51	FF55602	November 2015	September 2010
44-U-54	FF55496	November 2015	September 2010
44-U-55	FF55505	November 2015	September 2010
44-U-56	FF55539	November 2015	September 2010
44-U-57	FF55550	November 2015	September 2010
44-U-58	FF55494	November 2015	September 2010
44-U-59	FF55537	November 2015	September 2010
44-U-60	FF55554	November 2015	September 2010
44-U-61	FF55487	November 2015	September 2010
44-U-62	FF55488	November 2015	September 2010

(continued)

Sample Number	Cylinder Number	Hydrotest Date	Fill Date
44-U-64	FF55484	November 2015	September 2010
44-U-65	FF55606	November 2015	September 2010
44-U-67	FF55589	November 2015	September 2010
44-U-68	FF55544	November 2015	September 2010
44-U-69	FF55463	November 2015	September 2010

* * * * * End of Appendix A * * * * *