



National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1815a

Reference Fuel n-Heptane

(In cooperation with the Service des Matériaux
de Référence de the Bureau National
de Métrologie in Paris France)

This Standard Reference Material (SRM) is intended for use as a primary standard in the octane rating of motor and aviation fuels as specified in ASTM test methods (see Volume 05.04, Annual Book of ASTM Standards) and evaluating ASTM methods for chemical analysis of fuels by gas chromatography (D 2268). This SRM exceeds the ASTM specification for a reference fuel. The certified values are:

n-Heptane, purity by difference	99.987%
<u>Impurities</u>	
Total organics (other than n-Heptane)	0.011 ± 0.003%
Isooctane (2,2,4-Trimethylpentane)	0.007 ± 0.001%
Water	0.002 ± 0.001%

The lead concentration in this SRM is certified to be less than 10 µg/L.

The material for this SRM was obtained from the ELF Company in France, ampouled by the Laboratoire National d'Essais (LNE), and analyzed by NBS, LNE, and the Institut Français du Pétrole (IFP)

The homogeneity of this SRM, as determined by LNE using the gas chromatography technique, was judged to be satisfactory when the first three ampouled series and the two ampoules of the following series were discarded.

All three laboratories followed the gas chromatography technique specified in ASTM Method D 2268 to determine total organic impurities. The water content was determined by LNE using the classical Karl Fischer method.

The results obtained by the three laboratories are given below:

	<u>NBS</u>	<u>LNE</u>	<u>IFP</u>
Isooctane	0.006 ± 0.001%	0.006 ± 0.001%	0.008 ± 0.001%
Total organics	0.010 ± 0.002%	0.013 ± 0.004%	0.010 ± 0.001%
Water	-----	0.002 ± 0.001%	-----
Lead	<0.4 µg/L	5.4 ± 3.1 µg/L	<5 µg/L

The lead concentration was determined by NBS and LNE using various modifications of ASTM Method D 1368-77 while IFP used a variation of atomic absorption spectrometry (D 3237).

March 15, 1985
Gaithersburg, MD 20899

Stanley D. Rasberry, Chief
Office of Standard Reference Materials