

Standard Reference Material<sup>®</sup> 2619a  
Carbon Dioxide in Nitrogen  
(Nominal Amount-of-Substance Fraction 0.5 % mol/mol)  
Lot 30-F-XX

**CERTIFICATE OF ANALYSIS**

**Purpose:** The certified value delivered by this Standard Reference Material (SRM) is intended for use in producing metrologically traceable secondary standards for the calibration of instruments used for carbon dioxide determinations.

**Description:** This SRM is a primary gas mixture supplied in a DOT 3AL-specification aluminum (6061 alloy) cylinder with a water volume of 6 L. Mixtures are shipped with a nominal pressure exceeding 12.4 MPa (1800 psig), which provides the user with 0.73 m<sup>3</sup> (25.8 ft<sup>3</sup>) of useable mixture. The cylinder is the property of the purchaser and is equipped with a CGA-580 brass valve, which is the recommended outlet for this carbon dioxide mixture.

**Certified Value:** A NIST certified value is a value for which NIST has the highest confidence in that all known or suspected sources of bias and imprecision have been accounted [1]. This SRM mixture has been certified for carbon dioxide amount-of-substance fraction (mole fraction, sometimes termed “molar concentration”) [2]. This certified value is traceable to the International System of Units (SI) through the gravimetric primary standards and procedures used in the preparation of this mixture. The certified value given below applies to all cylinders in this lot.

Carbon Dioxide Mole Fraction: 0.4900 % mol/mol ± 0.0012 % mol/mol<sup>(a)</sup>

<sup>(a)</sup> The certified value is expressed as  $x \pm U_{95\%}(x)$ , where  $x$  is the value and  $U_{95\%}(x)$  is the expanded uncertainty of the value. The true value of the analyte lies within the interval  $x \pm U_{95\%}(x)$  with 95 % confidence. The uncertainty is expressed as an expanded uncertainty  $U = k u_c$  with  $u_c$  determined by experiment and a coverage factor  $k = 2$ . For guidance in propagating this uncertainty, see reference 3.

**Additional Information:** Appendix A contains a table of all cylinder Sample Numbers and their associated information for this lot.

**Period of Validity:** The certified value delivered by **SRM 2619a, Lot No. 30-F-XX** is valid within the measurement uncertainty specified until **24 July 2033**, provided the SRM is handled and stored in accordance with the instructions given in this certificate (see “Storage” and “Cylinder and Gas Handling Information”). The certification is nullified if the SRM is damaged, contaminated or otherwise modified, or if the internal pressure drops below 0.7 MPa (100 psig).

**Maintenance of Certified Value:** NIST will monitor this SRM over the period of its validity. If substantive technical changes occur that affect the certification, NIST will issue an amended certificate through the NIST SRM website (<https://www.nist.gov/srm>). Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (<https://www.nist.gov/srm>).

Carlos A. Gonzalez, Chief  
Chemical Sciences Division  
*Certificate Revision History on Page 2*

Steven J. Choquette, Director  
Office of Reference Materials

**Safety:** Consult the Safety Data Sheet (SDS) for hazard information.

**Storage:** This SRM should be stored under normal laboratory conditions within the temperature range of 15 °C to 30 °C.

**Cylinder and Gas Handling Information:** NIST recommends the use of a high purity, two stage pressure regulator with a stainless-steel diaphragm and CGA-580 outlet to safely reduce the pressure and to deliver this mixture to the instrument. The regulator should be purged to prevent accidental contamination of the SRM by repeatedly (minimum three times) opening the valve and pressurizing the regulator, then closing the valve and releasing the pressure safely into a vent line.

**Mixture Preparation:** The gas mixtures comprising this SRM lot were prepared in accordance with NIST technical specifications by a commercial specialty gas vendor under contract to NIST. The specifications stipulate that each SRM mixture be identical in carbon dioxide mole fraction and stable with time.

**Analytical Method:** Analyses of the carbon dioxide mole fraction for this lot of cylinders were conducted by comparison to a representative cylinder chosen from the lot, designated as the SRM lot standard (LS). The LS was compared to NIST primary gravimetric standards using Gas Chromatography with Thermal Conductivity Detector (GC-TCD). Each of the carbon dioxide mixtures that comprise this SRM lot was then compared to the LS using GC-TCD. Within the precision of the NIST measurements, all the cylinders comprising this SRM lot have identical carbon dioxide mole fractions.

**CAS Registry Numbers:** This SRM is certified for carbon dioxide in nitrogen. The relevant CAS Registry numbers for these components are: carbon dioxide CAS Registry 124-38-9; nitrogen (balance gas) CAS Registry 7727-37-9.

## NOTICE TO USERS

NIST strives to maintain the SRM inventory supply, but NIST cannot guarantee the continued or continuous supply of any specific SRM. Accordingly, NIST encourages the use of this SRM as a primary benchmark for the quality and accuracy of the user's in house reference materials and working standards. As such, the SRM should be used to validate the more routinely used reference materials in a laboratory. Comparisons between the SRM and in house reference materials or working measurement standards should take place at intervals appropriate to the conservation of the SRM and the stability of relevant in house materials. For further guidance on how this approach can be implemented, contact NIST by email at [srms@nist.gov](mailto:srms@nist.gov).

## REFERENCES

- [1] Beauchamp, C.R.; Camara, J.E.; Carney, J.; Choquette, S.J.; Cole, K.D.; DeRose, P.C.; Duerwer, D.L.; Epstein, M.S.; Kline, M.C.; Lippa, K.A.; Lucon, E.; Molloy, J.; Nelson, M.A.; Phinney, K.W.; Polakoski, M.; Possolo, A.; Sander, L.C.; Schiel, J.E.; Sharpless, K.E.; Toman, B.; Winchester, M.R.; Windover, D.; *Metrological Tools for the Reference Materials and Reference Instruments of the NIST Material Measurement Laboratory*; NIST Special Publication (NIST SP) 260-136, 2021 edition; U.S. Government Printing Office: Washington, DC (2021); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-136-2021.pdf> (accessed Feb 2026).
- [2] Thompson, A.; Taylor, B.N.; *Guide for the Use of the International System of Unit (SI)*; NIST Special Publication (NIST SP) 811; U.S. Government Printing Office: Washington, DC (2008); available at <https://www.nist.gov/pml/special-publication-811> (accessed Feb 2026).
- [3] Possolo, A.M.; *Evaluating, Expressing, and Propagating Measurement Uncertainty for NIST Reference Materials*; NIST Special Publication (NIST SP) 260-202; U.S. Government Printing Office: Washington, DC (2020); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-202.pdf> (accessed Feb 2026).

### If you use this SRM in published work, please reference:

Cecelski C, Harris K, Goodman C, Kimes W, Liu Q, Miller W, Carney J (2021) Certification of NIST Gas Mixture Standard Reference Materials. (National Institute of Standards and Technology, Gaithersburg, MD), NIST Special Publication (SP) 260-222. <https://doi.org/10.6028/NIST.SP.260-222>

**Certificate Revision History:** 25 February 2026 (Change of period of validity; editorial changes; format update); 11 August 2011 (Extension of certification period; editorial changes); 18 December 2007 (Original certification date).

*Certain commercial equipment, instruments, or materials may be identified in this Certificate of Analysis to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.*

*Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the Office of Reference Materials 100 Bureau Drive, Stop 2300, Gaithersburg, MD 20899-2300; telephone (301) 975-2200; e-mail [srminfo@nist.gov](mailto:srminfo@nist.gov); or the Internet at <https://www.nist.gov/srm>.*

**\* \* \* \* \* End of Certificate of Analysis \* \* \* \* \***

# APPENDIX A

The table contains all cylinder Sample Numbers and their associated information for this lot.

Sample Number	Cylinder Number	Hydrotest Date	Blend Date
30-F-03	CAL017231	June 2006	August 2006
30-F-04	CAL017233	June 2006	August 2006
30-F-06	CAL017251	June 2006	August 2006
30-F-07	CAL017250	June 2006	August 2006
30-F-08	CAL017230	June 2006	August 2006
30-F-09	CAL017292	June 2006	August 2006
30-F-10	CAL017237	June 2006	August 2006
30-F-11	CAL016999	June 2006	August 2006
30-F-12	CAL017224	June 2006	August 2006
30-F-13	CAL017282	June 2006	August 2006
30-F-15	CAL017264	June 2006	August 2006
30-F-16	CAL017249	June 2006	August 2006
30-F-17	CAL017085	June 2006	August 2006
30-F-18	CAL017241	June 2006	August 2006
30-F-19	CAL017257	June 2006	August 2006
30-F-20	CAL017260	June 2006	August 2006
30-F-21	CAL017262	June 2006	August 2006
30-F-22	CAL017079	June 2006	August 2006
30-F-23	CAL017083	June 2006	August 2006
30-F-24	CAL017238	June 2006	August 2006
30-F-25	CAL017119	June 2006	August 2006
30-F-26	CAL017039	June 2006	August 2006
30-F-27	CAL017242	June 2006	August 2006
30-F-29	CAL017041	June 2006	August 2006
30-F-30	CAL017256	June 2006	August 2006

\* \* \* \* \* End of Appendix A \* \* \* \* \*