

Standard Reference Material[®] 2624a
Carbon Dioxide in Nitrogen
(Nominal Amount-of-Substance Fraction 3 % mol/mol)
Lot 35-D-XX

CERTIFICATE OF ANALYSIS

Purpose: The certified value delivered by this Standard Reference Material (SRM) is intended for use in producing metrologically traceable secondary standards for the calibration of instruments used for carbon dioxide determinations.

Description: This SRM is a primary gas mixture supplied in a DOT 3AL-specification aluminum (6061 alloy) cylinder with a water volume of 6 L. Mixtures are shipped with a nominal pressure exceeding 12.4 MPa (1800 psig), which provides the user with 0.73 m³ (25.8 ft³) of useable mixture. The cylinder is the property of the purchaser and is equipped with a CGA-580 brass valve, which is the recommended outlet for this carbon dioxide mixture.

Certified Value: A NIST certified value is a value for which NIST has the highest confidence in that all known or suspected sources of bias and imprecision have been accounted [1]. This SRM mixture has been certified for carbon dioxide amount-of-substance fraction (mole fraction, sometimes termed “molar concentration”). This certified value is traceable to the International System of Units (SI) through the gravimetric primary standards and procedures used in the preparation of this mixture. The certified value given below applies to all cylinders in this lot.

Carbon Dioxide Mole Fraction: 2.938 % mol/mol $\pm$ 0.006 % mol/mol^(a)

^(a) The certified value is expressed as $x \pm U_{95\%}(x)$, where x is the value and $U_{95\%}(x)$ is the expanded uncertainty of the value. The true value of the analyte lies within the interval $x \pm U_{95\%}(x)$ with 95 % confidence. The uncertainty is expressed as an expanded uncertainty $U = ku_c$ with u_c determined by experiment and a coverage factor $k = 2$. For guidance in propagating this uncertainty, see reference 2.

Additional Information: Appendix A contains a table of all cylinder Sample Numbers and their associated information for this lot.

Period of Validity: The certified value delivered by **SRM 2624a, Lot No. 35-D-XX** is valid within the measurement uncertainty specified until **13 June 2032**, provided the SRM is handled and stored in accordance with the instructions given in this certificate (see “Storage” and “Cylinder and Gas Handling Information”). The certification is nullified if the SRM is damaged, contaminated or otherwise modified, or if the internal pressure drops below 0.7 MPa (100 psig).

Maintenance of Certified Value: NIST will monitor this SRM over the period of its validity. If substantive technical changes occur that affect the certification, NIST will issue an amended certificate through the NIST SRM website (<https://www.nist.gov/srm>). Before making use of any of the values delivered by this material, users should verify they have the most recent version of this documentation, available through the NIST SRM website (<https://www.nist.gov/srm>).

Marcela Najarro, Acting Chief
Chemical Sciences Division
Certificate Revision History on Page 2

Steven J. Choquette, Director
Office of Reference Materials

Safety: Consult the Safety Data Sheet (SDS) for hazard information.

Storage: This SRM should be stored under normal laboratory conditions within the temperature range of 15 °C to 30 °C.

Cylinder and Gas Handling Information: NIST recommends the use of a high purity, two stage pressure regulator with a stainless-steel diaphragm and CGA-580 outlet to safely reduce the pressure and to deliver this mixture to the instrument. The regulator should be purged to prevent accidental contamination of the SRM by repeatedly (minimum three times) opening the valve and pressurizing the regulator, then closing the valve and releasing the pressure safely into a vent line.

Mixture Preparation: The gas mixtures comprising this SRM lot were prepared in accordance with NIST technical specifications by a commercial specialty gas vendor under contract to NIST. The specifications stipulate that each SRM mixture be identical in carbon dioxide mole fraction and stable with time.

Analytical Method: Analyses of the carbon dioxide mole fraction for this lot of cylinders were conducted by comparison to a representative cylinder chosen from the lot, designated as the SRM lot standard (LS). The LS was compared to NIST primary gravimetric standards using Gas Chromatography with Thermal Conductivity Detector (GC-TCD). Each of the carbon dioxide mixtures that comprise this SRM lot was then compared to the LS using GC-TCD. Within the precision of the NIST measurements, all the cylinders comprising this SRM lot have identical carbon dioxide mole fractions.

CAS Registry Numbers: This SRM is certified for carbon dioxide in nitrogen. The relevant CAS Registry numbers for these components are: carbon dioxide CAS Registry 124-38-9; nitrogen (balance gas) CAS Registry 7727-37-9.

NOTICE TO USERS

NIST strives to maintain the SRM inventory supply, but NIST cannot guarantee the continued or continuous supply of any specific SRM. Accordingly, NIST encourages the use of this SRM as a primary benchmark for the quality and accuracy of the user's in house reference materials and working standards. As such, the SRM should be used to validate the more routinely used reference materials in a laboratory. Comparisons between the SRM and in house reference materials or working measurement standards should take place at intervals appropriate to the conservation of the SRM and the stability of relevant in house materials. For further guidance on how this approach can be implemented, contact NIST by email at srms@nist.gov.

REFERENCES

- [1] Beauchamp, C.R.; Camara, J.E.; Carney, J.; Choquette, S.J.; Cole, K.D.; DeRose, P.C.; Duewer, D.L.; Epstein, M.S.; Kline, M.C.; Lippa, K.A.; Lucon, E.; Molloy, J.; Nelson, M.A.; Phinney, K.W.; Polakoski, M.; Possolo, A.; Sander, L.C.; Schiel, J.E.; Sharpless, K.E.; Toman, B.; Winchester, M.R.; Windover, D.; *Metrological Tools for the Reference Materials and Reference Instruments of the NIST Material Measurement Laboratory*; NIST Special Publication (NIST SP) 260-136, 2021 edition; U.S. Government Printing Office: Washington, DC (2021); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-136-2021.pdf> (accessed Feb 2026).
- [2] Possolo, A.M.; *Evaluating, Expressing, and Propagating Measurement Uncertainty for NIST Reference Materials*; NIST Special Publication (NIST SP) 260-202; U.S. Government Printing Office: Washington, DC (2020); available at <https://nvlpubs.nist.gov/nistpubs/SpecialPublications/NIST.SP.260-202.pdf> (accessed Feb 2026).

If you use this SRM in published work, please reference:

Cecelski C, Harris K, Goodman C, Kimes W, Liu Q, Miller W, Carney J (2021) Certification of NIST Gas Mixture Standard Reference Materials. (National Institute of Standards and Technology, Gaithersburg, MD), NIST Special Publication (SP) 260-222. <https://doi.org/10.6028/NIST.SP.260-222>

Certificate Revision History: **23 February 2026** (Change of period of validity; updated format; editorial changes); **07 June 2017** (Change of expiration date; editorial changes); **28 July 2011** (Change of expiration date; editorial changes); **07 June 2007** (Extension of certification period); **01 June 2004** (Change in total uncertainty; change of expiration date; editorial changes); **05 September 2001** (Change of expiration date); **25 July 1996** (Original certification date).

Certain commercial equipment, instruments, or materials may be identified in this Certificate of Analysis to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.

Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the Office of Reference Materials 100 Bureau Drive, Stop 2300, Gaithersburg, MD 20899-2300; telephone (301) 975-2200; e-mail srminfo@nist.gov; or the Internet at <https://www.nist.gov/srm>.

* * * * * End of Certificate of Analysis * * * * *

APPENDIX A

The table contains all cylinder Sample Numbers and their associated information for this lot.

Sample Number	Cylinder Number	Hydrotest Date	Blend Date
35-D-03	CAL010458	May 1995	November 1995
35-D-04	CAL010445	May 1995	November 1995
35-D-05	CAL010464	May 1995	November 1995
35-D-06	CAL010758	May 1995	November 1995
35-D-07	CAL010415	May 1995	November 1995
35-D-08	CAL010389	May 1995	November 1995
35-D-09	CAL010424	May 1995	November 1995
35-D-11	CAL010476	May 1995	November 1995
35-D-12	CAL010385	May 1995	November 1995
35-D-13	CAL010764	May 1995	November 1995
35-D-14	CAL010472	May 1995	November 1995
35-D-15	CAL010460	May 1995	November 1995
35-D-16	CAL010846	May 1995	November 1995
35-D-17	CAL010818	May 1995	November 1995
35-D-18	CAL010742	May 1995	November 1995
35-D-20	CAL010387	May 1995	November 1995
35-D-21	CAL010812	May 1995	November 1995
35-D-22	CAL010430	May 1995	November 1995
35-D-23	CAL010738	May 1995	November 1995
35-D-24	CAL010457	May 1995	November 1995
35-D-25	CAL010379	May 1995	November 1995
35-D-26	CAL010462	May 1995	November 1995
35-D-27	CAL010770	May 1995	November 1995
35-D-29	CAL010482	May 1995	November 1995
35-D-30	CAL010403	May 1995	November 1995
35-D-31	CAL010746	May 1995	November 1995
35-D-32	CAL010475	May 1995	November 1995
35-D-33	CAL010760	May 1995	November 1995
35-D-35	CAL010816	May 1995	November 1995
35-D-37	CAL010808	May 1995	November 1995
35-D-38	CAL010852	May 1995	November 1995
35-D-39	CAL010478	May 1995	November 1995
35-D-40	CAL010784	May 1995	November 1995
35-D-41	CAL010790	May 1995	November 1995
35-D-42	CAL010824	May 1995	November 1995
35-D-43	CAL010399	May 1995	November 1995
35-D-44	CAP010798	May 1995	November 1995
35-D-45	CAL010796	May 1995	November 1995
35-D-46	CAL010454	May 1995	November 1995
35-D-47	CAL010377	May 1995	November 1995
35-D-49	CAL010470	May 1995	November 1995
35-D-50	CAL010456	May 1995	November 1995
35-D-51	CAL010806	May 1995	November 1995
35-D-52	CAL010778	May 1995	November 1995
35-D-53	CAL010748	May 1995	November 1995
35-D-54	CAL010466	May 1995	November 1995
35-D-55	CAL010449	May 1995	November 1995
35-D-56	CAL010792	May 1995	November 1995
35-D-57	CAL010788	May 1995	November 1995
35-D-58	CAL010744	May 1995	November 1995

* * * * * End of Appendix A * * * * *