



National Institute of Standards & Technology

Certificate

Standard Reference Material 2631

Nitric Oxide in Nitrogen

(Mobile-Source Emission Gas Standard)

Nominal Concentration 3000 $\mu\text{mol/mol}$ (ppm)

(In Cooperation with the Motor Vehicle Manufacturers Association)

This Standard Reference Material (SRM) is intended primarily for the calibration of instruments and the evaluation of methods used for the determination of nitric oxide in mobile-source emissions. It is not intended as a working standard, but rather as a primary standard to which the concentration of the nitric oxide in working standards may be related.

The SRM is supplied in an aluminum cylinder at a nominal pressure of 12.4 MPa (1800 psi) with a deliverable volume of approximately 0.85 m³ (30 cubic feet) at NTP. The cylinder conforms to DOT specifications and is equipped with a CGA-660 valve. The cylinder becomes the property of the purchaser.

Nitric oxide concentration: \pm $\mu\text{mol/mol}$ (ppm)

Cylinder Number: Sample Number:

The concentration of nitric oxide is relative to all other constituents of this gas mixture. The uncertainty shown is the estimated upper limit of error of the certified value at the 95 percent confidence level. This uncertainty includes the estimated inaccuracy of the NIST primary gravimetric standards, the imprecision of intercomparison of the primary gravimetric standards with the batch standards, and the imprecision of the comparison of the SRM with the batch standards.

Each cylinder of gas in the SRM lot is individually analyzed, and the certified value above applies to the cylinder identified by cylinder number and sample number in this certificate.

The certified value given above is valid for four years from the date of shipment from the National Institute of Standards and Technology (NIST). A validation sticker is supplied with the gas cylinder that specifies its certification period. Please affix this sticker to the cylinder upon receipt.

CAUTION: Care must be taken to avoid contamination of the sample during the use of the cylinder with any gas handling system.

The preparation of the primary gravimetric standards and the analytical measurements leading to the certification of this SRM were performed at NIST by W.J. Thorn and G.D. Mitchell.

The overall direction and coordination of the technical measurements leading to certification were performed under the chairmanship of W.D. Dorko and W. E. May, Chief of the NIST Organic Analytical Research Division.

The original research and development of this SRM was supplied by the Motor Vehicle Manufacturers Association of the United States, Inc. (MVMA), Detroit, MI.

Gaithersburg, MD 20899
July 3, 1991
(Revision of certificate dated 5-2-79)

William P. Reed, Chief
Standard Reference Materials Program

The technical and support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by T.E. Gills.

Material Preparation

This SRM is one of a group or "lot" of cylinders that was prepared commercially according to rigid NIST specifications to ensure that the lot is homogeneous and stable. A lot consists of a minimum of 50 cylinders. Each cylinder of gas within the lot is individually analyzed at the NIST for conformity to NIST specifications and is certified according to the NIST protocols and procedures.

Analysis

The nitric oxide content of this SRM was determined by comparison with NIST batch standards that had been previously intercompared with a set of primary gravimetric standards. The method of comparison utilized the chemiluminescent reaction of nitric oxide with ozone.

Representative samples from this SRM lot were analyzed for the presence of other nitrogen oxides. Samples were passed through a high temperature catalytic furnace that converts other oxides to nitric oxide. Under the conditions of the experiment, other oxides of nitrogen would have been detected at a level of 0.5 percent of the nitric oxide concentration. No other oxides of nitrogen were detected within the stated limits in the samples analyzed.

Stability

As stated on the first page of this certificate, the certified value for nitric oxide is valid for four years from the date of shipment from the NIST. Periodic reanalyses of representative samples from this lot will be performed during the four year period, and if a significant change in the certified value is observed, the purchaser will be notified.

Samples from similar gas mixtures have exhibited a change in constituent concentration when the cylinder pressure fell below 1.04 MPa (150 psi). Therefore, it is recommended that the SRM not be used after the pressure has fallen below 2.1 MPa (300 psi).

Reanalysis

The NIST will reanalyze this SRM for the original purchaser at a cost not to exceed the cost of similar SRMs available at the time of the request for reanalysis. The original purchaser should contact the NIST Organic Analytical Research Division (301) 975-3108 to arrange for reanalysis.