



National Institute of Standards & Technology

Certificate of Analysis

Standard Reference Material[®] 2631a

Nitric Oxide in Nitrogen

(Nominal Amount-of-Substance Fraction 3000 $\mu\text{mol/mol}$)

This certificate reports the certified values for Lot 47-F-XX.

This Standard Reference Material (SRM) is a primary gas mixture that, the amount-of-substance fraction expressed as concentration [1], may be related to secondary working standards. The SRM is intended for the calibration of instruments used for nitric oxide determinations and for other uses.

This SRM mixture is supplied in a DOT 3AL specification aluminum (6061 alloy) cylinder with a water volume of 6 L. Mixtures are shipped with a nominal pressure exceeding 12.4 MPa (1800 psig), which provides the user with 0.73 m³ (25.8 ft³) of useable mixture. The cylinder is the property of the purchaser and is equipped with a CGA-660 stainless steel valve, which is the recommended outlet for this nitric oxide mixture.

Certified Value: This SRM mixture has been certified for nitric oxide (NO) and total oxides of nitrogen (NO_x) concentrations. The certified value, given below, applies to the identified cylinder and NIST sample number.

Nitric Oxide (NO) Concentration:	2951.5 $\mu\text{mol/mol}$ \pm 14.8 $\mu\text{mol/mol}$
Total Oxides of Nitrogen (NO _x) Concentration:	2955.6 $\mu\text{mol/mol}$ \pm 14.8 $\mu\text{mol/mol}$

Cylinder Number: SAMPLE	NIST Sample Number: SAMPLE
Hydrotest Date: April 2005	Blend Date: April 2005

The uncertainty of the certified value includes the estimated uncertainties in the NIST standards, the analytical comparisons to the lot standard (LS), and the uncertainty of comparing the LS with each of the mixtures comprising this lot. This uncertainty is expressed as an expanded uncertainty $U = ku_c$, with u_c determined by experiment and a coverage factor $k = 2$. The true value for the nitric oxide amount-of-substance fraction is asserted to lie in the interval defined by the certified value $\pm U$ with a level of confidence of approximately 95 % [2].

Expiration of Certification: The certification of **SRM 2631a Lot No. 47-F-XX** is valid from this certificate issue date, within the measurement uncertainties specified until **13 March 2025**, provided the SRM is handled and stored in accordance with the instructions given in this certificate (see "Cylinder and Gas Handling Information"). The certification is nullified if the SRM is damaged, contaminated, or otherwise modified.

Cylinder and Gas Handling Information: NIST recommends the use of a high-purity, two-stage pressure regulator with a stainless steel diaphragm and CGA-660 outlet to safely reduce the pressure and to deliver this SRM mixture to the instrument. The regulator should be purged to prevent accidental contamination of the SRM by repeatedly (minimum three times) opening the valve and pressurizing the regulator, then closing the valve and releasing the pressure safely into a vent line. This SRM should not be used after the internal pressure drops below 0.7 MPa (100 psig). This SRM should be stored under normal laboratory conditions within the temperature range of 15 °C and 30 °C.

Analytical measurements leading to the certification of this current SRM lot were performed by M.E. Kelley of the NIST Chemical Sciences Division.

Carlos A. Gonzalez, Chief
Chemical Sciences Division

Gaithersburg, MD 20899
Certificate Issue Date: 29 March 2019
Certificate Revision History on Last Page

Steven J. Choquette, Director
Office of Reference Materials

Overall direction and coordination of the technical work required for certification of this SRM were performed by F.R. Guenther of the NIST Chemical Sciences Division.

Support aspects involved in the issuance of this SRM were coordinated through the NIST Office of Reference Materials.

Traceability: The measurand are the total concentration of nitric oxide in nitrogen and the total oxides of nitrogen in nitrogen. Certified values are respectively metrologically traceable to the SI unit of micromoles of nitric oxide per total moles of nitric oxide and nitrogen; and the SI unit of micromoles of total oxides of nitrogen per total moles of total oxides of nitrogen and nitrogen.

Maintenance of SRM Certification: NIST will monitor this SRM over the period of its certification. If substantive technical changes occur that affect the certification before the expiration of this certification, NIST will notify the purchaser. Registration (see attached sheet or register online) will facilitate notification.

Mixture Preparation: The gas mixtures comprising this SRM lot were prepared in accordance with NIST technical specifications by a commercial specialty gas vendor under contract to NIST. The specifications stipulate that each SRM mixture be identical in nitric oxide concentration and stable with time.

Analytical Methods: Analyses of the nitric oxide concentration for this lot of cylinders were conducted by comparing each cylinder mixture to a representative cylinder chosen from the lot, the lot standard (LS), using chemiluminescence. Assignment of the nitric oxide concentration to the LS was accomplished by comparison to primary gravimetric standards using chemiluminescence.

Homogeneity Analysis: Each of the nitric oxide mixtures that comprise this SRM lot was compared to the LS using chemiluminescence. An analysis of variance indicated that sample-to-sample nitric oxide concentration differences were statistically insignificant within the lot. Therefore, a single concentration has been assigned to the lot.

Nitric Oxide Concentration Value Assignment: The certified nitric oxide concentration for each SRM cylinder was computed from the assigned concentration for the lot standard and the homogeneity analysis.

Total Oxides of Nitrogen Concentration Value Assignment: The certified total oxides of nitrogen concentration for each SRM cylinder was computed from the assigned concentration for the lot standard and the homogeneity analysis.

CAS Registry Numbers: This SRM is certified for nitric oxide in nitrogen, and total oxides of nitrogen. The relevant CAS Registry numbers for these components are: nitric oxide CAS Registry 10102-43-9; nitrogen CAS Registry 7727-37-9.

REFERENCES

- [1] Thompson, A.; Taylor, B.N.; *Guide for the Use of the International System of Units (SI)*; NIST Special Publication 811; U.S. Government Printing Office: Washington, DC (2008); available at <https://www.nist.gov/pml/special-publication-811> (accessed Mar 2019).
- [2] JCGM 100:2008; *Evaluation of Measurement Data – Guide to the Expression of in Measurement* (GUM 1995 with Minor Corrections); Joint Committee for Guides in Metrology (JCGM) (2008); available at https://www.bipm.org/utls/common/documents/jcgm/JCGM_100_2008_E.pdf (accessed Mar 2019); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at <https://www.nist.gov/pml/nist-technical-note-1297> (accessed Mar 2019).

Certificate Revision History: 29 March 2019 (Change of expiration date; updated uncertainty values; editorial changes); 25 January 2010 (Extension of certification period); 30 June 2006 (Minor editorial change); 01 May 2006 (Original certificate date).

Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; fax (301) 948-3730; e-mail srminfo@nist.gov; or via the Internet at <https://www.nist.gov/srm>.