



Certificate of Analysis

Standard Reference Material 335

Basic Open-Hearth Steel, 0.1% Carbon

Analyst	Method	Carbon
1	Combustion-gravimetric ^a	Percent 0.093
2	Combustion-gravimetric ^b090
2	Combustion-chromatographic ^c093
2	Combustion-direct reading thermal conductivity ^c090
3	Combustion-gravimetric ^d090
4	Combustion-gravimetric ^b096
5	Combustion-volumetric093
5	Combustion-gravimetric ^e087
6	Combustion-gravimetric ^a093
	Average	0.092

^a 2-g sample.
^b Factor weight (2.73g) sample.
^c 1-g sample.

^d 3-g sample.
^e 5-g sample.

List of Analysts

- | | |
|--|--|
| <ol style="list-style-type: none"> 1. J. R. Baldwin, Division of Analytical Chemistry, Institute for Materials Research, National Bureau of Standards. 2. R. J. Bendure, L. C. Bartels, R. Chenaault, L. S. Steele, and J. S. Hayes, Armco Steel Corp., Middletown, Ohio. 3. R. W. Bley, Inland Steel Co., Indiana Harbor Works, East Chicago, Ind. | <ol style="list-style-type: none"> 4. R. H. Rouse, Bethlehem Steel Corp., Sparrows Point Plant, Sparrows Point, Md. 5. R. H. Elder, R. N. Smith and J. E. Hobby, American Cast Iron Pipe Co., Birmingham, Ala. 6. M. D. Cooper and W. R. Lee, General Motors Corp., Warren, Mich. |
|--|--|

The steel for the preparation of this Standard was furnished by Jones and Laughlin Steel Corp.

WASHINGTON, D. C.
 April 7, 1966

W. WAYNE MEINKE, Chief
 Office of Standard Reference Materials