

National Institute of Standards & Technology **Certificate**

Standard Reference Material® 4410H

Technetium-99m Radioactivity Standard

Lot Number 43

Ampoule 1

This Standard Reference Material (SRM) consists of a solution of a standardized and certified quantity of radioactive technetium-99m in a suitably stable and homogeneous matrix. It is intended primarily for the calibration of instruments that are used to measure radioactivity and for the monitoring of radiochemical procedures. A unit of SRM 4410H consists of approximately 5 mL of a solution, whose composition is specified in Tables 1 and 2, contained in a flame-sealed borosilicate-glass ampoule [1].

The certified **technetium-99m** massic activity value, at a **Reference Time of 1500 EST, 28 November 2017**, is: $(499.7 \pm 4.4) \text{ MBq} \cdot \text{g}^{-1}$

A NIST certified value, as used within the context of this certificate, is a value for which NIST has the highest confidence in its uncertainty assessment. It is a "measurement result" [2] obtained directly or indirectly from a "primary reference measurement procedure" [3]. The certified value is traceable to the derived SI unit, the becquerel (Bq).

Additional physical, chemical, and radiological properties for this SRM, as well as details on the standardization method, are given in Tables 1 and 2. Uncertainties for the certified quantities are expanded (k = 2). The uncertainties are calculated according to the ISO and NIST Guides [4,5]. Table 3 contains a specification of the components that comprise the uncertainty analyses.

Expiration of Certification: The certification of **SRM 4410H** is valid, within the measurement uncertainty specified, within its half-life-dependent useful lifetime, provided the SRM is handled in accordance with instructions given in this certificate (see "Instructions for Handling and Storage"). The certification is nullified if the SRM is damaged, contaminated, or otherwise modified.

Maintenance of Certification: NIST will monitor this SRM over the period of its certification. If substantive technical changes occur that affect the certification, NIST will notify the purchaser.

Radiological and Chemical Hazard: Consult the Safety Data Sheet (SDS), enclosed with the SRM shipment, for radiological and chemical hazard information.

This SRM was prepared in the Physical Measurement Laboratory, Radiation Physics Division, Radioactivity Group, B.E. Zimmerman, Acting Group Leader. The overall production, technical direction and physical measurement leading to certification were provided by W. Regits and K. Neal, Guest Researchers from NRMAP, Incorporated.

Support aspects involved in the issuance of this SRM were coordinated through the NIST Measurement Services Division.

Michael G. Mitch, Acting Chief Radiation Physics Division

Gaithersburg, Maryland 20899 April 2018 Steven J. Choquette, Chief Measurement Services Division

SRM 4410H, Lot 43 Page 1 of 4

Table 1. Certified Massic Activity of SRM 4410H, Lot 43, Ampoule 1

Radionuclide	Technetium-99m
Reference time	1500 EST, 28 November 2017
Massic activity of the solution	499.7 MBq•g ⁻¹
Relative expanded uncertainty $(k = 2)$	0.88 % ^(a)

^(a)The uncertainties on certified values are expanded uncertainties, $U = ku_c$. The quantity u_c is the combined standard uncertainty calculated according to the ISO and NIST Guides [4,5]. The combined standard uncertainty is multiplied by a coverage factor of k = 2 and was chosen to obtain an approximate 95 % level of confidence.

Table 2. Uncertified Information of SRM 4410H, Lot 43, Ampoule 1

Source description	Liquid in a flame-sealed 5 mL NIST borosilicate ampoule [1]	
Solution composition	Sodium pertechnetate in normal saline (0.9 % NaCl) solution	
Solution density	$(1.005 \pm 0.002) \text{ g} \cdot \text{mL}^{-1} \text{ at } 20.0 ^{\circ}\text{C}^{(a)}$	
Solution mass	$(5.0241 \pm 0.0003) g^{(a)}$	
Photon-emitting impurities	⁹⁹ Mo: (80 ± 29) Bq•g ^{-1 (b)}	
Half-lives used	$^{99\text{m}}$ Tc: $(6.0067 \pm 0.0010) \text{ h}^{(c)}$ 99 Mo: $(2.7479 \pm 0.0006) \text{ d}$	
Calibration method (and instruments)	Measurements of ionization current ratios relative to radium-226 reference sources using NIST pressurized " 4π " γ ionization chamber "B" calibrated with a technetium-99m solution whose activity was determined by the $4\pi(eX)$ - γ -anticoincidence efficiency-extrapolation technique	

⁽a) The stated uncertainty is two times the standard uncertainty.

0.4 s⁻¹•g⁻¹ for energies between 30 keV and 155 keV,

 $0.2 \text{ s}^{-1} \cdot \text{g}^{-1}$ for energies between 160 keV and 1480 keV, and

0.1 s⁻¹•g⁻¹ for energies between 1490 keV and 2000 keV,

provided that any impurity photons are separated by four keV or more from photons emitted in the decay of technetium-99m or molybdenum-99.

SRM 4410H, Lot 43

⁽b) The estimated lower limits of detection for photon-emitting impurities, expressed as massic photon emission rates, as of 4 December 2017 were:

⁽c) The stated uncertainty is the standard uncertainty. See reference 6.

Table 3. Uncertainty Evaluation for the Massic Activity of SRM 4410H, Lot 43

	Uncertainty component	Assessment Type ^(a)	Relative standard uncertainty contribution on massic activity of technetium-99m (%)
1	Ionization-chamber measurement precision on this solution; standard deviation of the mean for four sets of measurements on 12 ampoules $(n = 12)$	A	0.04
2	" 4π " γ ionization-chamber calibration factor	В	0.42
3	Decay correction for radium-226 reference source to correct the calibration factor (for half-life uncertainty of 0.44 %)	В	0.001
4	Radium-226 reference sources ratio	В	0.02
5	Radium reference source positioning	В	0.05
6	Electrometer response linearity	В	0.10
7	Gravimetric mass measurements	В	0.05
8	Correction for photon-emitting impurities	В	< 0.0001
9	Decay correction for technetium-99m (for half-life uncertainty of 0.02 %)	В	0.001
10	Detection limits for photon-emitting impurities	В	< 0.0001
Rela	tive combined standard uncertainty	0.44	
Relative expanded uncertainty $(k = 2)$			0.88

^(a)Type A denotes evaluation by statistical methods; Type B denotes evaluation by other methods.

SRM 4410H, Lot 43

INSTRUCTIONS FOR HANDLING AND STORAGE

Handling: If the ampoule is transported, it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) because of the radioactivity. The ampoule should be opened only by persons qualified to handle radioactive material. Appropriate shielding and/or distance should be used to minimize personnel exposure. Refer to the SDS for further information.

Storage: SRM 4410H should be stored and used at a temperature between 5 °C and 65 °C. The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material.

REFERENCES

- [1] NIST Physical Measurement Laboratory; *Storage and Handling of Radioactive Standard Reference Materials, Ampoule Specifications and Opening Procedure*; available at https://www.nist.gov/pml/radiation-physics/ampoule-specifications-and-opening-procedure (accessed November 2017).
- [2] JCGM 200:2012; International Vocabulary of Metrology Basic and General Concepts and Associated Terms (VIM) (2008 version with Minor Corrections), 3rd edition; Joint Committee for Guides in Metrology: BIPM, Sèvres Cedex, France; p. 19 (2012); available at http://www.bipm.org/utils/common/documents/jcgm/JCGM_200_2012.pdf (accessed November 2017).
- [3] JCGM 200:2012; *International Vocabulary of Metrology Basic and General Concepts and Associated Terms* (*VIM*) (2008 version with Minor Corrections), 3rd edition; Joint Committee for Guides in Metrology: BIPM, Sèvres Cedex, France; p. 18 (2012); available at http://www.bipm.org/utils/common/documents/jcgm/JCGM_200_2012.pdf (accessed November 2017).
- [4] JCGM 100:2008; *Guide to the Expression of Uncertainty in Measurement*; (ISO GUM 1995 with Minor Corrections), Joint Committee for Guides in Metrology: BIPM, Sèvres Cedex, France (2008); available at http://www.bipm.org/utils/common/documents/jcgm/JCGM_100_2008_E.pdf (accessed November 2017).
- [5] Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297, U.S. Government Printing Office: Washington, DC (1994); available at https://www.nist.gov/pml/nist-technical-note-1297 (accessed November 2017).
- [6] Bé, M., Chechev, A., Egorov, A., Morillon, C., *LNE-LNHB/CEA Table of Radionuclides*, ^{99m}Tc; (5 January 2012); available at http://www.nucleide.org/DDEP_WG/Nuclides/Tc-99m_tables.pdf (accessed November 2017).

Users of this SRM should ensure that the Certificate in their possession is current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; fax (301) 948-3730; e-mail srminfo@nist.gov; or via the internet at http://www.nist.gov/srm.

SRM 4410H, Lot 43 Page 4 of 4