



National Institute of Standards & Technology

Certificate

Standard Reference Material 4949C Iodine-129 Radioactivity Standard

This Standard Reference Material (SRM) consists of radioactive iodine-129 as sodium iodide, sodium hydroxide, and sodium sulphite dissolved in 5 mL of distilled water. The solution is contained in a flame-sealed NIST borosilicate-glass ampoule. The SRM is intended for the calibration of beta-particle counting instruments and for the monitoring of radiochemical procedures.

Radiological Hazard: The SRM ampoule contains iodine-129 with a total activity of approximately 17 kBq. Iodine-129 decays by beta-particle emission. None of the beta particles escape from the SRM ampoule. During the decay process, X-rays and gamma rays with energies from 4 keV to 40 keV are also emitted. Most of these photons escape from the SRM ampoule but their intensities are so small that they do not represent a significant radiation hazard. Appropriate shielding and/or distance should be used to minimize personnel exposure. The SRM should be used only by persons qualified to handle radioactive material.

Chemical Hazard: The SRM ampoule contains sodium hydroxide (NaOH) with a concentration of 0.01 mole per liter of water. The solution is mildly corrosive and represents a health hazard if it comes in contact with eyes or skin. If the ampoule is to be opened to transfer the solution, the recommended procedure is given on page 2. The ampoule should be opened only by persons qualified to handle both radioactive material and alkaline solution.

Storage and Handling: The SRM should be stored and used at a temperature between 5 and 65 °C. The solution in an unopened ampoule should remain stable and homogeneous for at least five (5) years after receipt. The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material. If the ampoule is transported, it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) because of both the radioactivity and the strong base.

Preparation: This Standard Reference Material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, J.M.R. Hutchinson, Group Leader. The overall technical direction and physical measurements leading to certification were provided by L.L. Lucas, formerly of the Radioactivity Group. The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program.

Lisa R. Karam, Deputy Chief
Ionizing Radiation Division

Gaithersburg, Maryland 20899
July 1995
Text revised and expiration date extended April 2006

Robert L. Watters, Jr., Chief
Measurement Services Division

Silica in the SRM Ampoule

The solution in the SRM ampoule has a pH of approximately 11 and will slowly etch microgram quantities of silica from the glass ampoule. You may be able to see this silica at times, as it has a density greater than that of the SRM solution and tends to settle. Tests have shown that this silica does not affect the iodine-129 concentration. If desired, the silica may be redispersed in the SRM solution by shaking well. Alternatively, if desired, the silica may be filtered out in the transfer process.

Recommended Procedure for Opening the SRM Ampoule

- 1) If the SRM solution is to be diluted, it is recommended that the diluting solution have a composition comparable to that of the SRM solution.
- 2) Wear eye protection, gloves, and protective clothing and work over a tray with absorbent paper in it. Work in a fume hood. In addition to the radioactive material, the solution contains strong base and is corrosive.
- 3) Shake the ampoule to wet all of the inside surface of the ampoule. Return the ampoule to the upright position.
- 4) Check that all of the liquid has drained out of the neck of the ampoule. If necessary, gently tap the neck to speed the process.
- 5) Holding the ampoule upright, score the narrowest part of the neck around its entire circumference with a scribe or diamond pencil.
- 6) Lightly wet the scored line. This reduces the crack propagation velocity and makes for a cleaner break.
- 7) Hold the ampoule upright with a paper towel, a wiper, or a support jig. Using a paper towel or wiper to avoid contamination, snap off the top of the ampoule by pressing the narrowest part of the neck away from you while pulling the tip of the ampoule towards you.
- 8) Transfer the solution from the ampoule using a pycnometer or a pipet with dispenser handle. NEVER PIPETTE BY MOUTH.
- 9) Seal any unused SRM solution in a flame-sealed glass ampoule, if possible, to minimize the evaporation loss.

See also reference [4]*.

PROPERTIES OF SRM 4949C

Certified values

Radionuclide	Iodine-129
Reference time	1200 EST, 21 March 1993
Massic activity of the solution [a]*	3451 Bq • g⁻¹
Relative expanded uncertainty (k=2)	0.64% [b][c]

Uncertified values

Physical Properties:			
Source description	Liquid in flame-sealed NIST borosilicate-glass ampoule		
Ampoule specifications	Body outside diameter	(16.5 ± 0.5) mm	
	Wall thickness	(0.60 ± 0.04) mm	
	Barium content	Less than 2.5%	
	Lead-oxide content	Less than 0.02%	
	Other heavy elements	Trace quantities	
Solution density	(1.003 ± 0.002) g • mL ⁻¹ at 21.2 °C [d]		
Solution mass	Approximately 5.0 g		
Chemical Properties:			
Solution composition	Chemical Formula	Concentration (mol•L ⁻¹)	Mass Fraction (g•g ⁻¹)
	H ₂ O	55	1.00
	NaOH	0.01	0.0004
	Na ₂ SO ₃	0.006	0.0008
	Na ¹²⁹ I	0.004	0.0006
Radiological Properties:			
Photon-emitting impurities	None detected [e]		
Half lives used	Iodine-129: (1.57 ± 0.04) × 10 ⁷ a [5][f]		
Calibration method and measuring instrument	NIST 4πβ(LS)-γ-anticoincidence counting system.		

EVALUATION OF THE UNCERTAINTY OF THE MASSIC ACTIVITY [b]*

Input Quantity x_i , the source of uncertainty (and individual uncertainty components where appropriate)	Method Used To Evaluate $u(x_i)$, the standard uncertainty of x_i (A) denotes evaluation by statistical methods (B) denotes evaluation by other methods	Relative Uncertainty Of Input Quantity, $u(x_i)/x_i$, (%) [g]	Relative Sensitivity Factor, $ \partial y/\partial x_i \cdot$ (x_i/y) [h]	Relative Uncertainty Of Output Quantity, $u_i(y)/y$, (%) [i]
Massic liquid-scintillation count rate, corrected for background and decay	Standard deviation of the mean for 6 sets of repeated measurements. (A)	0.12	1.0	0.12
Background	Standard deviation of the mean for 6 sets of repeated measurements. (A)	1.0	[j] 0.01	0.01
Mass calibration of the balance	Estimated from manufacturer's data (B)	0.05	1.0	0.05
Decay correction for iodine-129	Standard uncertainty of the half life	2.55 [k]	0.00 [m]	0.00
Decay-scheme data	Standard uncertainty of the probability of beta-particle emission to the excited state (B)	0.10	1.0	0.10
Live time [n]	Estimated (B)	0.10	1.0	0.10
Extrapolation of beta-particle-count-rate versus anticoincidence-gamma-ray-count-rate to zero anticoincidence-gamma-ray-count-rate	Estimated (B)	0.25	1.0	0.25
Photon-emitting impurities	Limit of detection (B) [p]	100	0.00002	0.002
Relative Combined Standard Uncertainty of the Output Quantity, $u_c(y)/y$, (%)				0.32
Coverage Factor, k				<u>x 2</u>
Relative Expanded Uncertainty of the Output Quantity, U/y , (%)				0.64

NOTES

- [a] **Massic activity** is the preferred name for the quantity activity divided by the total mass of the sample. See reference [1].
- [b] The reported value, y , of massic activity (activity per unit mass) at the reference time was not measured directly but was derived from measurements and calculations of other quantities. This can be expressed as $y = f(x_1, x_2, x_3, \dots, x_n)$, where f is a mathematical function derived from the assumed model of the measurement process. The value, x_i , used for each input quantity i has a **standard uncertainty**, $u(x_i)$, that generates a corresponding uncertainty in y , $u_i(y) \equiv |\partial y / \partial x_i| \cdot u(x_i)$, called a **component of combined standard uncertainty** of y . The **combined standard uncertainty** of y , $u_c(y)$, is the positive square root of the sum of the squares of the components of combined standard uncertainty. The combined standard uncertainty is multiplied by a **coverage factor** of $k = 2$ to obtain U , the **expanded uncertainty** of y .
- Since it can be assumed that the possible estimated values of the massic activity are approximately normally distributed with approximate standard deviation $u_c(y)$, the unknown value of the massic activity is believed to lie in the interval $y \pm U$ with a level of confidence of approximately 95 percent.
- For further information on the expression of uncertainties, see references [2] and [3].
- [c] The value of each component of combined standard uncertainty, and hence the value of the expanded uncertainty itself, is a best estimate based upon all available information, but is only approximately known. That is to say, the "uncertainty of the uncertainty" is large and not well known. This is true for uncertainties evaluated by statistical methods (e.g., the relative standard deviation of the standard deviation of the mean for the massic response is approximately 50%) and for uncertainties evaluated by other methods (which could easily be over estimated or under estimated by substantial amounts). The unknown value of the expanded uncertainty is believed to lie in the interval $U/2$ to $2U$ (i.e., within a factor of 2 of the estimated value).
- [d] The stated uncertainty is two times the standard uncertainty.
- [e] The estimated limit of detection for photon-emitting impurities, expressed as massic photon emission rates, is:
 $0.07 \text{ s}^{-1} \cdot \text{g}^{-1}$ for energies between 43 keV and 2700 keV.
- [f] The stated uncertainty is the standard uncertainty.
- [g] Relative standard uncertainty of the input quantity x_i .
- [h] The relative change in the output quantity y divided by the relative change in the input quantity x_i . If $|\partial y / \partial x_i| \cdot (x_i / y) = 1.0$, then a 1% change in x_i results in a 1% change in y . If $|\partial y / \partial x_i| \cdot (x_i / y) = 0.05$, then a 1% change in x_i results in a 0.05% change in y .
- [i] Relative component of combined standard uncertainty of output quantity y , rounded to two significant figures or less. The relative component of combined standard uncertainty of y is given by $u_i(y) / y \equiv |\partial y / \partial x_i| \cdot u(x_i) / y = |\partial y / \partial x_i| \cdot (x_i / y) \cdot u(x_i) / x_i$. The numerical values of $u(x_i) / x_i$, $|\partial y / \partial x_i| \cdot (x_i / y)$, and $u_i(y) / y$, all dimensionless quantities, are listed in columns 3, 4, and 5, respectively. Thus, the value in column 5 is equal to the value in column 4 multiplied by the value in column 3. The input quantities are independent, or very nearly so. Hence the covariances are zero or negligible.
- [j] $|\partial y / \partial x_i| \cdot (x_i / y) = (\text{average background count rate}) / (\text{average net sample count rate})$

- [k] The relative standard uncertainty of $\lambda \cdot t$ is determined by the relative standard uncertainty of λ (i.e., of the half life). The relative standard uncertainty of t is negligible.
- [m] $|\partial y / \partial x_i| \cdot (x_i / y) = |\lambda \cdot t|$
- [n] The live time is determined by counting the pulses from a gated crystal-controlled oscillator.
- [p] The standard uncertainty for each undetected impurity that might reasonably be expected to be present is estimated to be equal to the estimated limit of detection for that impurity, i.e. $u(x_i) / x_i = 100\%$. $|\partial y / \partial x_i| \cdot (x_i / y) = \{(\text{response per Bq of impurity}) / (\text{response per Bq of iodine-129})\} \cdot \{(\text{Bq of impurity}) / (\text{Bq of iodine-129})\}$. Thus, $u_i(y) / y$ is the relative change in y if the impurity were present with a massic activity equal to the estimated limit of detection.

REFERENCES

- [1] International Organization for Standardization (ISO), *ISO Standards Handbook - Quantities and Units*, 1993. Available from Global Engineering Documents, 12 Inverness Way East, Englewood, CO 80112, U.S.A. Telephone 1-800-854-7179.
- [2] International Organization for Standardization (ISO), *Guide to the Expression of Uncertainty in Measurement*, 1993 (corrected and reprinted, 1995). Available from Global Engineering Documents, 12 Inverness Way East, Englewood, CO 80112, U.S.A. Telephone 1-800-854-7179.
- [3] B. N. Taylor and C. E. Kuyatt, *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*, NIST Technical Note 1297, 1994. Available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20407, U.S.A.
- [4] National Council on Radiation Protection and Measurements Report No. 58, *A Handbook of Radioactivity Measurements Procedures*, Second Edition, 1985. Available from the National Council on Radiation Protection and Measurements, 7910 Woodmont Avenue, Bethesda, MD 20814 U.S.A.
- [5] Evaluated Nuclear Structure Data File (ENSDF), July 1995.