



**UNITED STATES DEPARTMENT OF COMMERCE**  
**National Institute of Standards and Technology**  
Gaithersburg, Maryland 20899-0001

DATE: 09 May 2016

**Product Identifier**

**SRM Number:** 1017b

**SRM Name:** Glass Beads - Particle Size Distribution  
(100  $\mu\text{m}$  to 400  $\mu\text{m}$  diameter range)

Under the U.S. Department of Labor, Occupational Safety and Health Administration (OSHA) 29 CFR 1910.1200, this Standard Reference Material (SRM) is NOT classified as a physical hazard or a health hazard, a simple asphyxiant, combustible dust, pyrophoric gas, or hazard not otherwise classified. There are no hazard pictograms, hazard statements or signal word associated with it. Safety Data Sheet information is not required. This document may be used in conjunction with your hazard communication program.

This material is formed to a specific shape or design during manufacture which under normal conditions of use does not release more than very small quantities, e.g., minute or trace amounts of a hazardous chemical (as determined under paragraph (d) of 1910.1200), and does not pose a physical hazard or health risk to employees.

**Description:** This SRM is intended primarily for use in evaluating and calibrating particle size measuring instruments covering the 100  $\mu\text{m}$  to 400  $\mu\text{m}$  range. Typical use is in the evaluation of wire-cloth test sieves in the range of 106  $\mu\text{m}$  (No. 140) through 355  $\mu\text{m}$  (No. 45). A unit of SRM 1017b consists of approximately 70 g of solid spherical soda-lime glass spheres contained in a glass bottle.

**Disposal:** SRM 1017b should be disposed of in accordance with local, state, and federal regulations.

**Transport Information:** This material is not regulated by the U.S. Department of Transportation (DOT) and/or International Air Transport Association (IATA).

**Disclaimer:** This document was prepared carefully, using current references. Users of this SRM should ensure that this document and the corresponding Certificate in their possession are current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; fax (301) 948-3730; e-mail [srmmsds@nist.gov](mailto:srmmsds@nist.gov); or via the Internet at <http://www.nist.gov/srm>.