



DATE: 14 September 2023

**Product Identifier**

**RM Number:** 8671

**RM Name:** NIST mAb, Humanized IgG1 $\kappa$  Monoclonal Antibody

Under the U.S. Department of Labor, Occupational Safety and Health Administration (OSHA) 29 CFR 1910.1200, this Reference Material (RM) is NOT classified as a physical hazard or a health hazard, a simple asphyxiant, combustible dust, pyrophoric gas, or hazard not otherwise classified. There are no hazard pictograms, hazard statements or signal word associated with it. Safety Data Sheet information is not required. This document may be used in conjunction with your hazard communication program.

**Description:** This RM is intended primarily for use in evaluating the performance of methods for determining physicochemical and biophysical attributes of monoclonal antibodies. It also provides a representative test molecule for development of novel technology for therapeutic protein characterization. This RM can be used for a variety of uses that may include system suitability tests, establishing method or instrument performance and variability, comparing changing analytical test methods, and assisting in method qualification. A unit of RM 8671 consists of one cryovial containing 800  $\mu$ L of 10 mg/mL IgG1 $\kappa$  monoclonal antibody in 12.5 mmol/L L-histidine, 12.5 mmol/L L-histidine HCl (pH 6.0).

**Additional Notes:** RM 8671 IS INTENDED FOR RESEARCH USE. NOT INTENDED FOR ANIMAL OR HUMAN CONSUMPTION, CLINICAL TESTING, OR THERAPEUTIC USE. As a general rule, personal protective equipment should always be worn during any laboratory procedure. This includes, but is not limited to safety goggles, gloves, and a laboratory jacket.

**Disposal:** RM 8671 should be disposed of in accordance with local, state, and federal regulations.

**Transport Information:** This material is not regulated by the U.S. Department of Transportation (DOT) and/or International Air Transport Association (IATA).

**Disclaimer:** The NIST information in this document is specific to the NIST product and is believed to be correct, based upon our current knowledge. This document may not necessarily be all inclusive and should be used only as a guide. NIST does not guarantee the accuracy or completeness of this information. The only official source for specific values and uncertainties is the certificate or report.

Users of this RM should ensure that this document and the corresponding Report of Investigation in their possession are current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; e-mail [srmmsds@nist.gov](mailto:srmmsds@nist.gov); or via the Internet at <https://www.nist.gov/srm>.