

309.12 - Force Calibration

This Standard Reference Material is intended for validating methods to determining the stiffness of atomic force microscope cantilevers.

PLEASE NOTE: The tables are presented to facilitate comparisons among a family of materials to help customers select the best SRM for their needs. For specific values and uncertainties, the certificate is the only official source.

SRM	Description	Unit Size	Certified Value
3461	Standard Reference Cantilevers for Atomic Force Microscopy Spring Constant Calibration	each	Certified Stiffness Values for each Cantilever

-
- Certified values are normal font
 - Non-certified or reference values are italicized
 - Non-certified values in parentheses are for information only